



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Joshua Smith  
Director

**ATTESTATION**


I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-002591, in compliance with the agency's Maintenance Policy and Procedures.

Signed \_\_\_\_\_

Joshua Smith

Name of Position: Director/Records Custodian



State of Oklahoma

County of Oklahoma

Signed or attested before me on 10/07/2025

by Joshua Smith.

Jordan Faith Thompson  
(Signature of notarial officer)

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing.



3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

BENCH CHECK DATE: 04/11/2022

SERIAL: 80-002591

## DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis on Certified Solutions (4)
- Guth Laboratories Certificate of Analysis (4)



**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

## CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION


This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-002591 on 04/11/2022 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 35521080A3-003
  - BOT analysis average test result: 0.081
  - Cylinder expiration date: 03/05/2024
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

| Simulator Model          |           |           |           |
|--------------------------|-----------|-----------|-----------|
| Guth 2100                | Guth 2100 | Guth 2100 | Guth 2100 |
| Serial #                 |           |           |           |
| DR3685                   | DR3694    | DR3594    | DR3754    |
| Concentration            |           |           |           |
| .02                      | .04       | .10       | .20       |
| LOT #                    |           |           |           |
| 21120                    | 21070     | 21190     | 20510     |
| Manufactured Date        |           |           |           |
| 4/7/2021                 | 3/2/2021  | 6/9/2021  | 12/3/2020 |
| Expiration Date          |           |           |           |
| 4/7/2023                 | 3/1/2023  | 6/8/2023  | 12/3/2022 |
| Solution Commission Date |           |           |           |
| 4/11/2022                | 4/11/2022 | 4/11/2022 | 4/11/2022 |

Completed by BOT Personnel:

  
Signature/Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

Certified by:  
  
Signature Name Joshua Smith

State Director  
Title

# STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

## CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT

|                                |                                 |                                       |
|--------------------------------|---------------------------------|---------------------------------------|
| BENCH CHECK DATE<br>04/11/2022 | TIME PROCEDURE BEGAN<br>10:11:1 | INSTRUMENT LOCATION<br>80-002591      |
| INSTRUMENT TYPE<br>Intoxilyzer | INSTRUMENT MODEL<br>I-8000      | INSTRUMENT SERIAL NUMBER<br>80-002591 |

### TEST RESULTS

#### Diagnostics

PASS

#### Mouth Alcohol Test

PASS

#### RFI Test

PASS

#### Abort Test

PASS

#### Improper Sample Test

PASS

#### Interferent Detect Test

PASS

#### No Sample Given Test

PASS

#### Insufficient Sample Test

PASS

#### Air Blank Contamination Test

PASS

#### Range Exceeded Test

PASS

#### Duplicate Breath Sampling Results

| Value | Test 1 | Test 2 | Control Test |
|-------|--------|--------|--------------|
| .02   | 0.019  | 0.018  | 0.019        |
| .04   | 0.040  | 0.039  | 0.039        |
| .10   | 0.100  | 0.099  | 0.099        |
| .20   | 0.201  | 0.197  | 0.197        |

PASS

### CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

|                                                                                                  |                        |
|--------------------------------------------------------------------------------------------------|------------------------|
| SIGNATURE<br> | DATE<br>4/11/2022      |
| NAME<br>PAPE, CHRISTOPHER                                                                        | PERMIT #<br>0000023020 |



**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
 Chairman

Post Office Box 36307  
 Oklahoma City, Oklahoma 73136-2307  
 Phone: (405) 425-2460 Fax: (405) 425-2490  
 www.bot.ok.gov

Joshua Smith  
 Director

**Certificate of Analysis – Dry Gas**

ALPHA INSTRUMENT 2

Intoxilyzer – Alcohol Analyzer

Model 8000


SN 80-004923

03/25/2022

11:19

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 11:19 |
| Cal Check       | 0.081  | 11:20 |
| Air Blank       | 0.000  | 11:20 |
| Cal Check       | 0.081  | 11:21 |
| Air Blank       | 0.000  | 11:21 |
| Cal Check       | 0.081  | 11:21 |
| Air Blank       | 0.000  | 11:22 |
| Cal Check       | 0.080  | 11:22 |
| Air Blank       | 0.000  | 11:23 |
| Cal Check       | 0.081  | 11:23 |
| Air Blank       | 0.000  | 11:24 |
| Cal Check       | 0.081  | 11:24 |
| Air Blank       | 0.000  | 11:24 |
| Cal Check       | 0.081  | 11:25 |
| Air Blank       | 0.000  | 11:25 |
| Cal Check       | 0.081  | 11:26 |
| Air Blank       | 0.000  | 11:26 |
| Cal Check       | 0.082  | 11:26 |
| Air Blank       | 0.000  | 11:27 |
| Cal Check       | 0.081  | 11:27 |
| Air Blank       | 0.000  | 11:28 |
| Cal Check Stats |        |       |
| Average         | 0.0810 |       |
| Std Dev         | 0.0004 |       |
| Rel Std Dev(%)  | 0.5819 |       |

**Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)**  
**0.08 g/210L = 208 PPM ÷ 2605\***

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |                                      |                                    |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------|------------------------------------|
| Manufacturer: <b>ILMO</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                | Manufacturer Cert. No.: <b>14549</b> | Cylinder Size: <b>105L</b>         |
| Lot No.: <b>35521080A3</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | Cylinder No.: <b>003</b>             | Expiration Date: <b>03/05/2024</b> |
| The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM ÷ 2605]. |                                      |                                    |
| Signature:                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |                                      | Date: <b>03/25/2022</b>            |
| Name: <b>Christopher Pape</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |                                      | Permit No.: <b>23020</b>           |



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
 217-245-2183 • Fax: 217-243-7634 • [www.ilmoproducts.com](http://www.ilmoproducts.com)

## Certificate of Analysis

**Certificate ID:** 14549  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 35521080A3  
**Expiration:** 3/5/2024

RCV'D  
 2/16/2022  
 CP

**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

| Component: | Reported<br>Concentration: | Analytical<br>Accuracy<br>(U, k=2): | Analytical<br>Method: |
|------------|----------------------------|-------------------------------------|-----------------------|
| Ethanol    | 208 ppm                    | +/-0.002 BAC(6/210L)                | NDIR                  |
| Nitrogen   | Balance                    | [5.2 ppm]                           |                       |

**Distributed by:**

CMI Inc.  
 316 East Ninth Street  
 Owensboro, KY 42303  
 Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

\*Traceable to:

Certified Reference Material - 262.4 µmol/mol

Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat,  
 ignition and direct sunlight. Do not allow storage  
 area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

02-01-2022  
 Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory

# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

7/16

Section 1

|                                          |  |                            |  |                                      |        |                           |                          |
|------------------------------------------|--|----------------------------|--|--------------------------------------|--------|---------------------------|--------------------------|
| ARREST DATE<br><b>04/11/2022</b>         |  | ARREST TIME<br><b>1015</b> |  | COUNTY #<br><b>55</b>                | CITY # | CITATION #                |                          |
| ARREST LOCATION<br><b>BOT</b>            |  |                            |  | CITY                                 |        | COUNTY<br><b>OKLAHOMA</b> |                          |
| SUBJECT NAME<br><b>BREATHTEST, DELTA</b> |  |                            |  | DATE OF BIRTH<br><b>01/01/1985</b>   |        | SEX<br><b>M</b>           | HEIGHT<br><b>510</b>     |
|                                          |  |                            |  |                                      |        | WEIGHT<br><b>200</b>      |                          |
| ADDRESS<br><b>3600 MLK</b>               |  |                            |  | CITY<br><b>OKLAHOMA CITY</b>         |        | STATE<br><b>OK</b>        | ZIP CODE<br><b>73111</b> |
| DRIVER LICENSE #<br><b>B083463042</b>    |  |                            |  | EXPIRATION DATE<br><b>10/31/2022</b> |        | STATE<br><b>OK</b>        | CLASS<br><b>D</b>        |
|                                          |  |                            |  |                                      |        | RESTRICTIONS              | ENDORSEMENTS             |
| VEHICLE MAKE<br><b>MOCK</b>              |  |                            |  | MODEL<br><b>MOCK</b>                 |        | TAG #<br><b>MOCK</b>      | STATE<br><b>OK</b>       |
|                                          |  |                            |  |                                      |        | CMV<br><b>N</b>           | PLACARD<br><b>N</b>      |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

**MOCK SUBJECT TEST**

Describe person's condition (odor, actions, etc.):

**ROUTINE MAINTENANCE**

## THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

|                                                                                                                                                             |                                |                                                                                                                                                                                                                                                                                                                                                          |                                                                                                                                                                                                                         |                                            |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------|
| EQUIPMENT<br><b>I-8000</b>                                                                                                                                  |                                | SERIAL NUMBER<br><b>80-002591</b>                                                                                                                                                                                                                                                                                                                        | SPECIALIST<br><b>BOARD OF TESTS</b>                                                                                                                                                                                     |                                            |
| STANDARD<br><b>Dry</b>                                                                                                                                      |                                | TARGET VALUE<br><b>0.080</b>                                                                                                                                                                                                                                                                                                                             | PERMIT NUMBER<br><b>23020</b>                                                                                                                                                                                           |                                            |
| MANUFACTURER<br><b>ILMO</b>                                                                                                                                 |                                | LOT #<br><b>35521080A3</b>                                                                                                                                                                                                                                                                                                                               | DATE CYLINDER INSTALLED<br><b>04/11/2022</b>                                                                                                                                                                            | DATE CYLINDER EXPIRES<br><b>03/05/2024</b> |
| OBSERVATION START<br><b>1030</b>                                                                                                                            | TEST DATE<br><b>04/11/2022</b> | <p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>Christopher Pape</i><br/>SIGNATURE OF OPERATOR</p> |                                                                                                                                                                                                                         |                                            |
| RESULT TABLE                                                                                                                                                |                                |                                                                                                                                                                                                                                                                                                                                                          |                                                                                                                                                                                                                         |                                            |
| Test                                                                                                                                                        | g/210L                         | Time                                                                                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                                         |                                            |
| Air Blank                                                                                                                                                   | 0.00                           | 10:55                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Subject Test                                                                                                                                                | 0.00                           | 10:56                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Breath Volume                                                                                                                                               | 1.953 LITERS                   |                                                                                                                                                                                                                                                                                                                                                          |                                                                                                                                                                                                                         |                                            |
| Air Blank                                                                                                                                                   | 0.00                           | 10:56                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Wait                                                                                                                                                        |                                | 10:58                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Air Blank                                                                                                                                                   | 0.00                           | 10:58                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Subject Test                                                                                                                                                | 0.00                           | 10:59                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Breath Volume                                                                                                                                               | 2.031 LITERS                   |                                                                                                                                                                                                                                                                                                                                                          |                                                                                                                                                                                                                         |                                            |
| Air Blank                                                                                                                                                   | 0.00                           | 10:59                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Cal Check                                                                                                                                                   | 0.08                           | 10:59                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| Air Blank                                                                                                                                                   | 0.00                           | 11:00                                                                                                                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                         |                                            |
| <p>NAME<br/><b>PAPE, CHRISTOPHER</b></p> <p>BADGE #<br/><b>0000000484</b></p> <p>PERMIT #<br/><b>0000023020</b></p> <p>AGENCY<br/><b>BOARD OF TESTS</b></p> |                                |                                                                                                                                                                                                                                                                                                                                                          |  <div style="border: 1px solid black; padding: 5px; text-align: center;"> <p>Test Result<br/>BrAC g/210L<br/><b>0.00</b></p> </div> |                                            |
|                                                                                                                                                             |                                |                                                                                                                                                                                                                                                                                                                                                          |                                                                                                                                                                                                                         |                                            |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below.

You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

|                                  |                                       |                                 |                            |
|----------------------------------|---------------------------------------|---------------------------------|----------------------------|
| NAME<br><b>PAPE, CHRISTOPHER</b> |                                       | AGENCY<br><b>BOARD OF TESTS</b> |                            |
| BADGE<br><b>0000000484</b>       | AGENCY ADDRESS<br><b>INTER-AGENCY</b> | ZIP CODE<br><b>00000</b>        | PHONE<br><b>4054252460</b> |

OTHER WITNESSES:

Section 5

| # | NAME | TITLE | ADDRESS | PHONE |
|---|------|-------|---------|-------|
| 1 |      |       |         |       |
| 2 |      |       |         |       |

OFFICER/DPS COPY

BOT Form AFF01 09/2021

# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

8/16

Section 1

|                                   |                               |                             |             |                                |
|-----------------------------------|-------------------------------|-----------------------------|-------------|--------------------------------|
| ARREST DATE<br>04/11/2022         | ARREST TIME<br>1015           | COUNTY #<br>55              | CITY #      | CITATION #                     |
| ARREST LOCATION<br>BOT            |                               | CITY                        |             | COUNTY<br>OKLAHOMA             |
| SUBJECT NAME<br>BREATHTEST, DELTA |                               | DATE OF BIRTH<br>01/01/1985 | SEX<br>M    | HEIGHT<br>510<br>WEIGHT<br>200 |
| ADDRESS<br>3600 MLK               |                               | CITY<br>OKLAHOMA CITY       | STATE<br>OK | ZIP CODE<br>73111              |
| DRIVER LICENSE #<br>B083463042    | EXPIRATION DATE<br>10/31/2022 | STATE<br>OK                 | CLASS<br>D  | RESTRICTIONS<br>ENDORSEMENTS   |
| VEHICLE MAKE<br>MOCK              | MODEL<br>MOCK                 | TAG #<br>MOCK               | STATE<br>OK | CMV<br>N<br>PLACARD<br>N       |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):

ROUTINE MAINTENANCE

## THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

| EQUIPMENT<br>I-8000                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     | SERIAL NUMBER<br>80-002591 | SPECIALIST<br>BOARD OF TESTS                                                                                                     |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------|----------------------------------------------------------------------------------------------------------------------------------|-------------------------------------|------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|------|--|-------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|-----------|------|-------|-----------|------|-------|---------------------------------------------------------------------------------------------------------------|--|
| STANDARD<br>Dry                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         | TARGET VALUE<br>0.080      | PERMIT NUMBER<br>23020                                                                                                           |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| MANUFACTURER<br>ILMO                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    | LOT #<br>35521080A3        | DATE CYLINDER INSTALLED<br>04/11/2022                                                                                            | DATE CYLINDER EXPIRES<br>03/05/2024 |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| OBSERVATION START<br>1030                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | TEST DATE<br>04/11/2022    | A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration. |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| RESULT TABLE                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |                            | I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.             |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| <table border="1"><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.00</td><td>10:55</td></tr><tr><td>Subject Test</td><td>0.00</td><td>10:56</td></tr><tr><td>Breath Volume</td><td>1.953 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>10:56</td></tr><tr><td>Wait</td><td></td><td>10:58</td></tr><tr><td>Air Blank</td><td>0.00</td><td>10:58</td></tr><tr><td>Subject Test</td><td>0.00</td><td>10:59</td></tr><tr><td>Breath Volume</td><td>2.031 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>10:59</td></tr><tr><td>Cal Check</td><td>0.08</td><td>10:59</td></tr><tr><td>Air Blank</td><td>0.00</td><td>11:00</td></tr></tbody></table> |                            | Test                                                                                                                             | g/210L                              | Time | Air Blank | 0.00 | 10:55 | Subject Test | 0.00 | 10:56 | Breath Volume | 1.953 LITERS |  | Air Blank | 0.00 | 10:56 | Wait |  | 10:58 | Air Blank | 0.00 | 10:58 | Subject Test | 0.00 | 10:59 | Breath Volume | 2.031 LITERS |  | Air Blank | 0.00 | 10:59 | Cal Check | 0.08 | 10:59 | Air Blank | 0.00 | 11:00 | <br>SIGNATURE OF OPERATOR |  |
| Test                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    | g/210L                     | Time                                                                                                                             |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Air Blank                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | 0.00                       | 10:55                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Subject Test                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            | 0.00                       | 10:56                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Breath Volume                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           | 1.953 LITERS               |                                                                                                                                  |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Air Blank                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | 0.00                       | 10:56                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Wait                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    |                            | 10:58                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Air Blank                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | 0.00                       | 10:58                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Subject Test                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            | 0.00                       | 10:59                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Breath Volume                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           | 2.031 LITERS               |                                                                                                                                  |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Air Blank                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | 0.00                       | 10:59                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Cal Check                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | 0.08                       | 10:59                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| Air Blank                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | 0.00                       | 11:00                                                                                                                            |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| NAME<br>PAPE, CHRISTOPHER                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |                            | Test Result<br>BrAC g/210L<br>0.00                                                                                               |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| BADGE #<br>0000000484                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |                            | PERMIT #<br>0000023020                                                                                                           |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |
| AGENCY<br>BOARD OF TESTS                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |                            |                                                                                                                                  |                                     |      |           |      |       |              |      |       |               |              |  |           |      |       |      |  |       |           |      |       |              |      |       |               |              |  |           |      |       |           |      |       |           |      |       |                                                                                                               |  |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below.

You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

|                                                                                       |                                |
|---------------------------------------------------------------------------------------|--------------------------------|
| Date: _____ Place (location when signed): _____ Signature of arresting officer: _____ |                                |
| NAME<br>PAPE, CHRISTOPHER                                                             | AGENCY<br>BOARD OF TESTS       |
| BADGE<br>0000000484                                                                   | AGENCY ADDRESS<br>INTER-AGENCY |
| ZIP CODE<br>00000                                                                     | PHONE<br>4054252460            |

ARRESTEE'S COPY

BOT Form AFF02 09/2021



9/16

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 11:33

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 11:34 |
| Cal Check       | 0.022  | 11:34 |
| Air Blank       | 0.000  | 11:35 |
| Cal Check       | 0.021  | 11:36 |
| Air Blank       | 0.000  | 11:36 |
| Cal Check       | 0.021  | 11:37 |
| Air Blank       | 0.000  | 11:37 |
| Cal Check       | 0.021  | 11:38 |
| Air Blank       | 0.000  | 11:38 |
| Cal Check       | 0.021  | 11:39 |
| Air Blank       | 0.000  | 11:40 |
| Cal Check       | 0.021  | 11:40 |
| Air Blank       | 0.000  | 11:41 |
| Cal Check       | 0.020  | 11:42 |
| Air Blank       | 0.000  | 11:42 |
| Cal Check       | 0.021  | 11:43 |
| Air Blank       | 0.000  | 11:43 |
| Cal Check       | 0.021  | 11:44 |
| Air Blank       | 0.000  | 11:45 |
| Cal Check       | 0.020  | 11:45 |
| Air Blank       | 0.000  | 11:46 |
| Cal Check Stats |        |       |
| Average         | 0.0209 |       |
| Std Dev         | 0.0005 |       |
| Rel Std Dev(%)  | 2.7158 |       |

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |                               |                            |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------|----------------------------|
| Manufacturer: Guth                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   | Manufactured Date: 04/07/2021 | Concentration: 0.02        |
| Lot No.: 21120                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       | Expiration Date: 04/07/2023   | Equipment Model: Guth 2100 |
| The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration. |                               |                            |
| Signature:                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        | Date: 03/08/2022              |                            |
| Name: Christopher Pape                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | Permit No.: 23020             |                            |



10/16

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer – Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 11:53

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 11:53 |
| Cal Check       | 0.041  | 11:54 |
| Air Blank       | 0.000  | 11:54 |
| Cal Check       | 0.041  | 11:55 |
| Air Blank       | 0.000  | 11:56 |
| Cal Check       | 0.042  | 11:56 |
| Air Blank       | 0.000  | 11:57 |
| Cal Check       | 0.042  | 11:58 |
| Air Blank       | 0.000  | 11:58 |
| Cal Check       | 0.042  | 11:59 |
| Air Blank       | 0.000  | 11:59 |
| Cal Check       | 0.041  | 12:00 |
| Air Blank       | 0.000  | 12:01 |
| Cal Check       | 0.041  | 12:01 |
| Air Blank       | 0.000  | 12:02 |
| Cal Check       | 0.041  | 12:02 |
| Air Blank       | 0.000  | 12:03 |
| Cal Check       | 0.041  | 12:04 |
| Air Blank       | 0.000  | 12:04 |
| Cal Check       | 0.042  | 12:05 |
| Air Blank       | 0.000  | 12:05 |
| Cal Check Stats |        |       |
| Average         | 0.0414 |       |
| Std Dev         | 0.0005 |       |
| Rel Std Dev(%)  | 1.2473 |       |

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |                                      |                                   |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------|-----------------------------------|
| Manufacturer: <b>Guth</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            | Manufactured Date: <b>03/01/2021</b> | Concentration: <b>0.04</b>        |
| Lot No.: <b>21070</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                | Expiration Date: <b>03/01/2023</b>   | Equipment Model: <b>Guth 2100</b> |
| The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration. |                                      |                                   |
| Signature<br>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     | Date: <b>03/08/2022</b>              |                                   |
| Name <b>Christopher Pape</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         | Permit No.: <b>23020</b>             |                                   |



11/16

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D., F-ABFT  
ChairmanJoshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-004923

03/08/2022

12:11

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 12:11 |
| Cal Check       | 0.101  | 12:12 |
| Air Blank       | 0.000  | 12:13 |
| Cal Check       | 0.101  | 12:13 |
| Air Blank       | 0.000  | 12:14 |
| Cal Check       | 0.101  | 12:15 |
| Air Blank       | 0.000  | 12:15 |
| Cal Check       | 0.102  | 12:16 |
| Air Blank       | 0.000  | 12:16 |
| Cal Check       | 0.101  | 12:17 |
| Air Blank       | 0.000  | 12:18 |
| Cal Check       | 0.101  | 12:18 |
| Air Blank       | 0.000  | 12:19 |
| Cal Check       | 0.101  | 12:20 |
| Air Blank       | 0.000  | 12:20 |
| Cal Check       | 0.101  | 12:21 |
| Air Blank       | 0.000  | 12:21 |
| Cal Check       | 0.101  | 12:22 |
| Air Blank       | 0.000  | 12:23 |
| Cal Check       | 0.101  | 12:23 |
| Air Blank       | 0.000  | 12:24 |
| Cal Check Stats |        |       |
| Average         | 0.1011 |       |
| Std Dev         | 0.0003 |       |
| Rel Std Dev (%) | 0.3127 |       |

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: Guth

Manufactured Date: 06/08/2021

Concentration: 0.10

Lot No.: 21190

Expiration Date: 06/08/2023

Equipment Model: Guth 2100

The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.

Signature

Date:

03/08/2022

Name Christopher Pape

Permit No.: 23020



12/16

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 12:28

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 12:29 |
| Cal Check       | 0.201  | 12:29 |
| Air Blank       | 0.000  | 12:30 |
| Cal Check       | 0.200  | 12:31 |
| Air Blank       | 0.000  | 12:31 |
| Cal Check       | 0.201  | 12:32 |
| Air Blank       | 0.000  | 12:33 |
| Cal Check       | 0.201  | 12:33 |
| Air Blank       | 0.000  | 12:34 |
| Cal Check       | 0.201  | 12:35 |
| Air Blank       | 0.000  | 12:35 |
| Cal Check       | 0.201  | 12:36 |
| Air Blank       | 0.000  | 12:36 |
| Cal Check       | 0.201  | 12:37 |
| Air Blank       | 0.000  | 12:38 |
| Cal Check       | 0.200  | 12:38 |
| Air Blank       | 0.000  | 12:39 |
| Cal Check       | 0.201  | 12:40 |
| Air Blank       | 0.000  | 12:40 |
| Cal Check       | 0.200  | 12:41 |
| Air Blank       | 0.000  | 12:41 |
| Cal Check Stats |        |       |
| Average         | 0.2007 |       |
| Std Dev         | 0.0004 |       |
| Rel Std Dev(%)  | 0.2406 |       |

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |                               |                            |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------|----------------------------|
| Manufacturer: Guth                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   | Manufactured Date: 12/03/2020 | Concentration: 0.20        |
| Lot No.: 20510                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       | Expiration Date: 12/03/2022   | Equipment Model: Guth 2100 |
| The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration. |                               |                            |
| Signature:                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        | Date: 03/08/2022              |                            |
| Name: Christopher Pape                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | Permit No.: 23020             |                            |

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 21120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 16, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0237% (w/vol) ethyl alcohol. The expiration date for this lot number is April 7, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08031603 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

RCV'D  
8/26/21  
CP

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**RCV'D  
8/26/21  
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 2, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0488%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 1, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**RCV'D  
8/26/21  
CP**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 21190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on June 9, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is June 8, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 20510 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 7, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2419% (w/vol) ethyl alcohol. The expiration date for this lot number is December 3, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN06231703 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

RCV'D  
8/26/21  
CP