

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing



3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

**BENCH CHECK DATE:** 04-27-2018

**SERIAL:** 80-002670

## **DOCUMENTATION WITHIN SERVICE PACKET:**

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)



Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,

A handwritten signature in black ink, appearing to read "J. Robert Blakeburn". The signature is written in a cursive style with some flourishes.

J. Robert Blakeburn (OBA11059)  
State Director of Tests for Alcohol  
and Drug Influence

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY  
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE <b>04/27/2018</b>	TIME PROCEDURE BEGAN <b>10:39</b>	INSTRUMENT LOCATION <b>80-002670</b>
INSTRUMENT TYPE <b>Intoxilyzer</b>	INSTRUMENT MODEL <b>I-8000</b>	INSTRUMENT SERIAL NUMBER <b>80-002670</b>

**TEST RESULTS**

**Diagnostics**

PASS

**Mouth Alcohol Test**

PASS

**RFI Test**

PASS

**Abort Test**

PASS

**Improper Sample Test**

PASS

**Interferent Detect Test**

PASS

**No Sample Given Test**

PASS

**Insufficient Sample Test**

PASS

**Air Blank Contamination Test**

PASS

**Range Exceeded Test**

PASS


**Duplicate Breath Sampling Results**

Value	Test 1	Test 2	Control Test
.02	0.016	0.017	0.018
.04	0.036	0.035	0.037
.10	0.093	0.095	0.096
.20	0.199	0.197	0.200

PASS

**CERTIFYING TECHNICIAN**

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE <b>04/27/2018</b>
NAME <b>SMITH, JOSHUA</b>	PERMIT # <b>0000016837</b>



3600 N. Martin Luther King Bldg #9  
 Oklahoma City, OK 73111  
 405-425-2460

# BENCH CHECK TECHNICIAN'S REPORT

Date: 4 / 27 / 18 Start Time: 1035 End Time: 1315

## INSTRUMENT

Intoxilyzer Model 8000 Serial # 80-002670

## GAS CANISTER

LOT # 02618080A3 # 004 EXP Date 3-5-2020


## REFERENCE

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR3591	DR3753	DR3594	DR3754
Concentration			
.02	.04	.10	.20
LOT #			
16400	16320	17130	17090
Manufactured Date			
12-12-2016	10-21-2016	04-03-2017	02-24-2017
Expiration Date			
12-12-2018	10-21-2018	04-03-2019	02-24-2019
Solution Commission Date			
<u>4-4-18</u>	<u>4-4-18</u>	<u>4-4-18</u>	<u>4-4-18</u>

## REASON FOR BENCH CHECK

- POST REPAIR     CYLINDER REPLACEMENT  
 TROUBLESHOOTING     ROUTINE MAINTENANCE

## COMMENTS



  
 Technician Signature / Permit # 16837  
 Printed Name Josh Smith

Certificate of Analysis - Dry Gas	
04-27-18	0.080
Date of Analysis	Labelled target value (g/210L)
# 004	0.0791
Cylinder #	Average test result
JOSH SMITH <i>[Signature]</i>	
BOT Technician Name and Signature	



BRT Form 5.0

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
 Tel: 217-243-2183 Fax: 217-243-7634 • www.ilmoproducts.com

In accordance with BOT Policy and Procedure Statement 09-21.0, Maintenance, the above referenced dry gas cylinder is suitable for use as an analytical standard.

ALPHA INSTRUMENT 2  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004923  
 04/27/2018 10:54

Test	g/210L	Time
Air Blank	0.000	10:55
Cal Check	0.079	10:55
Air Blank	0.000	10:55
Cal Check	0.079	10:56
Air Blank	0.000	10:56
Cal Check	0.079	10:57
Air Blank	0.000	10:57
Cal Check	0.079	10:57
Air Blank	0.000	10:58
Cal Check	0.079	10:58
Air Blank	0.000	10:59
Cal Check	0.080	10:59
Air Blank	0.000	11:00
Cal Check	0.079	11:00
Air Blank	0.000	11:00
Cal Check	0.079	11:01
Air Blank	0.000	11:01
Cal Check	0.079	11:02
Air Blank	0.000	11:02
Cal Check	0.079	11:02
Air Blank	0.000	11:03
Cal Check Stats		
Average	0.0791	
Std Dev	0.0003	
Rel Std Dev(%)	0.3997	

## Certificate of Analysis

Certificate ID: 10864  
 Part #: BAC105L080T  
 Cylinder Size: 105L  
 Lot Number: 02618080A3 ✓  
 Expiration: 3/5/2020 ✓

*RECD*  
*03/20/18*  
*[Signature]*

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.  
 316 East Ninth Street  
 Owensboro, KY 4230  
 Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

*[Signature]* 16837  
 Operator's Signature

\*NIST Traceable Reference Material  
 Cylinder No. CC274523 / Job No. 09160306  
 Certified 362.2 µmol/mol Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

*[Signature]*  
 Specialty Gas Lab Tech

02-16-18  
 Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2005 Accredited Laboratory



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE <b>04/27/2018</b>		ARREST TIME <b>1240</b>		COUNTY # <b>55</b>	CITY # <b>73</b>	CITATION # <b>MOCK</b>		
ARREST LOCATION <b>MOCK</b>					CITY <b>MOCK</b>	COUNTY <b>OKLAHOMA</b>		
SUBJECT NAME <b>MOCK, SUBJECT TEST</b>				DATE OF BIRTH <b>08/09/1961</b>	SEX <b>M</b>	RACE <b>W</b>	HEIGHT <b>510</b>	WEIGHT <b>165</b>
ADDRESS <b>3600 MARTIN LUTHER KING A</b>				CITY <b>OKLAHOMA CITY</b>		STATE <b>OK</b>	ZIP CODE <b>73133</b>	
DRIVER LICENSE # <b>T083454871</b>				EXPIRATION DATE <b>06/30/2019</b>	STATE <b>OK</b>	CLASS <b>A</b>	RESTRICTIONS <b>NONE</b>	ENDORSEMENTS <b>NONE</b>
VEHICLE MAKE <b>NONE</b>			MODEL <b>NONE</b>	TAG # <b>NONE</b>		STATE <b>OK</b>	CMV <b>N</b>	PLACARD <b>N</b>

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

**MOCK SUBJECT TEST. ROUTINE MAINTENANCE. NEW GAS CANISTER INSTALL.**

Describe person's condition (odor, actions, etc.):

**NONE**

### THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT <b>I-8000</b>		SERIAL NUMBER <b>80-002670</b>	SPECIALIST <b>BOARD OF TESTS</b>																																					
STANDARD <b>Dry</b>		TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>16837</b>																																					
MANUFACTURER <b>ILMO</b>		LOT # <b>02618080A3</b>	DATE CYLINDER INSTALLED <b>04/27/2018</b>	DATE CYLINDER EXPIRES <b>03/05/2020</b>																																				
OBSERVATION START <b>1255</b>	TEST DATE <b>04/27/2018</b>	MOUTHPIECE <b>G</b>	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p>																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>13:13</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:14</td></tr> <tr><td>  Breath Volume</td><td>1.628 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:14</td></tr> <tr><td>Wait</td><td></td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:16</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:17</td></tr> <tr><td>  Breath Volume</td><td>2.437 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:17</td></tr> <tr><td>Cal Check</td><td>0.07</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:18</td></tr> </tbody> </table>						Test	g/210L	Time	Air Blank	0.00	13:13	Subject Test	0.00	13:14	Breath Volume	1.628 LITERS		Air Blank	0.00	13:14	Wait		13:16	Air Blank	0.00	13:16	Subject Test	0.00	13:17	Breath Volume	2.437 LITERS		Air Blank	0.00	13:17	Cal Check	0.07	13:18	Air Blank	0.00
Test	g/210L	Time																																						
Air Blank	0.00	13:13																																						
Subject Test	0.00	13:14																																						
Breath Volume	1.628 LITERS																																							
Air Blank	0.00	13:14																																						
Wait		13:16																																						
Air Blank	0.00	13:16																																						
Subject Test	0.00	13:17																																						
Breath Volume	2.437 LITERS																																							
Air Blank	0.00	13:17																																						
Cal Check	0.07	13:18																																						
Air Blank	0.00	13:18																																						
NAME <b>SMITH, JOSHUA</b>			BADGE # <b>0000000435</b>		PERMIT # <b>0000016837</b>																																			
AGENCY <b>BOARD OF TESTS</b>																																								

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation** is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

**SERVING OF NOTICE:** I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **04/27/2018** Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **04/27/2018** Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME <b>SMITH, JOSHUA</b>		AGENCY <b>BOARD OF TESTS</b>	
BADGE <b>0000000435</b>	AGENCY ADDRESS <b>INTER-AGENCY</b>	ZIP CODE <b>00000</b>	PHONE <b>"4054252460"</b>

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

No Temporary License Issued:  
Test Below Legal Limit

OFFICER/DPS COPY  
BOT Form AFF01 07/2008

# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE <b>04/27/2018</b>		ARREST TIME <b>1240</b>		COUNTY # <b>55</b>	CITY # <b>73</b>	CITATION # <b>MOCK</b>			
ARREST LOCATION <b>MOCK</b>				CITY <b>MOCK</b>		COUNTY <b>OKLAHOMA</b>			
SUBJECT NAME <b>MOCK, SUBJECT TEST</b>					DATE OF BIRTH <b>08/09/1961</b>	SEX <b>M</b>	RACE <b>W</b>	HEIGHT <b>510</b>	WEIGHT <b>165</b>
ADDRESS <b>3600 MARTIN LUTHER KING A</b>				CITY <b>OKLAHOMA CITY</b>		STATE <b>OK</b>	ZIP CODE <b>73133</b>		
DRIVER LICENSE # <b>T083454871</b>				EXPIRATION DATE <b>06/30/2019</b>	STATE <b>OK</b>	CLASS <b>A</b>	RESTRICTIONS <b>NONE</b>	ENDORSEMENTS <b>NONE</b>	
VEHICLE MAKE <b>NONE</b>			MODEL <b>NONE</b>	TAG # <b>NONE</b>		STATE <b>OK</b>	CMV <b>N</b>	PLACARD <b>N</b>	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

**MOCK SUBJECT TEST. ROUTINE MAINTENANCE. NEW GAS CANISTER INSTALL.**

Describe person's condition (odor, actions, etc.):

**NONE**

**THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST**

**BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.**

Section 2

EQUIPMENT <b>I-8000</b>		SERIAL NUMBER <b>80-002670</b>	SPECIALIST <b>BOARD OF TESTS</b>																																								
STANDARD <b>Dry</b>		TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>16837</b>																																								
MANUFACTURER <b>ILMO</b>		LOT # <b>02618080A3</b>	DATE CYLINDER INSTALLED <b>04/27/2018</b>	DATE CYLINDER EXPIRES <b>03/05/2020</b>																																							
OBSERVATION START <b>1255</b>	TEST DATE <b>04/27/2018</b>	MOUTHPIECE <b>G</b>	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td colspan="3">-----</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:13</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:14</td></tr> <tr><td>  Breath Volume</td><td>1.628 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:14</td></tr> <tr><td>Wait</td><td></td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:16</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:17</td></tr> <tr><td>  Breath Volume</td><td>2.437 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:17</td></tr> <tr><td>Cal Check</td><td>0.07</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:18</td></tr> </tbody> </table>					Test	g/210L	Time	-----			Air Blank	0.00	13:13	Subject Test	0.00	13:14	Breath Volume	1.628 LITERS		Air Blank	0.00	13:14	Wait		13:16	Air Blank	0.00	13:16	Subject Test	0.00	13:17	Breath Volume	2.437 LITERS		Air Blank	0.00	13:17	Cal Check	0.07	13:18	Air Blank	0.00	13:18
Test	g/210L	Time																																									
-----																																											
Air Blank	0.00	13:13																																									
Subject Test	0.00	13:14																																									
Breath Volume	1.628 LITERS																																										
Air Blank	0.00	13:14																																									
Wait		13:16																																									
Air Blank	0.00	13:16																																									
Subject Test	0.00	13:17																																									
Breath Volume	2.437 LITERS																																										
Air Blank	0.00	13:17																																									
Cal Check	0.07	13:18																																									
Air Blank	0.00	13:18																																									
RESULT TABLE			NAME <b>SMITH, JOSHUA</b>																																								
			BADGE # <b>0000000435</b>	PERMIT # <b>0000016837</b>																																							
			AGENCY <b>BOARD OF TESTS</b>																																								

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation** is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

**SERVING OF NOTICE:** I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **04/27/2018** Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **04/27/2018** Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME <b>SMITH, JOSHUA</b>		AGENCY <b>BOARD OF TESTS</b>	
BADGE <b>0000000435</b>	AGENCY ADDRESS <b>INTER-AGENCY</b>	ZIP CODE <b>00000</b>	PHONE <b>"4054252460"</b>

No Temporary License Issued:  
Test Below Legal Limit

ARRESTEE'S COPY  
BOT Form AFF02 07/2008



Certificate of Analysis - Wet Bath	
05-05-17	.020
Date of Analysis	Labeled target value (g/210L)
16400	.0195
Lot Number	Average test result (g/210L)
Jbst Smith	
BOT Technician Name and Signature	



**GUTH LABORATORIES, INC.**  
 G, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

INTOXILYZER 8000  
 Instrument Initialization  
 8133.12  
 14:48 05/05/2017

Certified Alcohol Reference Solution for Simulator

Rec'd  
 4/01/2017  
 #46837

Random Samples of Lot Number 16400 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 13, 2016, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain

ALPHA INSTRUMENT 2  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004923  
 05/05/2017 14:56

Test	g/210L	Time
Air Blank	0.000	14:57
Cal Check	0.020	14:58
Air Blank	0.000	14:58
Cal Check	0.020	14:59
Air Blank	0.000	14:59
Cal Check	0.019	15:00
Air Blank	0.000	15:01
Cal Check	0.020	15:01
Air Blank	0.000	15:02
Cal Check	0.020	15:03
Air Blank	0.000	15:03
Cal Check	0.020	15:04
Air Blank	0.000	15:04
Cal Check	0.019	15:05
Air Blank	0.000	15:06
Cal Check	0.019	15:06
Air Blank	0.000	15:07
Cal Check	0.019	15:08
Air Blank	0.000	15:08
Cal Check	0.019	15:09
Air Blank	0.000	15:09
Cal Check Stats		
Average	0.0195	
Std Dev	0.0005	
Rel Std Dev(%)	2.7026	

0.0241% (w/vol) ethyl alcohol. The expiration date for this lot number is December 12, 2018 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

*Ted L. Pauley*  
 Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

*NIST Traceability:*  
 Testing was conducted using Cerilliant Reference Standard lot number FN11241503 whose values are traceable to NIST.  
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

#46837

**Certificate of Analysis - Wet Bath**

Date of Analysis	3-8-2017	Labelled target value (g/210L)	.040
Lot Number	16320	Average test result (g/210L)	.0409
JSL Smith		[Signature]	
BRT Technician Name and Signature			



BRT Form 6.0

In accordance with BRT - 2.1.0 (Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

**TORIES, INC.**  
 BURG, PA 17111-4511 • TELEPHONE: 717-664-6470

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **16320** of  
 Alcohol Reference Solution for Simulator were analyzed by

*REC'D 7/11/16 4:58 PM*

ALPHA INSTRUMENT 2  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 03/08/2017

gas chromatography on **October 25, 2016**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain

Test g/210L Time number is **October 21, 2018** at 11:59 PM.

**0.0495%** (w/vol) ethyl alcohol. The expiration date for this lot

Air Blank	0.000	11:16	When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of <b>0.040 g/210L +/- 3%</b> . ✓
Cal Check	0.041	11:17	
Air Blank	0.000	11:18	
Cal Check	0.041	11:19	
Air Blank	0.000	11:19	
Cal Check	0.040	11:20	
Air Blank	0.000	11:20	
Cal Check	0.041	11:21	
Air Blank	0.000	11:21	
Cal Check	0.041	11:22	
Air Blank	0.000	11:22	The alcohol and water used in this solution were free of test interfering substances.
Cal Check	0.041	11:22	
Air Blank	0.000	11:23	
Cal Check	0.041	11:23	
Air Blank	0.000	11:24	
Cal Check	0.041	11:24	
Air Blank	0.000	11:25	
Cal Check	0.040	11:25	
Air Blank	0.000	11:26	
Cal Check	0.042	11:27	
Air Blank	0.000	11:27	
Cal Check	0.041	11:28	
Air Blank	0.000	11:29	
Cal Check Stats			
Average	0.0409		
Std Dev	0.0005		
Rel Std Dev(%)	1.3878		

*Ted L. Pauley*  
 Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

*NIST Traceability:*  
 Testing was conducted using Cerilliant Reference Standard lot number **EN1218501** whose values are traceable to NIST.  
 All balances are calibrated annually by an outside agency using NIST traceable weights.  
 Calibration verification is done prior to each use utilizing NIST traceable weights.

*11/08/17*



Certificate of Analysis - Wet Bath

Date of Analysis	04/15/2017	Labelled target value (g/210L)	1.10
Lot Number	17130	Average test result (g/210L)	1.1015
BRT Technician Name and Signature <i>Josh Swartz</i>			



BRT form 6.0

**LABORATORIES, INC.**  
SEBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
SN 80-004923  
06/15/2017  
Certified Alcohol Reference Solution for Simulator

*Rec'd  
04/09/2017*

Random Samples of Lot Number 17130 of g/210L

Test	g/210L	Time
Air Blank	0.000	16:26
Cal Check	0.102	16:26
Air Blank	0.000	16:27
Cal Check	0.102	16:28
Air Blank	0.000	16:28
Cal Check	0.101	16:29
Air Blank	0.000	16:29
Cal Check	0.102	16:30
Air Blank	0.000	16:31
Cal Check	0.101	16:31
Air Blank	0.000	16:32
Cal Check	0.101	16:33
Air Blank	0.000	16:33
Cal Check	0.102	16:34
Air Blank	0.000	16:34
Cal Check	0.101	16:35
Air Blank	0.000	16:35
Cal Check	0.102	16:36
Air Blank	0.000	16:37
Cal Check	0.101	16:38
Air Blank	0.000	16:38
Average	0.1015	
Std Dev	0.0005	
Rel Std Dev(%)	0.5192	

Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on April 6, 2017, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1212% (w/vol) ethyl alcohol. The expiration date for this lot number is April 3, 2019 at 11:59 PM. ✓

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%. ✓

The alcohol and water used in this solution were free of test interfering substances.

*Ted L. Pauley*  
Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*[Signature]*  
Operator's Signature  
16837

NIST Traceability:  
Testing was conducted using Certified Reference Standard lot number FN08051301 whose values are traceable to NIST.  
All balances are calibrated annually by an outside agency using NIST traceable weights.  
Calibration verification is done prior to each use utilizing NIST traceable weights.

