

Service Overview

Board of Tests for Alcohol and Drug Influence

Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 01-23-2017

SERIAL: 80-002673

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- CMI Certificate of Calibration 01-04-2017
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- CMI Service Evaluation Form
- CMI Work Order
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Administrative Office at:
Board of Tests for Alcohol and Drug Influence
Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307

Certificate of Calibration and Operation

This is to certify that the calibration of **INTOXILYZER 8000**, serial number 80-002673, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R 126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.

In addition the operational procedures and functions of this Intoxilyzer have also been tested and found to be in compliance with the Oklahoma Board of Tests for Alcohol and Drug Influence - Rules and Regulations OAC Title 40: 25-1-2 and 30-1-3 and is hereby certified for use as an evidential breath testing device within this state.

CERTIFICATION TYPE	DATE
Manufacturer Calibration	09-11-07
Board of Tests Bench Check	06-03-08
Board of Tests Bench Check	10-06-08
Board of Tests Bench Check	08-11-09
Board of Tests Bench Check	03-22-10
Manufacturer Calibration	10-06-11
Board of Tests Bench Check	10-24-11
Board of Tests Bench Check	10-04-13
Board of Tests Bench Check	09-11-15
Manufacturer Calibration	01-04-17
Board of Tests Bench Check	01-23-17



Presentation of this form certifies that the Intoxilyzer, SN: 80-002673, functioned properly at the time of the breath test and hereby certifies the breath test result as valid.

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)
State Director of Tests for Alcohol
and Drug Influence

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 01/23/2017	TIME PROCEDURE BEGAN 10:14	INSTRUMENT LOCATION 80-002673
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-002673

TEST RESULTS

Diagnostics
PASS

Mouth Alcohol Test
PASS

RFI Test
PASS

Abort Test
PASS

Improper Sample Test
PASS

Interferent Detect Test
PASS

Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.018	0.019	0.019
.04	0.037	0.037	0.038

Certificate of Calibration

*This is to certify the calibration of **Intoxilyzer**® serial number 80-002673, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58, No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).*

Date January 4, 2017 Signed [Signature]
Technician



316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev.A

I
th
SIGNATURE
NAME
SMI



3600 N. Martin Luther King Bldg #9
 Oklahoma City, OK 73111
 405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 01 / 23 / 17 Start Time: 1000 End Time: 1430

INSTRUMENT

Intoxilyzer Model 8000 Serial # 80-002673

GAS CANISTER

LOT # 32316080A3 # 028 EXP Date 01-05-2019


REFERENCE

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR3591	DR3753	DR3594	DR3754
Concentration			
.02	.04	.10	.20
LOT #			
15160	15080	15220	15240
Manufactured Date			
07-14-2015	04-01-2015	09-28-2015	11-11-2015
Expiration Date			
07-14-2017	04-01-2017	09-28-2017	11-11-2017
Solution Commission Date			
<u>01-05-17</u>	<u>01-05-17</u>	<u>01-05-17</u>	<u>01-05-17</u>

REASON FOR BENCH CHECK

- POST REPAIR
 CYLINDER REPLACEMENT
 TROUBLESHOOTING
 ROUTINE MAINTENANCE

COMMENTS

 16837
 Technician Signature / Permit #

Josh Smith
 Printed Name

Certificate of Analysis - Dry Gas	
Date of Analysis	01-23-17
Labelled target value (g/210L)	.080
Cylinder #	028
Average test result	.0795
BOT Technician Name and Signature	Josh Smith 16837



BRT Form 5.0
ISO/IEC 17025:2005 Accredited Laboratory

ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 01/23/2017 08:40

Test	g/210L	Time
Air Blank	0.000	08:41
Cal Check	0.079	08:41
Air Blank	0.000	08:42
Cal Check	0.079	08:42
Air Blank	0.000	08:43
Cal Check	0.080	08:43
Air Blank	0.000	08:43
Cal Check	0.079	08:44
Air Blank	0.000	08:44
Cal Check	0.079	08:45
Air Blank	0.000	08:45
Cal Check	0.080	08:45
Air Blank	0.000	08:46
Cal Check	0.080	08:46
Air Blank	0.000	08:47
Cal Check	0.080	08:47
Air Blank	0.000	08:48
Cal Check	0.080	08:48
Air Blank	0.000	08:49
Cal Check	0.079	08:49
Air Blank	0.000	08:50
Cal Check Stats		
Average	0.0795	
Std Dev	0.0005	
Rel Std Dev(%)	0.6629	

Certificate of Analysis

Certificate ID: 9789 ✓
 Part #: BAC105L080T
 Cylinder Size: 105L
 Lot Number: 32316080A3 ✓
 Expiration: 1/5/2019 ✓

REC'D
 01-23-17

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	209 PPM	+/- 0.002 or 2%	NDIR
Nitrogen	BALANCE	BAC whichever is greater	

[Handwritten Signature]
 Operator's Signature

*NIST Standard Reference Material
 Cylinder No. CC103723 / Job No. 09160202
 Certified 212.8 µmol/mol Ethanol in Nitrogen
 for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Handwritten Signature]
 Specialty Gas Lab Tech

12/15/16
 Date



Distributed by:

CMI Inc.
 316 East Ninth Street
 Owensboro, KY 42303
 Phone 866-835-0690
www.alcoholtest.com

CMI INC.

Service Evaluation Form

**This form MUST be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.**

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number **731362** (contact Customer Service)
Name: **Board of Tests** Phone: **(405) 425 2460**
Fax: **(405) 425 2490** Email: **bot@dps.state.ok.us**

2. Bill to Address: **Board of Tests** Ship to Address: **Board of Tests**
 3600 N ML King, Bldg 9 **3600 N ML King, Bldg 9**
 OKC, OK 73111 **OKC, OK 73111**

3. Serial Number: **80-002673** Instrument Model: **Intoxilyzer 500 / 8000**

4. Detailed Description of Problem:
Unit is failing to purge. DVM's are not within tolerance and odd. 3:13332 9:13112

Hazardous Material Warning! - DO NOT return gas cylinders with instrument!

I Authorize Repairs Up To: All \$250 \$500 \$750 Other \$ _____
Purchase Order Number (attach a copy of P.O. if applicable) _____

Authorized By:

Ship to:

Joshua Smith Technician

Name (Please Print) Title

CMI, Inc

 10/03/2016

Attn: Service Dept.

Signature Date

316 East Ninth Street

Owensboro, KY 42303

Call: 405-425-2460 fax 405-425-2490

No, please send estimate before repairs are made.

Note: An estimate will be faxed before performing any repairs and may cause delays in service.
An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.



SERVICE WORK ORDER

DATE	10-10-16
RCVD VIA	LIPS

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

339886

BILL TO:
ATTN:
CITY STATE ZIP
CUSTOMER NO. 731362
PHONE

SHIP TO: JK Board of Tests
ATTN:
3600 N Mill King
Bldg 9
CITY STATE ZIP OKC OK 73111
PHONE 405-425-2460
FAX 405-425-2490

MODEL # 8000	SERIAL # 80-002673	BILL CODE 1
--------------	--------------------	-------------

EXTRA PARTS RCVD
 131, saw shelf, no tanking

DESCRIPTION OF PROBLEM
 Unit is failing to purge. DVM's are not within tolerance.

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION
 Preamp gains need reset.
 Needs recalibration.

BATTERY- EXT N/A V; INT 2.9 V

VERIFIED PROBLEM?
 Yes

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	923002	Diode, SMD, 1N4148
1	021357	Solenoid assy, I8000
1	690040	Battery, 3Volt Li.
1	402565	O-ring, neoprene
1	471201	O-ring, buna-n

WORK PERFORMED:
 Replaced parts listed, cleaned cell extrusion
 Updates completed
 Recalibrated & final tested.

Miscellaneous
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:
 130

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
1111	000	000	000
1/2 HOURS	HOURS	HOURS	HOURS
6	0.5	1.0	1.0
DATE	DATE	DATE	DATE
Dec 29 2016	Jan 3, 17	Jan 4, 17	Jan 4, 17

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 01/23/2017		ARREST TIME 1415		COUNTY # 55		CITY # 13		CITATION # MOCK	
ARREST LOCATION MOCK				CITY MOCK			COUNTY OKLAHOMA		
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 08/09/1961		SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY			STATE OK		ZIP CODE 73133	
DRIVER LICENSE # T083454871			EXPIRATION DATE 06/30/2019		STATE OK	CLASS D	RESTRICTIONS NONE	ENDORSEMENTS NONE	
VEHICLE MAKE MOCK		MODEL MOCK			TAG # MOCK		STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. POST REPAIR, NEW GAS CANISTER INSTALL.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-002673		SPECIALIST BOARD OF TESTS																																			
STANDARD Dry		TARGET VALUE 0.080		PERMIT NUMBER 16837																																			
MANUFACTURER ILMO		LOT # 32316080A3		DATE CYLINDER INSTALLED 01/23/2017	DATE CYLINDER EXPIRES 01/05/2019																																		
OBSERVATION START 1425	TEST DATE 01/23/2017	MOUTHPIECE G	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p>_____ SIGNATURE OF OPERATOR</p>																																				
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>15:06</td></tr> <tr><td>Subject Test</td><td>0.09</td><td>15:06</td></tr> <tr><td>Breath Volume</td><td>1.351 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:07</td></tr> <tr><td>Wait</td><td></td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:09</td></tr> <tr><td>Subject Test</td><td>0.09</td><td>15:10</td></tr> <tr><td>Breath Volume</td><td>1.421 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:10</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>15:11</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:11</td></tr> </tbody> </table>						Test	g/210L	Time	Air Blank	0.00	15:06	Subject Test	0.09	15:06	Breath Volume	1.351 LITERS		Air Blank	0.00	15:07	Wait		15:09	Air Blank	0.00	15:09	Subject Test	0.09	15:10	Breath Volume	1.421 LITERS		Air Blank	0.00	15:10	Cal Check	0.08	15:11	Air Blank
Test	g/210L	Time																																					
Air Blank	0.00	15:06																																					
Subject Test	0.09	15:06																																					
Breath Volume	1.351 LITERS																																						
Air Blank	0.00	15:07																																					
Wait		15:09																																					
Air Blank	0.00	15:09																																					
Subject Test	0.09	15:10																																					
Breath Volume	1.421 LITERS																																						
Air Blank	0.00	15:10																																					
Cal Check	0.08	15:11																																					
Air Blank	0.00	15:11																																					
NAME SMITH, JOSHUA		BADGE # 0000000435	PERMIT # 0000016837																																				
AGENCY BOARD OF TESTS		AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000																																				
		PHONE "4054252460"																																					



Test Result
BrAC g/210L
0.09

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification.** Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: 01/23/2017 Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Date: 01/23/2017 Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

Temporary License Issued

ARRESTEE'S COPY
BOT Form AFF02 07/2008

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 01/23/2017		ARREST TIME 1415		COUNTY # 55		CITY # 13		CITATION # MOCK		
ARREST LOCATION MOCK				CITY MOCK			COUNTY OKLAHOMA			
SUBJECT NAME MOCK, SUBJECT TEST					DATE OF BIRTH 08/09/1961		SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A				CITY OKLAHOMA CITY			STATE OK		ZIP CODE 73133	
DRIVER LICENSE # T083454871				EXPIRATION DATE 06/30/2019		STATE OK	CLASS D	RESTRICTIONS NONE	ENDORSEMENTS NONE	
VEHICLE MAKE MOCK			MODEL MOCK			TAG # MOCK		STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. POST REPAIR, NEW GAS CANISTER INSTALL.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-002673		SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080		PERMIT NUMBER 16837																																					
MANUFACTURER ILMO		LOT # 32316080A3		DATE CYLINDER INSTALLED 01/23/2017	DATE CYLINDER EXPIRES 01/05/2019																																				
OBSERVATION START 1425	TEST DATE 01/23/2017	MOUTHPIECE G	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p>																																						
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>15:06</td></tr> <tr><td>Subject Test</td><td>0.09</td><td>15:06</td></tr> <tr><td>Breath Volume</td><td>1.351 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:07</td></tr> <tr><td>Wait</td><td></td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:09</td></tr> <tr><td>Subject Test</td><td>0.09</td><td>15:10</td></tr> <tr><td>Breath Volume</td><td>1.421 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:10</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>15:11</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:11</td></tr> </tbody> </table>						Test	g/210L	Time	Air Blank	0.00	15:06	Subject Test	0.09	15:06	Breath Volume	1.351 LITERS		Air Blank	0.00	15:07	Wait		15:09	Air Blank	0.00	15:09	Subject Test	0.09	15:10	Breath Volume	1.421 LITERS		Air Blank	0.00	15:10	Cal Check	0.08	15:11	Air Blank	0.00	15:11
Test	g/210L	Time																																							
Air Blank	0.00	15:06																																							
Subject Test	0.09	15:06																																							
Breath Volume	1.351 LITERS																																								
Air Blank	0.00	15:07																																							
Wait		15:09																																							
Air Blank	0.00	15:09																																							
Subject Test	0.09	15:10																																							
Breath Volume	1.421 LITERS																																								
Air Blank	0.00	15:10																																							
Cal Check	0.08	15:11																																							
Air Blank	0.00	15:11																																							
RESULT TABLE			<div style="border: 1px solid black; padding: 5px; display: inline-block;"> <p>Test Result BrAC g/210L 0.09</p> </div>																																						
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td colspan="2">NAME SMITH, JOSHUA</td></tr> <tr><td>BADGE # 0000000435</td><td>PERMIT # 0000016837</td></tr> <tr><td colspan="2">AGENCY BOARD OF TESTS</td></tr> </table>						NAME SMITH, JOSHUA		BADGE # 0000000435	PERMIT # 0000016837	AGENCY BOARD OF TESTS																															
NAME SMITH, JOSHUA																																									
BADGE # 0000000435	PERMIT # 0000016837																																								
AGENCY BOARD OF TESTS																																									

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification.** Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **01/23/2017** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Date: **01/23/2017** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

Temporary License Issued

STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
Receipt for Driver License and Temporary Driving Permit

.02 or Greater BrAC for under 21 yoa, 0.08 or Greater BrAC for 21 yoa and older, or REFUSAL

DATE OF ARREST 01/23/2017		TIME OF ARREST 1415	
NAME (LAST, FIRST, MIDDLE) MOCK, SUBJECT TEST			
ADDRESS 3600 MARTIN LUTHER KING A			
CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73133
DRIVER LICENSE NUMBER T083454871	STATE OK	BIRTHDATE 08/09/1961	SEX M
		RACE W	WEIGHT 165
			HEIGHT 510
CLASS D	EXPIRATION 06/30/2019	RESTRICTIONS NONE	ENDORSEMENTS NONE

BrAC
0.09



DATE ISSUED: 01/23/2017

This form is valid as a driver permit for a period of thirty (30) days from the date it was issued.


Anyone using this form thirty (30) days after issuance may be driving under suspension. All law enforcement officers may determine if the holder is driving under suspension by requesting a Driver's License status through OLETS. This receipt and permit is issued to individuals who have refused to take or have taken a breath test to determine Alcohol Content and the results indicated a Breath Alcohol Content as specified under the Implied Consent Law and Title 47, 11-902, and/or 11-906.4.

VIOLATOR'S REVOCATION NOTICE

Upon receipt of a written blood or breath test report reflecting that the arrested person, if under twenty-one (21) years of age, had any measurable quantity of alcohol in the person's blood or breath, or, if the arrested person is twenty-one (21) years of age or older, a blood or breath alcohol concentration of eight-hundredths (0.08) or more, or refused the State's test, accompanied by a sworn report from a law enforcement officer that the officer had reasonable grounds to believe the arrested person had been operating or was in actual physical control of a motor vehicle while under the influence of alcohol as prohibited by law, the Department shall revoke or deny the driving privilege of the arrested person for a period as provided by Section 6-205.1 of Title 47. Revocation or denial of the driving privilege of the arrested person shall become effective thirty (30) days after the arrested person is given written notice thereof by the officer as provided in this section or by the Department as provided in Section 2-116 of Title 47.

Upon the written request of the person whose driving privilege has been revoked or denied by this notice, the Department of Public Safety shall grant the person an opportunity to be heard if the request is received by the Department within fifteen (15) days after the notice. The sworn report of the officer, together with the results of any test or tests, shall be deemed true, absent any facial deficiency, should the requesting person fail to appear at the scheduled hearing.

ARRESTING OFFICER	
SIGNATURE	BADGE NUMBER 000000435
SMITH, JOSHUA	AGENCY BOARD OF TESTS

Certificate of Analysis - Wet Bath	
12-12-2016	.020
Date of Analysis	Labelled target value (g/210L)
15160	.0194
Lot Number	Average test result (g/210L)
 16837 Joshua Smith	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 ISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

REC'D
 09-24-2015
 #16837

Certified Alcohol Reference Solution for Simulator


ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 12/12/2016 13:25


Random Samples of Lot Number 15160 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 15, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0240% (w/vol) ethyl alcohol. The expiration date for this lot number is July 14, 2017 at 11:59 PM.

Test	g/210L	Time
Air Blank	0.000	13:25
Cal Check	0.019	13:26
Air Blank	0.000	13:26
Cal Check	0.019	13:27
Air Blank	0.000	13:28
Cal Check	0.020	13:28
Air Blank	0.000	13:29
Cal Check	0.019	13:30
Air Blank	0.000	13:30
Cal Check	0.019	13:31
Air Blank	0.000	13:31
Cal Check	0.020	13:32
Air Blank	0.000	13:33
Cal Check	0.019	13:33
Air Blank	0.000	13:34
Cal Check	0.020	13:35
Air Blank	0.000	13:35
Cal Check	0.020	13:36
Air Blank	0.000	13:36
Cal Check	0.019	13:37
Air Blank	0.000	13:38
Cal Check Stats		
Average	0.0194	
Std Dev	0.0005	
Rel Std Dev(%)	2.6617	

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.


 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

 12/12/16

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN070212-01 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



BRT Form 6.0

In accordance with BRT - 2.1.0 (Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

Certificate of Analysis - Wet Bath	
Date of Analysis	12-12-2016
Labelled target value (g/210L)	.040
Lot Number	15080
Average test result (g/210L)	.0396
BRT Technician Name and Signature JOSH SMITH	
Lot Number 16837	



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

REC'D
09-24-2015
J #16837

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 2, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0476%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 1, 2017** at 11:59 PM.

Intoxilyzer - Intoxilyzer
Model 8000 SN 80-002591
12/12/2016 14:18

Test	g/210L	Time
Air Blank	0.000	14:18
Cal Check	0.040	14:19
Air Blank	0.000	14:19
Cal Check	0.040	14:20
Air Blank	0.000	14:21
Cal Check	0.040	14:21
Air Blank	0.000	14:22
Cal Check	0.040	14:22
Air Blank	0.000	14:23
Cal Check	0.040	14:23
Air Blank	0.000	14:23
Cal Check	0.039	14:24
Air Blank	0.000	14:24
Cal Check	0.040	14:25
Air Blank	0.000	14:26
Cal Check	0.039	14:26
Air Blank	0.000	14:27
Cal Check	0.039	14:28
Air Blank	0.000	14:28
Cal Check	0.040	14:29
Air Blank	0.000	14:29
Cal Check	0.039	14:30
Air Blank	0.000	14:31
Cal Check Stats		
Average	0.0396	
Std Dev	0.0005	
Rel Std Dev(%)	1.3040	

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN120110-04** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

J 12-12-16



BRT Form 6.0

Certificate of Analysis - Wet Bath	
12-12-2016	1.100
Date of Analysis	Labelled target value (g/210L)
15220	1.0999
Lot Number	Average test result (g/210L)
JOSEPH SMITH	1.0999
BOT Technician Name and Signature	

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.



GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

RECAL'D
11-4-2015
#16837

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 30, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 28, 2017** at 11:59 PM.

Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
12/12/2016 14:04

Test	g/210L	Time
Air Blank	0.000	14:04
Cal Check	0.100	14:05
Air Blank	0.000	14:05
Cal Check	0.100	14:06
Air Blank	0.000	14:07
Cal Check	0.100	14:07
Air Blank	0.000	14:08
Cal Check	0.100	14:09
Air Blank	0.000	14:09
Cal Check	0.099	14:10
Air Blank	0.000	14:10
Cal Check	0.100	14:11
Air Blank	0.000	14:12
Cal Check	0.100	14:12
Air Blank	0.000	14:13
Cal Check	0.100	14:14
Air Blank	0.000	14:14
Cal Check	0.100	14:15
Air Blank	0.000	14:16
Cal Check	0.100	14:16
Air Blank	0.000	14:17
Cal Check Stats		
Average	0.0999	
Std Dev	0.0003	
Rel Std Dev(%)	0.3165	

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.


12-12-16



BRT Form 6.0

Certificate of Analysis - Wet Bath	
12-12-16	.200
Date of Analysis	Labelled target value (g/210L)
15240	.2006
Lot Number	Average test result (g/210L)
JOE SMITH #16837	
BRT Technician Name and Signature	

In accordance with BRT - 2.1.0 Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

REC'D
11-4-2015
#16837

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15240** of Alcohol Reference Solution for Simulator were analyzed by

gas chromatography on **November 12, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2418%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 11, 2017** at 11:59 PM.

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
12/12/2016 13:42

Test	g/210L	Time
Air Blank	0.000	13:42
Cal Check	0.202	13:43
Air Blank	0.000	13:43
Cal Check	0.201	13:44
Air Blank	0.000	13:45
Cal Check	0.201	13:45
Air Blank	0.000	13:46
Cal Check	0.201	13:47
Air Blank	0.000	13:47
Cal Check	0.200	13:48
Air Blank	0.000	13:49
Cal Check	0.200	13:49
Air Blank	0.000	13:50
Cal Check	0.201	13:51
Air Blank	0.000	13:51
Cal Check	0.200	13:52
Air Blank	0.000	13:52
Cal Check	0.199	13:53
Air Blank	0.000	13:54
Cal Check	0.201	13:54
Air Blank	0.000	13:55
Cal Check Stats		
Average	0.2006	
Std Dev	0.0008	
Rel Std Dev(%)	0.4203	

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

12-12-16

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN03211401 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Kenneth E. Blick, Ph.D.
Chairman

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

Kevin Behrens
State Director

Date: 09/25/2015

As of this date, Intoxilyzer 8000 instrument serial no. 80-002673 has been deployed to Garvin County SO in place of 80-003359.

Joshua C. Smith
Intoxilyzer Technician / Instructor
Board of Tests for Alcohol and Drug Influence
Phone (405) 425-2460
Fax (405) 425-2490

jsmith@dps.state.ok.us