



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman


Joshua Smith  
Director

**ATTESTATION**

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-003187 , in compliance with the agency's Maintenance Policy and Procedures.

This packet consists of 13 pages.

Signed 

Name of Position: Director/Records Custodian

Date of Attestation: 08/11/2020

# Service Overview

Board of Tests for Alcohol and Drug Influence

*Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing.*



3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

**BENCH CHECK DATE:** 08-11-2020

**SERIAL:** 80-003187

## **DOCUMENTATION WITHIN SERVICE PACKET:**

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)



Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)  
State Director of Tests for Alcohol  
and Drug Influence



3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

## **BENCH CHECK TECHNICIAN'S REPORT**

**Date:** 08/11/2020

**Start Time:** 13:10

**End Time:** 14:00

### **INSTRUMENT**

Intoxilyzer Model

Serial #

8000

80-003187

### **GAS CANISTER**

LOT #

EXP Date

14020080A2-052

07/05/2022

### **REFERENCE**

| Simulator Model          |            |            |            |
|--------------------------|------------|------------|------------|
| Guth 2100                | Guth 2100  | Guth 2100  | Guth 2100  |
| Serial #                 |            |            |            |
| DR2834                   | DR3700     | DR2455     | DR3470     |
| Concentration            |            |            |            |
| .02                      | .04        | .10        | .20        |
| LOT #                    |            |            |            |
| 19050                    | 20060      | 20190      | 20160      |
| Manufactured Date        |            |            |            |
| 02/04/2019               | 02/10/2020 | 04/06/2020 | 03/18/2020 |
| Expiration Date          |            |            |            |
| 02/04/2021               | 02/10/2022 | 04/06/2022 | 03/18/2022 |
| Solution Commission Date |            |            |            |
| 8/10/2020                | 8/10/2020  | 8/10/2020  | 8/10/2020  |

### **REASON FOR BENCH CHECK**

- POST REPAIR     CYLINDER REPLACEMENT  
 TROUBLESHOOTING     ROUTINE MAINTENANCE

### **COMMENTS**

|  |
|--|
|  |
|  |
|  |
|  |

*Christopher Pape 23020*

Technician Signature / Permit #

Christopher Pape #23020

Printed Name

**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE  
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

|                                |                                 |                                       |
|--------------------------------|---------------------------------|---------------------------------------|
| BENCH CHECK DATE<br>08/11/2020 | TIME PROCEDURE BEGAN<br>13:10:1 | INSTRUMENT LOCATION<br>PONCA CITY PD  |
| INSTRUMENT TYPE<br>Intoxilyzer | INSTRUMENT MODEL<br>I-8000      | INSTRUMENT SERIAL NUMBER<br>80-003187 |

**TEST RESULTS**

**Diagnostics**

PASS

**Mouth Alcohol Test**

PASS

**RFI Test**

PASS

**Abort Test**

PASS

**Improper Sample Test**

PASS

**Interferent Detect Test**

PASS

**No Sample Given Test**

PASS

**Insufficient Sample Test**

PASS

**Air Blank Contamination Test**

PASS

**Range Exceeded Test**

PASS


**Duplicate Breath Sampling Results**

| Value | Test 1 | Test 2 | Control Test |
|-------|--------|--------|--------------|
| .02   | 0.018  | 0.018  | 0.019        |
| .04   | 0.038  | 0.037  | 0.040        |
| .10   | 0.098  | 0.097  | 0.100        |
| .20   | 0.198  | 0.197  | 0.198        |

PASS

**CERTIFYING TECHNICIAN**

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

|  |                        |
|--|------------------------|
| SIGNATURE<br> 23020 | DATE<br>8/11/2020      |
| NAME<br>PAPE, CHRISTOPHER  | PERMIT #<br>0000023020 |

**Certificate of Analysis - Dry Gas**

|  |                                |
|--|--------------------------------|
| 7/24/2020                                      | 0.080                          |
| Date of Analysis                               | Labelled target value (g/210L) |
| 0.52   | 0.081                          |
| Cylinder #                                     | Average test result            |
| Christopher Pape <i>Christopher Pape</i> 23020 |                                |
| BOT Technician Name and Signature              |                                |



7/13

ALPHA INSTRUMENT 2  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004923  
 07/24/2020 15:57

In accordance with BOT Policy and Procedure Statement BRT-2.2.0, Maintenance, the above referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
 217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 15:57 |
| Cal Check       | 0.080  | 15:57 |
| Air Blank       | 0.000  | 15:58 |
| Cal Check       | 0.081  | 15:58 |
| Air Blank       | 0.000  | 15:59 |
| Cal Check       | 0.081  | 15:59 |
| Air Blank       | 0.000  | 16:00 |
| Cal Check       | 0.080  | 16:00 |
| Air Blank       | 0.000  | 16:01 |
| Cal Check       | 0.081  | 16:01 |
| Air Blank       | 0.000  | 16:02 |
| Cal Check       | 0.080  | 16:02 |
| Air Blank       | 0.000  | 16:02 |
| Cal Check       | 0.081  | 16:02 |
| Air Blank       | 0.000  | 16:03 |
| Cal Check       | 0.081  | 16:03 |
| Air Blank       | 0.000  | 16:04 |
| Cal Check       | 0.081  | 16:04 |
| Air Blank       | 0.000  | 16:05 |
| Cal Check       | 0.080  | 16:05 |
| Air Blank       | 0.000  | 16:05 |
| Cal Check Stats |        |       |
| Average         | 0.0806 |       |
| Std Dev         | 0.0005 |       |
| Rel Std Dev(%)  | 0.6406 |       |

**Certificate of Analysis**

*RCV'D*  
*7/24/2020*  
*CP*

Certificate ID: 13021 ✓  
 Part #: BAC105L080T  
 Cylinder Size: 105L  
 Lot Number: 14020080A2 ✓  
 Expiration: 7/5/2022 ✓

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

| Component: | Reported Concentration: | Analytical Accuracy (U, k=2): | Analytical Method: |
|------------|-------------------------|-------------------------------|--------------------|
| Ethanol    | 208 ppm                 | +/-0.002 BAC(G/210L)          | NDIR               |
| Nitrogen   | Balance                 | [5.2 ppm]                     |                    |

**Distributed by:**

CMI Inc.  
 316 East Ninth Street  
 Owensboro, KY 4231  
 Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

*Christopher Pape*  
 Operator's Signature

\*Traceable to:  
 Certified Reference Material - 262.4 µmol/mol  
 Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

*[Signature]*  
 Specialty Gas Lab Tech

*[Signature]*

06-15-2020  
 Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

8/13

Section 1

|  |  |                            |                                      |                                    |                    |                           |                  |                          |                      |
|--|--|----------------------------|--------------------------------------|------------------------------------|--------------------|---------------------------|------------------|--------------------------|----------------------|
| ARREST DATE<br><b>08/11/2020</b>             |  | ARREST TIME<br><b>1315</b> |                                      | COUNTY #<br><b>55</b>              |                    | CITY #                    |                  | CITATION #               |                      |
| ARREST LOCATION<br><b>BOT</b>                |  |                            |                                      | CITY                               |                    | COUNTY<br><b>OKLAHOMA</b> |                  |                          |                      |
| SUBJECT NAME<br><b>NORTH, GAYLON LEVERNE</b> |  |                            |                                      | DATE OF BIRTH<br><b>08/09/1961</b> |                    | SEX<br><b>M</b>           | RACE<br><b>W</b> | HEIGHT<br><b>510</b>     | WEIGHT<br><b>165</b> |
| ADDRESS<br><b>3600 MARTIN LUTHER KING A</b>  |  |                            | CITY<br><b>OKLAHOMA CITY</b>         |                                    |                    | STATE<br><b>OK</b>        |                  | ZIP CODE<br><b>73133</b> |                      |
| DRIVER LICENSE #<br><b>T083454871</b>        |  |                            | EXPIRATION DATE<br><b>06/30/2019</b> |                                    | STATE<br><b>OK</b> | CLASS<br><b>A</b>         | RESTRICTIONS     | ENDORSEMENTS<br><b>M</b> |                      |
| VEHICLE MAKE<br><b>MOCK</b>                  |  | MODEL<br><b>MOCK</b>       |                                      | TAG #<br><b>MOCK</b>               |                    | STATE<br><b>OK</b>        | CMV<br><b>N</b>  | PLACARD<br><b>N</b>      |                      |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.  
(Describe driving behavior or circumstances):



Describe person's condition (odor, actions, etc.):

**MOCK SUBJECT TEST, ROUTINE MAINTENANCE, INSTALL NEW CYLINDER**

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

|                                  |                                |                                   |   |  |  |
|----------------------------------|--------------------------------|-----------------------------------|---|--|--|
| EQUIPMENT<br><b>I-8000</b>       |                                | SERIAL NUMBER<br><b>80-003187</b> |   | SPECIALIST<br><b>BOARD OF TESTS</b>          |  |
| STANDARD<br><b>Dry</b>           |                                | TARGET VALUE<br><b>0.080</b>      |   | PERMIT NUMBER<br><b>23020</b>                |  |
| MANUFACTURER<br><b>ILMO</b>      |                                | LOT #<br><b>14020080A2</b>        |   | DATE CYLINDER INSTALLED<br><b>08/11/2020</b> | DATE CYLINDER EXPIRES<br><b>07/05/2022</b> |
| OBSERVATION START<br><b>1330</b> | TEST DATE<br><b>08/11/2020</b> | MOUTHPIECE<br><b>G</b>            | A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.  |  |  |
| RESULT TABLE                     |                                |                                   | <br>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.<br><br>SIGNATURE OF OPERATOR |  |  |
| Test                             | g/210L                         | Time                              |   |  |  |
| Air Blank                        | 0.00                           | 13:53                             |   |  |  |
| Subject Test                     | 0.00                           | 13:53                             |   |  |  |
| Breath Volume                    | 1.441 LITERS                   |                                   |   |  |  |
| Air Blank                        | 0.00                           | 13:53                             |   |  |  |
| Wait                             |                                | 13:55                             |   |  |  |
| Air Blank                        | 0.00                           | 13:56                             |   |  |  |
| Subject Test                     | 0.00                           | 13:56                             |   |  |  |
| Breath Volume                    | 1.652 LITERS                   |                                   |   |  |  |
| Air Blank                        | 0.00                           | 13:56                             |   |  |  |
| Cal Check                        | 0.07                           | 13:57                             |   |  |  |
| Air Blank                        | 0.00                           | 13:57                             |   |  |  |
| NAME<br><b>PAPE, CHRISTOPHER</b> |                                |                                   | Test Result<br>BrAC g/210L<br><b>0.00</b>   |  |  |
| BADGE #<br><b>0000000484</b>     |                                | PERMIT #<br><b>0000023020</b>     |   |  |  |
| AGENCY<br><b>BOARD OF TESTS</b>  |                                |                                   |   |  |  |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

Section 3

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **08/11/2020** Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct." Section 4

Date: **08/11/2020** Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

|                                  |                                       |                                 |                          |                            |  |
|----------------------------------|---------------------------------------|---------------------------------|--------------------------|----------------------------|--|
| NAME<br><b>PAPE, CHRISTOPHER</b> |                                       | AGENCY<br><b>BOARD OF TESTS</b> |                          |                            |  |
| BADGE<br><b>0000000484</b>       | AGENCY ADDRESS<br><b>INTER-AGENCY</b> |                                 | ZIP CODE<br><b>00000</b> | PHONE<br><b>4054252460</b> |  |

OTHER WITNESSES:

Section 5

|   |      |       |         |       |
|---|------|-------|---------|-------|
| 1 | NAME | TITLE | ADDRESS | PHONE |
| 2 | NAME | TITLE | ADDRESS | PHONE |

No Temporary License Issued:  
Test Below Legal Limit

OFFICER/DPS COPY  
BOT Form AFF01 03/2020



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

9/13

Section 1

|  |  |                            |                              |                                      |        |                           |                   |                          |                          |                     |
|--|--|----------------------------|------------------------------|--------------------------------------|--------|---------------------------|-------------------|--------------------------|--------------------------|---------------------|
| ARREST DATE<br><b>08/11/2020</b>             |  | ARREST TIME<br><b>1315</b> |                              | COUNTY #<br><b>55</b>                | CITY # | CITATION #                |                   |                          |                          |                     |
| ARREST LOCATION<br><b>BOT</b>                |  |                            |                              | CITY                                 |        | COUNTY<br><b>OKLAHOMA</b> |                   |                          |                          |                     |
| SUBJECT NAME<br><b>NORTH, GAYLON LEVERNE</b> |  |                            |                              | DATE OF BIRTH<br><b>08/09/1961</b>   |        | SEX<br><b>M</b>           | RACE<br><b>W</b>  | HEIGHT<br><b>510</b>     | WEIGHT<br><b>165</b>     |                     |
| ADDRESS<br><b>3600 MARTIN LUTHER KING A</b>  |  |                            | CITY<br><b>OKLAHOMA CITY</b> |                                      |        | STATE<br><b>OK</b>        |                   | ZIP CODE<br><b>73133</b> |                          |                     |
| DRIVER LICENSE #<br><b>T083454871</b>        |  |                            |                              | EXPIRATION DATE<br><b>06/30/2019</b> |        | STATE<br><b>OK</b>        | CLASS<br><b>A</b> | RESTRICTIONS             | ENDORSEMENTS<br><b>M</b> |                     |
| VEHICLE MAKE<br><b>MOCK</b>                  |  |                            | MODEL<br><b>MOCK</b>         |                                      |        | TAG #<br><b>MOCK</b>      |                   | STATE<br><b>OK</b>       | CMV<br><b>N</b>          | PLACARD<br><b>N</b> |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):


Describe person's condition (odor, actions, etc.):

**MOCK SUBJECT TEST, ROUTINE MAINTENANCE, INSTALL NEW CYLINDER**

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

|                                  |                                |                                   |   |  |  |
|----------------------------------|--------------------------------|-----------------------------------|---|--|--|
| EQUIPMENT<br><b>I-8000</b>       |                                | SERIAL NUMBER<br><b>80-003187</b> | SPECIALIST<br><b>BOARD OF TESTS</b>   |  |  |
| STANDARD<br><b>Dry</b>           |                                | TARGET VALUE<br><b>0.080</b>      | PERMIT NUMBER<br><b>23020</b>   |  |  |
| MANUFACTURER<br><b>ILMO</b>      |                                | LOT #<br><b>14020080A2</b>        | DATE CYLINDER INSTALLED<br><b>08/11/2020</b>  | DATE CYLINDER EXPIRES<br><b>07/05/2022</b> |  |
| OBSERVATION START<br><b>1330</b> | TEST DATE<br><b>08/11/2020</b> | MOUTHPIECE<br><b>G</b>            | <p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>Christopher Pape</i> 2020<br/>SIGNATURE OF OPERATOR</p> |  |  |
| RESULT TABLE                     |                                |                                   |   |  |  |
| Test                             | g/210L                         | Time                              |   |  |  |
| Air Blank                        | 0.00                           | 13:53                             |   |  |  |
| Subject Test                     | 0.00                           | 13:53                             |   |  |  |
| Breath Volume                    | 1.441 LITERS                   |                                   |   |  |  |
| Air Blank                        | 0.00                           | 13:53                             |   |  |  |
| Wait                             |                                | 13:55                             |   |  |  |
| Air Blank                        | 0.00                           | 13:56                             |   |  |  |
| Subject Test                     | 0.00                           | 13:56                             |   |  |  |
| Breath Volume                    | 1.652 LITERS                   |                                   |   |  |  |
| Air Blank                        | 0.00                           | 13:56                             |   |  |  |
| Cal Check                        | 0.07                           | 13:57                             |   |  |  |
| Air Blank                        | 0.00                           | 13:57                             |   |  |  |
| NAME<br><b>PAPE, CHRISTOPHER</b> |                                |                                   | <div style="border: 1px solid black; padding: 5px;"> <p>Test Result<br/>BrAC g/210L<br/><b>0.00</b></p> </div>  |  |  |
| BADGE #<br><b>0000000484</b>     |                                | PERMIT #<br><b>0000023020</b>     |   |  |  |
| AGENCY<br><b>BOARD OF TESTS</b>  |                                |                                   |   |  |  |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

Section 3

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **08/11/2020** Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **08/11/2020** Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

|                                  |                                       |                                 |                            |  |
|----------------------------------|---------------------------------------|---------------------------------|----------------------------|--|
| NAME<br><b>PAPE, CHRISTOPHER</b> |                                       | AGENCY<br><b>BOARD OF TESTS</b> |                            |  |
| BADGE<br><b>0000000484</b>       | AGENCY ADDRESS<br><b>INTER-AGENCY</b> | ZIP CODE<br><b>00000</b>        | PHONE<br><b>4054252460</b> |  |

No Temporary License Issued:  
Test Below Legal Limit

ARRESTEE'S COPY  
BOT Form AFF02 03/2020

Certificate of Analysis - Wet Bath



10/13

|                  |                               |
|------------------|-------------------------------|
| 12-10-19         | 0.020                         |
| Date of Analysis | Labeled target value (g/210L) |
| 19050            | 0.020                         |
| Lot Number       | Average Result (g/210L)       |

JOSH SMITH

BOT Technician Name and Signature

GUTH LABORATORIES, INC.  
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

Rec'd  
12/2019  
[Signature]

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
12/10/2019

SN 80-004923  
14:52

Certified Alcohol Reference Solution for Simulator

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.009  | 14:53 |
| Cal Check       | 0.019  | 14:53 |
| Air Blank       | 0.000  | 14:54 |
| Cal Check       | 0.020  | 14:55 |
| Air Blank       | 0.000  | 14:55 |
| Cal Check       | 0.019  | 14:56 |
| Air Blank       | 0.000  | 14:56 |
| Cal Check       | 0.020  | 14:57 |
| Air Blank       | 0.000  | 14:58 |
| Cal Check       | 0.019  | 14:58 |
| Air Blank       | 0.000  | 14:59 |
| Cal Check       | 0.020  | 15:00 |
| Air Blank       | 0.000  | 15:00 |
| Cal Check       | 0.019  | 15:01 |
| Air Blank       | 0.000  | 15:01 |
| Cal Check       | 0.020  | 15:02 |
| Air Blank       | 0.000  | 15:02 |
| Cal Check       | 0.020  | 15:03 |
| Air Blank       | 0.000  | 15:04 |
| Cal Check       | 0.020  | 15:04 |
| Air Blank       | 0.000  | 15:05 |
| Cal Check Stats |        |       |
| Average         | 0.0196 |       |
| Std Dev         | 0.0005 |       |
| Rel Std Dev(%)  | 2.6345 |       |

Random Samples of Lot Number 19050 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on February 5, 2019, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0236% (w/vol) ethyl alcohol. The expiration date for this lot number is February 4, 2021 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

[Signature]  
16837  
Operator's Signature


[Signature]  
Ted L. Pauley, President  
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08031603 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**Certificate of Analysis – Wet Bath**

|  |                                |
|--|--------------------------------|
| 7/22/2020  | 0.040                          |
| Date of Analysis   | Labelled target value (g/210L) |
| 20060  | 0.040                          |
| Lot Number   | Average test result (g/210L)   |
|  |                                |
| BOT Technician Name and Signature  |                                |



11/13

**GUTH LABORATORIES, INC.**  
 530 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

RCV'D  
6/1/20  
CP


Random Samples of Lot Number **20060** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 11, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0483%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 10, 2022** at 11:59 PM.

IR UNIT SUB  
 Intoxilyzer - Alcohol Analyzer  
 #1 8000 SN 80-003400  
 12/2020 14:57

|             | g/210L | Time  |
|-------------|--------|-------|
| Blank       | 0.000  | 14:57 |
| Check       | 0.040  | 14:58 |
| Blank       | 0.000  | 14:59 |
| Check       | 0.039  | 14:59 |
| Blank       | 0.000  | 15:00 |
| Check       | 0.039  | 15:01 |
| Blank       | 0.000  | 15:01 |
| Check       | 0.039  | 15:02 |
| Blank       | 0.000  | 15:02 |
| Check       | 0.039  | 15:03 |
| Blank       | 0.000  | 15:04 |
| Check       | 0.039  | 15:04 |
| Blank       | 0.000  | 15:05 |
| Check       | 0.040  | 15:06 |
| Blank       | 0.000  | 15:06 |
| Check       | 0.040  | 15:07 |
| Blank       | 0.000  | 15:07 |
| Check       | 0.040  | 15:08 |
| Blank       | 0.000  | 15:09 |
| Check       | 0.040  | 15:09 |
| Blank       | 0.000  | 15:10 |
| Check Stats |        |       |
| Age         | 0.0395 |       |
| Dev         | 0.0005 |       |
| Std Dev(%)  | 1.3342 |       |

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.



Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

  
 Operator's Signature

**NIST Traceability:**  
 Testing was conducted using Cerilliant Reference Standard lot number **FN12181501** whose values are traceable to NIST.  
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath



12/13

|                                   |                                |
|-----------------------------------|--------------------------------|
| 06/16/20                          | .100                           |
| Date of Analysis                  | Labelled target value (g/210L) |
| 20190                             | .098                           |
| Lot Number                        | Average test result (g/210L)   |
| Josh Smith                        |                                |
| BOT Technician Name and Signature |                                |

GUTH LABORATORIES, INC.

BRT Form 6.0 RISC JRG, PA 17111-4511 • TELEPHONE: 717-864-5470

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

Rec'd  
06/11/2020

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT  
Intoxilyzer – Alcohol Analyzer  
Model 8000 SN 80-002591  
06/16/2020 16:18

Certified Alcohol Reference Solution for Simulator


| Test            | g/210L | Time  |
|-----------------|--------|-------|
| Air Blank       | 0.000  | 16:18 |
| Cal Check       | 0.099  | 16:19 |
| Air Blank       | 0.000  | 16:20 |
| Cal Check       | 0.098  | 16:20 |
| Air Blank       | 0.000  | 16:21 |
| Cal Check       | 0.097  | 16:22 |
| Air Blank       | 0.000  | 16:22 |
| Cal Check       | 0.098  | 16:23 |
| Air Blank       | 0.000  | 16:24 |
| Cal Check       | 0.098  | 16:24 |
| Air Blank       | 0.000  | 16:25 |
| Cal Check       | 0.098  | 16:26 |
| Air Blank       | 0.000  | 16:26 |
| Cal Check       | 0.098  | 16:27 |
| Air Blank       | 0.000  | 16:27 |
| Cal Check       | 0.097  | 16:28 |
| Air Blank       | 0.000  | 16:29 |
| Cal Check       | 0.098  | 16:29 |
| Air Blank       | 0.000  | 16:30 |
| Cal Check       | 0.098  | 16:31 |
| Air Blank       | 0.000  | 16:31 |
| Cal Check Stats |        |       |
| Average         | 0.0979 |       |
| Std Dev         | 0.0005 |       |
| Rel Std Dev(%)  | 0.5798 |       |

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

  
Operator's Signature 16837

  
Ted L. Pauley, President  
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

| Certificate of Analysis - Wet Bath |                               |
|------------------------------------|-------------------------------|
| 07/07/2020                         | 0.200                         |
| Date of Analysis                   | Labeled Weight value (g/210L) |
| 20160                              | 0.197                         |
| Lot Number                         | Average test result (g/210L)  |
| Christopher Pope                   | 0.23030                       |
| BOT Technician Name and Signature  |                               |



**GUTH LABORATORIES, INC.**  
 650 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

### CERTIFICATE OF ANALYSIS

*Rec'd  
05/20/2020  
CP*

Certified Alcohol Reference Solution for Simulator

PHR UNIT SUB  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SH 80-003400  
 7/07/2020 15:25

| Test            | g/210L | Time  |
|-----------------|--------|-------|
| ir Blank        | 0.000  | 15:26 |
| al Check        | 0.196  | 15:26 |
| ir Blank        | 0.000  | 15:27 |
| al Check        | 0.196  | 15:28 |
| ir Blank        | 0.000  | 15:28 |
| al Check        | 0.197  | 15:29 |
| ir Blank        | 0.000  | 15:29 |
| al Check        | 0.196  | 15:30 |
| ir Blank        | 0.000  | 15:31 |
| al Check        | 0.198  | 15:31 |
| ir Blank        | 0.000  | 15:32 |
| al Check        | 0.198  | 15:33 |
| ir Blank        | 0.000  | 15:33 |
| al Check        | 0.197  | 15:34 |
| ir Blank        | 0.000  | 15:35 |
| al Check        | 0.198  | 15:35 |
| ir Blank        | 0.000  | 15:36 |
| al Check        | 0.198  | 15:37 |
| ir Blank        | 0.000  | 15:37 |
| al Check        | 0.197  | 15:38 |
| ir Blank        | 0.000  | 15:38 |
| al Check Stats  |        |       |
| Average         | 0.1971 |       |
| Std Dev         | 0.0008 |       |
| Rel. Std Dev(%) | 0.4442 |       |

Random Samples of Lot Number **20160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 20, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2427% (w/vol) ethyl alcohol**. The expiration date for this lot number is **March 18, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

*Christopher Pope*  
 Operator's Signature

*Ted L. Pauley*  
 Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN08101505 whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.