



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT
Chairman

Joshua Smith
Director

ATTESTATION

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-003300, in compliance with the agency's Maintenance Policy and Procedures.

This packet consists of 12 pages.

Signed _____

Name of Position: Director/Records Custodian

Date of Attestation: 09/24/2021

Supplemented 02/17/2022

Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 09-24-2021

SERIAL: 80-003300

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D., F-ABFT
Chairman

Joshua Smith
Director

MEMORANDUM

DATE: February 11, 2022

TO: FILE

A scrivener's error was identified on BOT Form 4.0 Certificate of Maintenance, Calibration, and Operation as a result of multiple sessions of drafting. The form contained in this maintenance packet has been amended and updated to include the intended information. The agency apologizes for any inconvenience this may have caused.

A handwritten signature in blue ink, appearing to be "J. Smith", is written over the printed name and title.

Joshua Smith
State Director

Cc: Intoxilyzer 8000 maintenance folder



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D.,
F-ABFT
Chairman

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

Joshua Smith
Director

CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION


This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-003300 on 09/24/2021 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
 - Manufacturer: ILMO Unit: 208 ppm / 105 L
 - Cylinder Lot No. - Canister No.: 19021080A2-013
 - BOT analysis average test result: 0.082
 - Cylinder expiration date: 09/05/2023
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model					
Guth 2100		Guth 2100		Guth 2100	
Serial #					
DR3685		DR3753		DR3594	
Concentration					
.02		.04		.10	
LOT #					
20070		20060		20190	
Manufactured Date					
02/13/2020		02/10/2020		04/06/2020	
Expiration Date					
02/13/2022		02/10/2022		04/06/2022	
Solution Commission Date					
09/03/2021		09/03/2021		09/23/2021	

Completed by BOT Personnel:

 christopher Pope
Signature/Name

Breath-Alcohol Testing Program Administrator
Title

Certified by:

 _____
Signature Name

State Director
Title

**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 09/24/2021	TIME PROCEDURE BEGAN 10:11:3	INSTRUMENT LOCATION 80-003300
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-003300

TEST RESULTS

Diagnostics

PASS

Mouth Alcohol Test

PASS

RFI Test

PASS

Abort Test

PASS

Improper Sample Test

PASS

Interferent Detect Test

PASS

No Sample Given Test

PASS

Insufficient Sample Test

PASS

Air Blank Contamination Test

PASS

Range Exceeded Test

PASS

Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.018	0.018	0.019
.04	0.037	0.037	0.037
.10	0.100	0.098	0.099
.20	0.201	0.199	0.200

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 9/24/2021
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020

Certificate of Analysis - Dry Gas

9/23/2021	0.080
Date of Analysis	Labelled target value (g/210L)
013	0.082
Cylinder #	Average test result
Christopher Pope <i>[Signature]</i>	
BOT Technician Name and Signature	



6/12

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
09/23/2021 06:36

In accordance with BOT Maintenance Policy and Procedure Statement, the above referenced dry gas cylinder is suitable for use as an external reference with the following information:
7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.co

Test	g/210L	Time
Air Blank	0.000	08:36
Cal Check	0.081	08:36
Air Blank	0.000	08:37
Cal Check	0.082	08:37
Air Blank	0.000	08:38
Cal Check	0.082	08:38
Air Blank	0.000	08:39
Cal Check	0.082	08:39
Air Blank	0.000	08:40
Cal Check	0.082	08:40
Air Blank	0.000	08:40
Cal Check	0.082	08:41
Air Blank	0.000	08:41
Cal Check	0.082	08:42
Air Blank	0.000	08:42
Cal Check	0.082	08:42
Air Blank	0.000	08:43
Cal Check	0.082	08:43
Air Blank	0.000	08:44
Cal Check	0.082	08:44
Air Blank	0.000	08:45
Cal Check Stats		
Average	0.0819	
Std Dev	0.0003	
Rel Std Dev(%)	0.3661	

Certificate of Analysis

Certificate ID: 14091 ✓
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 19021080A2 ✓
Expiration: 9/5/2023 ✓

*RCVD
9/16/21
CP*

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by
CMI Inc.
316 East Ninth
Owensboro, KY
Phone 866-83
www.alcoholt.com

[Signature]
Operator's Signature

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech
[Signature]
Troy Nelson

08-12-2021
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

7/12

Section 1

ARREST DATE 09/24/2021		ARREST TIME 1100		COUNTY # 55	CITY #	CITATION #		
ARREST LOCATION BOT				CITY		COUNTY OKLAHOMA		
SUBJECT NAME NORTH, GAYLON LEVERNE				DATE OF BIRTH 08/09/1961	SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73133		
DRIVER LICENSE # T083454871			EXPIRATION DATE 06/30/2019	STATE OK	CLASS A	RESTRICTIONS	ENDORSEMENTS M	
VEHICLE MAKE MOCK		MODEL MOCK		TAG # MOCK		STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST

Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-003300	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO		LOT # 19021080A2	DATE CYLINDER INSTALLED 09/24/2021	DATE CYLINDER EXPIRES 09/05/2023																																				
OBSERVATION START 1110	TEST DATE 09/24/2021	MOUTHPIECE G	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>Christopher Pape</i> SIGNATURE OF OPERATOR</p>																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>11:28</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>11:28</td></tr> <tr><td> Breath Volume</td><td>2.085 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:29</td></tr> <tr><td>Wait</td><td></td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:31</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>11:31</td></tr> <tr><td> Breath Volume</td><td>2.066 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:32</td></tr> <tr><td>Cal Check</td><td>0.07</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:33</td></tr> </tbody> </table>					Test	g/210L	Time	Air Blank	0.00	11:28	Subject Test	0.00	11:28	Breath Volume	2.085 LITERS		Air Blank	0.00	11:29	Wait		11:31	Air Blank	0.00	11:31	Subject Test	0.00	11:31	Breath Volume	2.066 LITERS		Air Blank	0.00	11:32	Cal Check	0.07	11:32	Air Blank	0.00	11:33
Test	g/210L	Time																																						
Air Blank	0.00	11:28																																						
Subject Test	0.00	11:28																																						
Breath Volume	2.085 LITERS																																							
Air Blank	0.00	11:29																																						
Wait		11:31																																						
Air Blank	0.00	11:31																																						
Subject Test	0.00	11:31																																						
Breath Volume	2.066 LITERS																																							
Air Blank	0.00	11:32																																						
Cal Check	0.07	11:32																																						
Air Blank	0.00	11:33																																						
RESULT TABLE			NAME PAPE, CHRISTOPHER																																					
			BADGE # 0000000484	PERMIT # 0000023020																																				
			AGENCY BOARD OF TESTS																																					

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

Section 3

SERVING OF NOTICE: I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **09/24/2021** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **09/24/2021** Place (location when signed): _____ Signature of arresting officer: _____

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

No Temporary License Issued:
Test Below Legal Limit

OFFICER/DPS COPY
BOT Form AFF01 03/2020

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

8/12

Section 1

ARREST DATE 09/24/2021		ARREST TIME 1100		COUNTY # 55	CITY #	CITATION #		
ARREST LOCATION BOT				CITY		COUNTY OKLAHOMA		
SUBJECT NAME NORTH, GAYLON LEVERNE				DATE OF BIRTH 08/09/1961	SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73133		
DRIVER LICENSE # T083454871			EXPIRATION DATE 06/30/2019	STATE OK	CLASS A	RESTRICTIONS	ENDORSEMENTS M	
VEHICLE MAKE MOCK		MODEL MOCK		TAG # MOCK	STATE OK	CMV N	PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST

Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-003300	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO		LOT # 19021080A2	DATE CYLINDER INSTALLED 09/24/2021	DATE CYLINDER EXPIRES 09/05/2023																																				
OBSERVATION START 1110	TEST DATE 09/24/2021	MOUTHPIECE G	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>Christopher Pape</i> SIGNATURE OF OPERATOR</p>																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>11:28</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>11:28</td></tr> <tr><td> Breath Volume</td><td>2.085 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:29</td></tr> <tr><td>Wait</td><td></td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:31</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>11:31</td></tr> <tr><td> Breath Volume</td><td>2.066 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:32</td></tr> <tr><td>Cal Check</td><td>0.07</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>11:33</td></tr> </tbody> </table>					Test	g/210L	Time	Air Blank	0.00	11:28	Subject Test	0.00	11:28	Breath Volume	2.085 LITERS		Air Blank	0.00	11:29	Wait		11:31	Air Blank	0.00	11:31	Subject Test	0.00	11:31	Breath Volume	2.066 LITERS		Air Blank	0.00	11:32	Cal Check	0.07	11:32	Air Blank	0.00	11:33
Test	g/210L	Time																																						
Air Blank	0.00	11:28																																						
Subject Test	0.00	11:28																																						
Breath Volume	2.085 LITERS																																							
Air Blank	0.00	11:29																																						
Wait		11:31																																						
Air Blank	0.00	11:31																																						
Subject Test	0.00	11:31																																						
Breath Volume	2.066 LITERS																																							
Air Blank	0.00	11:32																																						
Cal Check	0.07	11:32																																						
Air Blank	0.00	11:33																																						
NAME PAPE, CHRISTOPHER			BADGE # 0000000484																																					
AGENCY BOARD OF TESTS			PERMIT # 0000023020																																					

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below.

Section 3

You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

SERVING OF NOTICE: I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **09/24/2021** Signature of officer: _____

Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **09/24/2021** Place (location when signed): _____

Signature of arresting officer: _____

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

No Temporary License Issued:
Test Below Legal Limit

ARRESTEE'S COPY
BOT Form AFFD2 03/2020

Certificate of Analysis – Wet Bath	
8/11/2020	0.020
Date of Analysis	Labelled tank value (g/210L)
20070	0.020
Lot Number	Average result (g/210L)
christopher kape	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 150 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

ACV'D
 6/1/2020
 CP

ALPHA INSTRUMENT 2
 Intoxilyzer – Alcohol Analyzer
 Model 8000 SN 00-004923
 08/11/2020 09:42

Test	g/210L	Time
Air Blank	0.000	09:42
Cal Check	0.020	09:43
Air Blank	0.000	09:43
Cal Check	0.020	09:44
Air Blank	0.000	09:45
Cal Check	0.020	09:45
Air Blank	0.000	09:46
Cal Check	0.020	09:47
Air Blank	0.000	09:47
Cal Check	0.021	09:48
Air Blank	0.000	09:48
Cal Check	0.020	09:49
Air Blank	0.000	09:49
Cal Check	0.020	09:50
Air Blank	0.000	09:51
Cal Check	0.020	09:51
Air Blank	0.000	09:52
Cal Check	0.020	09:53
Air Blank	0.000	09:53
Cal Check	0.019	09:54
Air Blank	0.000	09:54
Cal Check Stats		
Average	0.0200	
Std Dev	0.0004	
Rel Std Dev(%)	2.3569	

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 17, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0242%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 13, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.020 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Christopher Kape
 Operator's Signature

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number **FN08031603** whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath	
7/22/2020	0.040
Date of Analysis	Labelled target value (g/210L)
20060	0.040
Lot Number	Average test result (g/210L)
Christopher Pauley	0.0483
BRT Technician Name and Signature	



GUTH LABORATORIES, INC.
 100 NORTH 500 STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

RCV'D
6/1/20
CP

Random Samples of Lot Number **20060** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 11, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0483%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 10, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

ALPHA WIT 508
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 00-003400
 07/22/2020 14:57

Test	g/210L	Time
Air Blank	0.000	14:57
Cal Check	0.040	14:58
Air Blank	0.000	14:59
Cal Check	0.039	14:59
Air Blank	0.000	15:00
Cal Check	0.039	15:01
Air Blank	0.000	15:01
Cal Check	0.039	15:02
Air Blank	0.000	15:02
Cal Check	0.039	15:03
Air Blank	0.000	15:04
Cal Check	0.039	15:04
Air Blank	0.000	15:05
Cal Check	0.040	15:06
Air Blank	0.000	15:06
Cal Check	0.040	15:07
Air Blank	0.000	15:07
Cal Check	0.040	15:08
Air Blank	0.000	15:09
Cal Check	0.040	15:09
Air Blank	0.000	15:10
Cal Check Stats		
Average	0.0395	
Std Dev	0.0005	
Rel Std Dev(%)	1.3342	

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

Christopher Pauley
 Operator's Signature

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number **FN12181501** whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath

06/16/20	.100
Date of Analysis	Labelled target value (g/210L)
20190	.098
Lot Number	As found test result (g/210L)
John Smith	
BOT Technician Name and Signature	



11/12

GUTH LABORATORIES, INC.

1001 N. 11th St., Erie, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

Rec'd
06/11/2020
[Signature]

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000
06/16/2020

SN 80-002591
16:18

Certified Alcohol Reference Solution for Simulator

Test	g/210L	Time
Air Blank	0.000	16:18
Cal Check	0.099	16:19
Air Blank	0.000	16:20
Cal Check	0.098	16:20
Air Blank	0.000	16:21
Cal Check	0.097	16:22
Air Blank	0.000	16:22
Cal Check	0.098	16:23
Air Blank	0.000	16:24
Cal Check	0.098	16:24
Air Blank	0.000	16:25
Cal Check	0.098	16:26
Air Blank	0.000	16:26
Cal Check	0.098	16:27
Air Blank	0.000	16:27
Cal Check	0.097	16:28
Air Blank	0.000	16:29
Cal Check	0.098	16:29
Air Blank	0.000	16:30
Cal Check	0.098	16:31
Air Blank	0.000	16:31
Cal Check Stats		
Average	0.0979	
Std Dev	0.0005	
Rel Std Dev(%)	0.5798	

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

[Signature]
Operator's Signature 16037

[Signature]
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath	
07/07/2020	0.200
Date of Analysis	Labeled (g/dl) value (g/210L)
20160	0.197
Lot Number	Average test result (g/210L)
Christopher Pauley	0.200
BOY Technician Name and Signature	



GUTH LABORATORIES, INC.
 650 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

BRT Form 6.0

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

*Rec'd
05/2020
CP*

Certified Alcohol Reference Solution for Simulator

LPHR UNIT SUB
 Intoxilyzer – Alcohol Analyzer
 Model 8000 SI 00-003400
 7/07/2020 15:25

Random Samples of Lot Number **20160** of

Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 20, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2427% (w/vol) ethyl alcohol. The expiration date for this lot number is March 18, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Test	g/210L	Time
Ir Blank	0.000	15:26
al Check	0.196	15:26
Ir Blank	0.000	15:27
al Check	0.196	15:28
Ir Blank	0.000	15:28
al Check	0.197	15:29
Ir Blank	0.000	15:29
al Check	0.196	15:30
Ir Blank	0.000	15:31
al Check	0.198	15:31
Ir Blank	0.000	15:32
al Check	0.198	15:33
Ir Blank	0.000	15:33
al Check	0.197	15:34
Ir Blank	0.000	15:35
al Check	0.198	15:35
Ir Blank	0.000	15:36
al Check	0.198	15:37
Ir Blank	0.000	15:37
al Check	0.197	15:38
Ir Blank	0.000	15:38
al Check Stats		
Average	0.1971	
Std Dev	0.0008	
Rel Std Dev(%)	0.4442	

Christopher Pauley
 Operator's Signature

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08101505 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.