



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

[www.bot.ok.gov](http://www.bot.ok.gov)

Joshua Smith  
Director

**ATTESTATION**

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-003403, in compliance with the agency's Maintenance Policy and Procedures.

Signed \_\_\_\_\_  
*JS*

Name of Position: Director/Records Custodian

Date of Attestation: 04/26/2023

AMENDED 09/19/2023



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
[www.bot.ok.gov](http://www.bot.ok.gov)

Joshua Smith  
Director

9/19/2023

It was brought to our attention of a clerical error with the documentation of the 0.04 reference standard in the maintenance packet. The maintenance packet has been amended to include the correct documentation.

A handwritten signature in blue ink, appearing to read "Christopher Pape".

Christopher Pape  
Breath Alcohol Testing Program Administrator



**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
 Chairman

Post Office Box 36307  
 Oklahoma City, Oklahoma 73136-2307  
 Phone: (405) 425-2460 Fax: (405) 425-2490  
 www.bot.ok.gov

Joshua Smith  
 Director

**Certificate of Analysis – Reference Solution**


ALPHA INSTRUMENT  
 Intoxilyzer – Alcohol Analyzer  
 Model 8000  
 10/24/2022

SN 80-002591  
 15:45

Test	g/210L	Time
Air Blank	0.000	15:45
Cal Check	0.039	15:46
Air Blank	0.000	15:47
Cal Check	0.039	15:47
Air Blank	0.000	15:48
Cal Check	0.039	15:49
Air Blank	0.000	15:49
Cal Check	0.039	15:50
Air Blank	0.000	15:50
Cal Check	0.039	15:51
Air Blank	0.000	15:52
Cal Check	0.039	15:52
Air Blank	0.000	15:53
Cal Check	0.039	15:54
Air Blank	0.000	15:54
Cal Check	0.039	15:55
Air Blank	0.000	15:56
Cal Check	0.039	15:56
Air Blank	0.000	15:57
Cal Check	0.040	15:57
Air Blank	0.000	15:58
Cal Check Stats		
Average	0.0391	
Std Dev	0.0003	
Rel Std Dev(%)	0.8087	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: Guth	Manufactured Date: 10/4/2021	Concentration: 0.040
Lot No.: 21410	Expiration Date: 9/30/2023	Equipment Model: Guth 2100
<p>The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.</p>		
Signature	Date: 10/24/2022	
Name	Permit No.: 23020	
 Christopher Pape		



# GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

RCVD  
5/9/2022  
CP

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21410** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 4, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0485%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 30, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*  
Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose values are traceable to NIST.  
All balances are calibrated annually by an outside agency using NIST traceable weights.  
Calibration verification is done prior to each use utilizing NIST traceable weights.

5/20



Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing.

# Service Overview

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

**BENCH CHECK DATE:** 4/26/2023

**SERIAL:** 80-003403

## **DOCUMENTATION WITHIN SERVICE PACKET:**

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis on Certified Solutions (4)
- Guth Laboratories Certificate of Analysis (4)



6/20

# STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

## CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION


This certifies the Intoxilyzer 8000 referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-003403 on 4/26/2023 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 28021080A2-053
  - BOT analysis average test result: 0.081
  - Cylinder expiration date: 12/5/2023
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model	Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #				Guth 2100
DR2834	DR3694	DR2455	DR3470	
Concentration	.02	.04	.10	.20
LOT #	21340	21410	22080	22050
Manufactured Date	8/24/2021	9/30/2021	3/7/2022	2/7/2022
Expiration Date	8/24/2023	9/30/2023	3/7/2024	2/7/2024
Solution Commission Date	4/25/2023	4/25/2023	4/25/2023	4/25/2023

Completed by BOT Personnel:

  
Signature Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

  
Signature Name Joshua Smith

State Director  
Title

**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE  
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 04/26/2023	TIME PROCEDURE BEGAN 15:29:5	INSTRUMENT LOCATION 80-003403
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-003403

**TEST RESULTS**

**Diagnostics**  
PASS

**Mouth Alcohol Test**  
PASS

**RFI Test**  
PASS

**Abort Test**  
PASS

**Improper Sample Test**  
PASS

**Interferent Detect Test**  
PASS

**No Sample Given Test**  
PASS

**Insufficient Sample Test**  
PASS

**Air Blank Contamination Test**  
PASS

**Range Exceeded Test**  
PASS


**Duplicate Breath Sampling Results**

Value	Test 1	Test 2	Control Test
.02	0.017	0.018	0.018
.04	0.037	0.037	0.038
.10	0.096	0.096	0.096
.20	0.193	0.192	0.192

PASS

**CERTIFYING TECHNICIAN**

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 4/26/23
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020



8/20

# STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Joshua Smith  
Director

## Certificate of Analysis – Dry Gas

### ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000

03/20/2023

SN 80-002591

13:59

Test	g/210L	Time
Air Blank	0.000	13:59
Cal Check	0.081	14:00
Air Blank	0.000	14:00
Cal Check	0.081	14:00
Air Blank	0.000	14:01
Cal Check	0.081	14:01
Air Blank	0.000	14:02
Cal Check	0.081	14:02
Air Blank	0.000	14:03
Cal Check	0.081	14:03
Air Blank	0.000	14:04
Cal Check	0.081	14:04
Air Blank	0.000	14:04
Cal Check	0.081	14:04
Air Blank	0.000	14:05
Cal Check	0.081	14:05
Air Blank	0.000	14:06
Cal Check	0.081	14:06
Air Blank	0.000	14:06
Cal Check	0.081	14:06
Air Blank	0.000	14:07
Cal Check	0.081	14:07
Air Blank	0.000	14:08
Cal Check Stats		
Average	0.0809	
Std Dev	0.0003	
Rel Std Dev (%)	0.3908	

Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)  
0.08 g/210L = 208 PPM ÷ 2605\*

Manufacturer: **ILMO** Manufacturer Cert. No.: **14351**

Cylinder Size: **105L**

Lot No.: **28021080A2** Cylinder No.: **053**

Expiration Date: **12/05/2023**

The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. \* See Dubowski, K.M. and Essary N., *Journal of Anal. Toxicol.* (1996) 20(6): pp. 484-491. [g/210L = PPM ÷2605].

Signature

Date: **03/20/2023**

Name **Christopher Pape**

Permit No.: **23020**





7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
 217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

### Certificate of Analysis

*RCV:D  
 12/19/21  
 CP*

**Certificate ID:** 14351  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 28021080A2  
**Expiration:** 12/5/2023

**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(6/210L)	NDIR
Nitrogen	Balance	[5.2 ppm]	

**Distributed by:**  
**CMI Inc.**  
 316 East Ninth Street  
 Owensboro, KY 42303  
 Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

\*Traceable to:  
 Certified Reference Material - 262.4 µmol/mol  
 Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

*[Signature]*  
 Specialty Gas Lab Tech  
*[Signature]*

11-09-2021  
 Issuance Date

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

# OFFICER'S IMPAIRED DRIVING AFFIDAVIT

10/20

Section 1

ARREST DATE <b>04/26/2023</b>	ARREST TIME <b>1600</b>	COUNTY # <b>55</b>	CITY #	CITY	CITATION #
ARREST LOCATION <b>BOT</b>					
SUBJECT NAME <b>BREATHTEST, DELTA</b>					
ADDRESS <b>3600 MLK</b>		CITY <b>OKLAHOMA CITY</b>		COUNTY <b>OKLAHOMA</b>	
DRIVER LICENSE # <b>B083463042</b>		DATE OF BIRTH <b>01/01/1985</b>		SEX <b>M</b>	
VEHICLE MAKE <b>MOCK</b>		EXPIRATION DATE <b>10/31/2022</b>		STATE <b>OK</b>	
MODEL <b>MOCK</b>		TAG # <b>MOCK</b>		RESTRICTIONS <b>73111</b>	
		STATE <b>OK</b>		ENDORSEMENTS	
		CIVIL <b>N</b>		PLACARD <b>N</b>	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, ramp, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):  
**MOCK SUBJECT TEST**



Describe person's condition (odor, actions, etc.):

**ROUTINE MAINTENANCE, INSTALLED NEW CYLINDER**

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT <b>I-8000</b>	SERIAL NUMBER <b>80-003403</b>	SPECIALIST <b>BOARD OF TESTS</b>	DATE CYLINDER EXPIRES <b>12/05/2023</b>																																				
STANDARD <b>DIY</b>	TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>23020</b>																																					
MANUFACTURER <b>ILMO</b>	LOT # <b>28021080A2</b>	DATE CYLINDER INSTALLED <b>04/26/2023</b>																																					
OBSERVATION START <b>1605</b>	TEST DATE <b>04/26/2023</b>	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr> <td>Air Blank</td> <td>0.00</td> <td>16:28</td> </tr> <tr> <td>Subject Test</td> <td>0.00</td> <td>16:29</td> </tr> <tr> <td>Breath Volume</td> <td>2.093 LITERS</td> <td></td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>16:29</td> </tr> <tr> <td>Wait</td> <td></td> <td>16:31</td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>16:31</td> </tr> <tr> <td>Subject Test</td> <td>0.00</td> <td>16:32</td> </tr> <tr> <td>Breath Volume</td> <td>2.109 LITERS</td> <td></td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>16:32</td> </tr> <tr> <td>Cal Check</td> <td>0.08</td> <td>16:33</td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>16:33</td> </tr> </tbody> </table>				Test	g/210L	Time	Air Blank	0.00	16:28	Subject Test	0.00	16:29	Breath Volume	2.093 LITERS		Air Blank	0.00	16:29	Wait		16:31	Air Blank	0.00	16:31	Subject Test	0.00	16:32	Breath Volume	2.109 LITERS		Air Blank	0.00	16:32	Cal Check	0.08	16:33	Air Blank	0.00	16:33
Test	g/210L	Time																																					
Air Blank	0.00	16:28																																					
Subject Test	0.00	16:29																																					
Breath Volume	2.093 LITERS																																						
Air Blank	0.00	16:29																																					
Wait		16:31																																					
Air Blank	0.00	16:31																																					
Subject Test	0.00	16:32																																					
Breath Volume	2.109 LITERS																																						
Air Blank	0.00	16:32																																					
Cal Check	0.08	16:33																																					
Air Blank	0.00	16:33																																					
																																							
<div style="border: 1px solid black; padding: 5px; display: inline-block;">                 Test Result BrAC g/210L <b>0.00</b> </div>																																							
SIGNATURE OF OPERATOR 																																							
NAME <b>PAPE, CHRISTOPHER</b>		PERMIT # <b>0000023020</b>																																					
BADGE # <b>000000484</b>		AGENCY <b>BOARD OF TESTS</b>																																					

### What happens next?

1. The Arresting Officer is required to submit the Officer's Impaired Driving Affidavit to the State of Oklahoma.
2. Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
3. The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma. To confirm or change your address on file, contact Service Oklahoma.
4. The Revocation and/or Disqualification will be effective 55 days from the date of the Order, including 10 days allowed for mailing, 47 O.S. §§2-116, 753, 754.
5. The Order of Revocation and/or Disqualification will contain important information about your rights to appeal the revocation and/or disqualification of your driving privileges. To appeal the applicable Order of Revocation and/or Disqualification your petition must be filed with the District Court of the County in which you were arrested within 40 days of the date of the order, including 10 days allowed for mailing, 47 O.S. §§2-116, 6-211.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

Section 3

NAME <b>PAPE, CHRISTOPHER</b>	AGENCY <b>BOARD OF TESTS</b>
BADGE <b>000000484</b>	AGENCY ADDRESS <b>INTER-AGENCY</b>
	ZIP CODE <b>00000</b>
	PHONE <b>4054252460</b>

OTHER WITNESSES:

Section 5

NAME <b>1</b>	TITLE	ADDRESS	PHONE
NAME <b>2</b>	TITLE	ADDRESS	PHONE

# OFFICER'S IMPAIRED DRIVING AFFIDAVIT

11/20

Section 1

ARREST DATE <b>04/26/2023</b>	ARREST TIME <b>1600</b>	COUNTY # <b>55</b>	CITY #	CITY	CITATION #
ARREST LOCATION <b>BOT</b>					
SUBJECT NAME <b>BREATHTEST, DELTA</b>		COUNTY <b>OKLAHOMA</b>		HEIGHT <b>510</b>	
ADDRESS <b>3600 MLK</b>		DATE OF BIRTH <b>01/01/1985</b>		WEIGHT <b>200</b>	
DRIVER LICENSE # <b>B083463042</b>		CITY <b>OKLAHOMA CITY</b>		STATE <b>OK</b>	
VEHICLE MAKE <b>MOCK</b>		EXPIRATION DATE <b>10/31/2022</b>		RESTRICTIONS <b>73111</b>	
MODEL <b>MOCK</b>		TAG # <b>MOCK</b>		ENDORSEMENTS	
		STATE <b>OK</b>		STATE <b>OK</b>	
		CIVIL <b>N</b>		PLACARD <b>N</b>	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):  
**MOCK SUBJECT TEST**

Describe person's condition (odor, actions, etc.):

## ROUTINE MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST  
BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT <b>I-8000</b>	SERIAL NUMBER <b>80-003403</b>	SPECIALIST <b>BOARD OF TESTS</b>
STANDARD <b>DIY</b>	TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>23020</b>
MANUFACTURER <b>ILMO</b>	LOT # <b>28021080A2</b>	DATE CYLINDER INSTALLED <b>04/26/2023</b>
OBSERVATION START <b>1605</b>	TEST DATE <b>04/26/2023</b>	DATE CYLINDER EXPIRES <b>12/05/2023</b>

**RESULT TABLE**

Test	Time	g/210L
Air Blank	0.00	16:28
Subject Test	0.00	16:29
Breath Volume	2.093 LITERS	
Air Blank	0.00	16:29
Wait		16:31
Air Blank	0.00	16:31
Subject Test	0.00	16:31
Breath Volume	2.109 LITERS	
Air Blank	0.00	16:32
Cal Check	0.08	16:33
Air Blank	0.00	16:33

A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.



Test Result  
BrAC g/210L  
**0.00**

I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.

*Christopher Pape*  
SIGNATURE OF OPERATOR

NAME <b>PAPE, CHRISTOPHER</b>
BADGE # <b>000000484</b>
PERMIT # <b>0000023020</b>
AGENCY <b>BOARD OF TESTS</b>

Section 3

### What happens next?

- The Arresting Officer is required to submit the Officer's Impaired Driving Affidavit to the State of Oklahoma.
- Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
- The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma. To confirm or change your address on file, contact Service Oklahoma.
- The Revocation and/or Disqualification will be effective 55 days from the date of the Order, including 10 days allowed for mailing. 47 O.S. §§2-116, 753, 754.
- The Order of Revocation and/or Disqualification will contain important information about your rights to appeal the revocation and/or disqualification of your driving privileges. To appeal the applicable Order of Revocation and/or Disqualification your petition must be filed with the District Court of the County in which you were arrested within 40 days of the date of the order, including 10 days allowed for mailing. 47 O.S. §§2-116, 6-211.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME <b>PAPE, CHRISTOPHER</b>	AGENCY <b>BOARD OF TESTS</b>
BADGE <b>000000484</b>	PHONE <b>4054252460</b>
AGENCY ADDRESS <b>INTER - AGENCY</b>	ZIP CODE <b>00000</b>

Section 4



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman  
Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director


**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002591  
09/23/2022 09:37

Test	g/210L	Time
Air Blank	0.000	09:37
Cal Check	0.019	09:38
Air Blank	0.000	09:39
Cal Check	0.019	09:39
Air Blank	0.000	09:40
Cal Check	0.019	09:41
Air Blank	0.000	09:41
Cal Check	0.019	09:42
Air Blank	0.000	09:42
Cal Check	0.019	09:43
Air Blank	0.000	09:44
Cal Check	0.019	09:44
Air Blank	0.000	09:45
Cal Check	0.019	09:46
Air Blank	0.000	09:46
Cal Check	0.019	09:47
Air Blank	0.000	09:47
Cal Check	0.019	09:48
Air Blank	0.000	09:49
Cal Check	0.019	09:49
Air Blank	0.000	09:50
Cal Check Stats		
Average	0.0190	
Std Dev	0.0000	
Rel Std Dev (%)	0.0000	

**NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.**

Manufacturer: Guth	Manufactured Date: 8/24/2021	Concentration: 0.02
Lot No.: 21340	Expiration Date: 8/24/2023	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature	Date: 9/23/2022	
Name Christopher Pape	Permit No.: 23020	


**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

 RCVD  
 5/9/2022  
 CP

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21340** ✓ of  
 Alcohol Reference Solution for Simulator were analyzed by  
 gas chromatography on **August 25, 2021**, using a Perkin Elmer Gas  
 Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.0242%** (w/vol) ethyl alcohol. The expiration date for this lot  
 number is **August 24, 2023** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
 34°C +/- .2°C, this solution will give a breath alcohol  
 analysis instrument reading of **0.020 g/210L +/- 3%**.

The alcohol and water used in this solution were  
 free of test interfering substances.

Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

*NIST Traceability:  
 Testing was conducted using Cerilliant Reference Standard lot number **FN05122004** whose  
 values are traceable to NIST.  
 All balances are calibrated annually by an outside agency using NIST traceable weights.  
 Calibration verification is done prior to each use utilizing NIST traceable weights.*



14/20

# STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

## Certificate of Analysis – Reference Solution

ALPHA INSTRUMENT Z  
Intoxilyzer – Alcohol Analyzer SN 80-004923  
Model 8000  
03/08/2022 11:53

Test	g/210L	Time
Air Blank	0.000	11:53
Cal Check	0.041	11:54
Air Blank	0.000	11:54
Cal Check	0.041	11:55
Air Blank	0.000	11:56
Cal Check	0.042	11:56
Air Blank	0.000	11:57
Cal Check	0.042	11:58
Air Blank	0.000	11:58
Cal Check	0.042	11:59
Air Blank	0.000	11:59
Cal Check	0.041	12:00
Air Blank	0.000	12:01
Cal Check	0.041	12:01
Air Blank	0.000	12:02
Cal Check	0.041	12:02
Air Blank	0.000	12:03
Cal Check	0.041	12:04
Air Blank	0.000	12:04
Cal Check	0.042	12:05
Air Blank	0.000	12:05
Cal Check Stats		
Average	0.0414	
Std Dev	0.0005	
Rel Std Dev(%)	1.2473	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.	
Manufacturer: <b>Guth</b>	Manufactured Date: <b>03/01/2021</b>
Lot No.: <b>21070</b>	Expiration Date: <b>03/01/2023</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.	
Signature 	Date: <b>03/08/2022</b>
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>
Concentration: <b>0.04</b> Equipment Model: <b>Guth 2100</b>	



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

ACU'D  
8/26/21  
CP

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 2, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0488%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 1, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose values are traceable to NIST.  
All balances are calibrated annually by an outside agency using NIST traceable weights.  
Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**


ALPHA INSTRUMENT  
Intoxilyzer – Alcohol Analyzer SN 80-002591  
Model 8000  
06/27/2022 11:40

Test	g/210L	Time
Air Blank	0.000	11:40
Cal Check	0.100	11:41
Air Blank	0.000	11:41
Cal Check	0.100	11:42
Air Blank	0.000	11:43
Cal Check	0.100	11:43
Air Blank	0.000	11:44
Cal Check	0.100	11:45
Air Blank	0.000	11:45
Cal Check	0.100	11:46
Air Blank	0.000	11:47
Cal Check	0.101	11:47
Air Blank	0.000	11:48
Cal Check	0.101	11:49
Air Blank	0.000	11:49
Cal Check	0.101	11:50
Air Blank	0.000	11:50
Cal Check	0.100	11:51
Air Blank	0.000	11:52
Cal Check	0.100	11:52
Air Blank	0.000	11:53
Cal Check Stats		
Average	0.1003	
Std Dev	0.0004	
Rel Std Dev(%)	0.4815	

**NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.**

Manufacturer: Guth	Manufactured Date: 03/07/2022	Concentration: 0.10
Lot No.: 22080	Expiration Date: 03/07/2024	Equipment Model: Guth 2100
<p>The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.</p>		
Signature	Date: 06/27/2022	
Name Christopher Pape	Permit No.: 23020	




**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

 RCVD  
 5/9/2022  
 CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22080** of

Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 9, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 7, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

*NIST Traceability:*  
 Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.  
 All balances are calibrated annually by an outside agency using NIST traceable weights.  
 Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT  
Intoxilyzer - Alcohol Analyzer SN 80-002591  
Model 8000 04/19/2023 10:57

Test	g/210L	Time
Air Blank	0.000	10:58
Cal Check	0.198	10:58
Air Blank	0.000	10:59
Cal Check	0.198	11:00
Air Blank	0.000	11:00
Cal Check	0.199	11:01
Air Blank	0.000	11:02
Cal Check	0.199	11:02
Air Blank	0.000	11:03
Cal Check	0.198	11:04
Air Blank	0.000	11:04
Cal Check	0.199	11:05
Air Blank	0.000	11:06
Cal Check	0.199	11:06
Air Blank	0.000	11:07
Cal Check	0.199	11:08
Air Blank	0.000	11:08
Cal Check	0.199	11:08
Air Blank	0.000	11:09
Cal Check	0.199	11:10
Air Blank	0.000	11:11
Cal Check Stats		
Average	0.1987	
Std Dev	0.0004	
Rel Std Dev (%)	0.2431	

<b>NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.</b>	
Manufacturer: Guth	Manufactured Date: 2/7/2022
Lot No.: 22050	Expiration Date: 2/7/2024
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.	
Signature	Date: 4/19/2023
Name Christopher Pape	Permit No.: 23020



# GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22050** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 9, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2437%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 7, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

RCVD  
2/23/23  
CP

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:  
Testing was conducted using Cerilliant Reference Standard lot number FN06231703 whose values are traceable to NIST.  
All balances are calibrated annually by an outside agency using NIST traceable weights.  
Calibration verification is done prior to each use utilizing NIST traceable weights.*

Location Returned From  
Antlers PD

Unit Serial #  
80-003403

Date:  
3/7/23

Inventory Process  
34520080A1-021

Unit checked in to BOT  
 New Unit Checked out

Unlink Gas Canister from Unit  
 Remove old cylinder from unit (After Lab Evaluation)

Old cylinder disposed in inventory  
 Old cylinder archived

Check in Procedure  
 Notes from the Field  
 Routine Swap  Issues in the Field

Connect unit to ethernet connection 172.019.004. 141  
 Download any stored records 2  
 Change unit location to Serial number  
 Update Unit Date and time

Previous time: 3/20/2023 14:01  
 New Time: 3/20/2023 14:02

Execute a new Forms load on the unit

Lab Evaluation  
 Breath Hose Hold Vacuum?:  Test Chamber Hold Vacuum?:

DVM  
 3µm 12474 Chamber Temp 47.00 Snap Test ok?   
 9µm 12972 Breath Hose Temp 45 RFI Detected?

Peripheral Tests  
 Speaker (F5)  Power LED Green (F6)  Red (F7)  Fan (F9)   
 Gas Regulator (F11)  Pump (F10)  Receipt paper

Atmospheric Sensor  
 Pressure 977 Correction Factor 1.03 Flow Sensor Tank Sensor  
 Resistance 128 Value 995  
 Is regulator close to Tank Sensor Value?  /  No

Sent to CMI (if needed)  
 Date Sent \_\_\_\_\_ Reason Sent \_\_\_\_\_  
 Invoice Sent \_\_\_\_\_  
 Date Returned \_\_\_\_\_

Repairs needed:

Repairs Made:

Ready for Bench Check?

Inventory Process  
 Linked Cylinder to Unit  
 Changed Unit Status  
 Bench Check Report Completed

Bench Check Technician Report  
 Mock Subject Test Completed  
 Certificate of Calibration and Operation  
 Service Overview Completed

To complete if unit sent to CMI  
 CMI Workorder  
 CMI Certificate of Calibration  
 CMI Service Evaluation Form