



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

ATTESTATION

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-004926 , in compliance with the agency's Maintenance Policy and Procedures.

This packet consists of 11 pages.

Signed 

Name of Position: Director/Records Custodian

Date of Attestation: 02/17/2022

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing



2/11

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

BENCH CHECK DATE: 2/17/2022

SERIAL: 80-004926

## DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis on Certified Solutions (4)



# STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

## CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION

This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.


- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-004926 on 2/17/2022 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 34520080A1-026
  - BOT analysis average test result: 0.081
  - Cylinder expiration date: 2/5/2023
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model	Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #	DR3685	DR3694	DR3594	DR3754
Concentration	.02	.04	.10	.20
LOT #	21120	21070	21190	20160
Manufactured Date	4/7/2021	3/2/2021	6/9/2021	3/18/2020
Expiration Date	4/7/2023	3/1/2023	6/8/2023	3/18/2022
Solution Commission Date	1/31/2022	1/31/2022	1/31/2022	1/31/2022

Completed by BOT Personnel:

  
Signature/Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

Certified by:  
  
Signature Joshua Smith  
Name

State Director  
Title



**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE  
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 02/17/2022	TIME PROCEDURE BEGAN 11:22:3	INSTRUMENT LOCATION 80-004926
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-004926

**TEST RESULTS**

**Diagnostics**

PASS

**Mouth Alcohol Test**

PASS

**RFI Test**

PASS

**Abort Test**

PASS

**Improper Sample Test**

PASS

**Interferent Detect Test**

PASS

**No Sample Given Test**

PASS

**Insufficient Sample Test**

PASS

**Air Blank Contamination Test**

PASS

**Range Exceeded Test**

PASS


**Duplicate Breath Sampling Results**

Value	Test 1	Test 2	Control Test
.02	0.020	0.019	0.020
.04	0.038	0.037	0.041
.10	0.098	0.096	0.099
.20	0.197	0.192	0.199

PASS

**CERTIFYING TECHNICIAN**

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 2/17/2022
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020

**Certificate of Analysis – Dry Gas**



5/11

1/21/2021	0.080
Date of Analysis	Labelled target value (g/210L)
026	0.081
Cylinder #	Average test result
Christopher Papo <i>[Signature]</i> 23020	
BOT Technician Name and Signature	



ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
01/21/2021 13:37

In accordance with BOT Policy and Procedure Statement BRT-2.2.0, Maintenance of the above referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.  
**7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790**  
**217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com**

Test	g/210L	Time
Air Blank	0.000	13:37
Cal Check	0.080	13:38
Air Blank	0.000	13:38
Cal Check	0.081	13:39
Air Blank	0.000	13:39
Cal Check	0.081	13:39
Air Blank	0.000	13:40
Cal Check	0.081	13:40
Air Blank	0.000	13:41
Cal Check	0.080	13:41
Air Blank	0.000	13:42
Cal Check	0.080	13:42
Air Blank	0.000	13:42
Cal Check	0.081	13:43
Air Blank	0.000	13:43
Cal Check	0.081	13:44
Air Blank	0.000	13:44
Cal Check	0.081	13:44
Air Blank	0.000	13:45
Cal Check	0.081	13:45
Air Blank	0.000	13:46
Cal Check Stats		
Average	0.0807	
Std Dev	0.0004	
Rel Std Dev(%)	0.5985	

**Certificate of Analysis**

**Certificate ID:** 13552 ✓  
**Part #:** BAC105L080T\*  
**Cylinder Size:** 105L  
**Lot Number:** 34520080A1 ✓  
**Expiration:** 2/5/2023 ✓

*RCVD  
1/21/2021  
CP*

**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L)	NDIR
Nitrogen	Balance	[5.2 ppm]	

**Distributed by:**

**CMI Inc.**  
316 East Ninth St  
Owensboro, KY  
Phone 866-835-1111  
[www.alcoholtes.com](http://www.alcoholtes.com)

*[Signature]* 2020  
Operator's Signature

\*Traceable to:  
Certified Reference Material - 262.4 µmol/mol  
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

*[Signature]*  
Specialty Gas Lab Tech

*[Signature]*

01-11-2021  
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

6/11

Section 1

ARREST DATE 02/17/2022		ARREST TIME 1215		COUNTY # 55		CITY #		CITATION #			
ARREST LOCATION BOT				CITY			COUNTY OKLAHOMA				
SUBJECT NAME BREATHTEST, ALPHA				DATE OF BIRTH 01/01/1980		SEX M		HEIGHT 510		WEIGHT 150	
ADDRESS 3600 MARTIN LUTHER KING A				CITY OKLAHOMA CITY			STATE OK		ZIP CODE 73111		
DRIVER LICENSE # B083462869				EXPIRATION DATE 10/31/2022		STATE OK		CLASS D		RESTRICTIONS	ENDORSEMENTS
VEHICLE MAKE MOCK			MODEL MOCK			TAG # MOCK		STATE OK		CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST

Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE

### THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-004926		SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080		PERMIT NUMBER 23020																																					
MANUFACTURER ILMO		LOT # 34520080A1		DATE CYLINDER INSTALLED 02/17/2022																																					
				DATE CYLINDER EXPIRES 02/05/2023																																					
OBSERVATION START 1220		TEST DATE 02/17/2022		<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>Christopher Pape</i> SIGNATURE OF OPERATOR</p>																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>12:45</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>12:46</td></tr> <tr><td>  Breath Volume</td><td>2.101 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:46</td></tr> <tr><td>Wait</td><td></td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:48</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>12:49</td></tr> <tr><td>  Breath Volume</td><td>2.121 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:50</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:51</td></tr> </tbody> </table>						Test	g/210L	Time	Air Blank	0.00	12:45	Subject Test	0.00	12:46	Breath Volume	2.101 LITERS		Air Blank	0.00	12:46	Wait		12:48	Air Blank	0.00	12:48	Subject Test	0.00	12:49	Breath Volume	2.121 LITERS		Air Blank	0.00	12:50	Cal Check	0.08	12:50	Air Blank	0.00	12:51
Test	g/210L	Time																																							
Air Blank	0.00	12:45																																							
Subject Test	0.00	12:46																																							
Breath Volume	2.101 LITERS																																								
Air Blank	0.00	12:46																																							
Wait		12:48																																							
Air Blank	0.00	12:48																																							
Subject Test	0.00	12:49																																							
Breath Volume	2.121 LITERS																																								
Air Blank	0.00	12:50																																							
Cal Check	0.08	12:50																																							
Air Blank	0.00	12:51																																							
NAME PAPE, CHRISTOPHER		BADGE # 000000484		PERMIT # 000023020																																					
AGENCY BOARD OF TESTS																																									



Test Result BrAC g/210L <h1 style="margin: 0;">0.00</h1>
--

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED.** Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below.

Section 3

You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

7/11

Section 1

ARREST DATE 02/17/2022		ARREST TIME 1215		COUNTY # 55		CITY #		CITATION #	
ARREST LOCATION BOT				CITY			COUNTY OKLAHOMA		
SUBJECT NAME BREATHTEST, ALPHA				DATE OF BIRTH 01/01/1980		SEX M		HEIGHT 510	WEIGHT 150
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY			STATE OK		ZIP CODE 73111	
DRIVER LICENSE # B083462869			EXPIRATION DATE 10/31/2022		STATE OK	CLASS D	RESTRICTIONS	ENDORSEMENTS	
VEHICLE MAKE MOCK		MODEL MOCK		TAG # MOCK		STATE OK	CMV N	PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE

### THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

#### BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-004926		SPECIALIST BOARD OF TESTS							
STANDARD Dry		TARGET VALUE 0.080		PERMIT NUMBER 23020							
MANUFACTURER ILMO		LOT # 34520080A1		DATE CYLINDER INSTALLED 02/17/2022	DATE CYLINDER EXPIRES 02/05/2023						
OBSERVATION START 1220	TEST DATE 02/17/2022	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>[Signature]</i> SIGNATURE OF OPERATOR</p>									
RESULT TABLE											
Test	g/210L	Time									
Air Blank	0.00	12:45									
Subject Test	0.00	12:46									
Breath Volume	2.101 LITERS										
Air Blank	0.00	12:46									
Wait		12:48									
Air Blank	0.00	12:48									
Subject Test	0.00	12:49									
Breath Volume	2.121 LITERS										
Air Blank	0.00	12:50									
Cal Check	0.08	12:50									
Air Blank	0.00	12:51									
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2">NAME PAPE, CHRISTOPHER</td> </tr> <tr> <td>BADGE # 0000000484</td> <td>PERMIT # 0000023020</td> </tr> <tr> <td colspan="2">AGENCY BOARD OF TESTS</td> </tr> </table>						NAME PAPE, CHRISTOPHER		BADGE # 0000000484	PERMIT # 0000023020	AGENCY BOARD OF TESTS	
NAME PAPE, CHRISTOPHER											
BADGE # 0000000484	PERMIT # 0000023020										
AGENCY BOARD OF TESTS											
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2">Test Result BrAC g/210L</td> </tr> <tr> <td colspan="2" style="font-size: 2em;">0.00</td> </tr> </table>						Test Result BrAC g/210L		0.00			
Test Result BrAC g/210L											
0.00											

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED.** Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below.

**You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.**

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460



Certificate of Analysis – Wet Bath

1/31/2022	0.020
Date of Analysis	Labeled to (g) value (g/210L)
21120	0.022
Lot Number	Assigned to result (g/210L)
Christopher Pope BOT Technician Name and Signature	



8/11

**GUTH LABORATORIES, INC.**  
 690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BOT Maintenance Policy and Procedure Statement, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

RCU'D  
 8/26/21  
 CP

**Certified Alcohol Reference Solution for Simulator**

ALPHA INSTRUMENT 2  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 00-004923  
 1/31/2022 08:32

Random Samples of Lot Number 21120 of

Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 16, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0237% (w/vol) ethyl alcohol. The expiration date for this lot number is April 7, 2023 at 11:59 PM.

test	g/210L	Time
r Blank	0.000	08:32
al Check	0.022	08:33
r Blank	0.000	08:34
al Check	0.022	08:34
r Blank	0.000	08:35
al Check	0.022	08:35
r Blank	0.000	08:36
al Check	0.021	08:37
r Blank	0.000	08:37
al Check	0.021	08:38
r Blank	0.000	08:38
al Check	0.022	08:39
r Blank	0.000	08:40
al Check	0.021	08:40
r Blank	0.000	08:41
al Check	0.022	08:42
r Blank	0.000	08:42
al Check	0.021	08:43
r Blank	0.000	08:43
al Check	0.021	08:44
r Blank	0.000	08:45

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Check Stats  
 Average 0.0215  
 Std Dev 0.0005  
 Std Dev(%) 2.4512

*Christopher Pope*  
 Operator's Signature

*Ted L. Pauley*  
 Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN08031603 whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Certificate of Analysis – Wet Bath	
12/13/2021	0.040
Date of Analysis	Labelled Target value (g/210L)
21070	0.042
Lot Number	Assigned test result (g/210L)
Christopher Pauley	
BOT Technician Name and Signature	



**GUTH LABORATORIES, INC.**

500 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

RCV'D  
8/26/21  
CP

In accordance with BOT Maintenance Policy and Procedure Statement, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

### CERTIFICATE OF ANALYSIS

INTOXILYZER 8000
Instrument Initialization
8133.12
09:27 12/01/2021

### Certified Alcohol Reference Solution for Simulator

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
12/13/2021 12:09

Random Samples of Lot Number **21070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 2, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0488%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 1, 2023** at 11:59 PM.

Test	g/210L	Time
Air Blank	0.000	12:09
Cal Check	0.042	12:10
Air Blank	0.000	12:10
Cal Check	0.042	12:11
Air Blank	0.000	12:12
Cal Check	0.042	12:12
Air Blank	0.000	12:13
Cal Check	0.042	12:14
Air Blank	0.000	12:14
Cal Check	0.042	12:15
Air Blank	0.000	12:15
Cal Check	0.042	12:16
Air Blank	0.000	12:17
Cal Check	0.042	12:17
Air Blank	0.000	12:18
Cal Check	0.042	12:18
Air Blank	0.000	12:19
Cal Check	0.042	12:20
Air Blank	0.000	12:20
Cal Check	0.042	12:21
Air Blank	0.000	12:21
Cal Check Stats		
Average	0.0420	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

Operator's Signature

#### NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**Certificate of Analysis – Wet Bath**



10/11

12/13/2021	0.100
Date of Analysis	Labeled (g) net value (g/210L)
21190	0.103
Lot Number	Weight (g) (g/210L)
Christopher [Signature] 500 NORTH 57th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470	
BOT Technician Name and Signature	

BRT Form 6.0

In accordance with BOT Maintenance Policy and Procedure Statement, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

RCV'D  
8/26/21  
CP

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
12/13/2021 12:45

**Certified Alcohol Reference Solution for Simulator**

Test	g/210L	Time
Air Blank	0.000	12:46
Cal Check	0.107	12:46
Air Blank	0.000	12:47
Cal Check	0.103	12:48
Air Blank	0.000	12:48
Cal Check	0.102	12:49
Air Blank	0.000	12:49
Cal Check	0.102	12:50
Air Blank	0.000	12:51
Cal Check	0.102	12:51
Air Blank	0.000	12:52
Cal Check	0.102	12:53
Air Blank	0.000	12:53
Cal Check	0.102	12:54
Air Blank	0.000	12:54
Cal Check	0.102	12:55
Air Blank	0.000	12:56
Cal Check	0.102	12:56
Air Blank	0.000	12:57
Cal Check	0.101	12:58
Air Blank	0.000	12:58
Cal Check Stats		
Average	0.1025	
Std Dev	0.0016	
Rel Std Dev(%)	1.6096	

Random Samples of Lot Number 21190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on June 9, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is June 8, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

[Signature]  
Operator's Signature

[Signature]  
Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Certificate of Analysis - wet Barn

07/07/2020	0.200
Date of Analysis	Labeled target value (g/210L)
20160	0.197
Lot Number	Average test result (g/210L)
Christopher Pipe	0.23030
BOT Technician Name and Signature	



11/11

**GUTH LABORATORIES, INC.**  
 690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

*Rec'd 05/20/20 CP*

**Certified Alcohol Reference Solution for Simulator**

LPMA UNIT SUB  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SH 80-003400  
 7/07/2020 15:25

Test	g/210L	Time
Ir Blank	0.000	15:26
al Check	0.196	15:26
Ir Blank	0.000	15:27
al Check	0.196	15:28
Ir Blank	0.000	15:28
al Check	0.197	15:29
Ir Blank	0.000	15:29
al Check	0.196	15:30
Ir Blank	0.000	15:31
al Check	0.198	15:31
Ir Blank	0.000	15:32
al Check	0.198	15:33
Ir Blank	0.000	15:33
al Check	0.197	15:34
Ir Blank	0.000	15:35
al Check	0.198	15:35
Ir Blank	0.000	15:36
al Check	0.198	15:37
Ir Blank	0.000	15:37
al Check	0.197	15:38
Ir Blank	0.000	15:38
al Check Stats		
Average	0.1971	
Std Dev	0.0008	
Rel Std Dev(%)	0.4442	

Random Samples of Lot Number **20160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 20, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2427% (w/vol) ethyl alcohol**. The expiration date for this lot number is March 18, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

*[Signature]*  
 Operator's Signature

*[Signature]*  
 Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

**NIST Traceability:**  
 Testing was conducted using Cerilliant Reference Standard lot number FN08101505 whose values are traceable to NIST.  
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.