

Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing.



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 08/05/2015

SERIAL: 80-005136

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- ILMO Certificate of Analysis
- ILMO MSDS sheet (2)
- Mock Subject Test
- GUTH Certificates of Analysis on Certified Solutions (4)
- BOT Certificates of Analysis on Certified Solutions (4)

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 08/05/2015	TIME PROCEDURE BEGAN 15:38	INSTRUMENT LOCATION THE VILLAGE PD
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-005136

TEST RESULTS

Diagnostics
PASS

Mouth Alcohol Test
PASS

RFI Test
PASS

Abort Test
PASS

Improper Sample Test
PASS

Interferent Detect Test
PASS

No Sample Given Test
PASS

Insufficient Sample Test
PASS

Air Blank Contamination Test
PASS

Range Exceeded Test
PASS


Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.019	0.020	0.020
.04	0.040	0.040	0.040
.10	0.098	0.099	0.100
.20	0.197	0.197	0.199

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 08-05-2015
NAME SMITH, JOSHUA	PERMIT # 0000016837



3600 N. Martin Luther King Bldg #9
 Oklahoma City, OK 73111
 405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 8 / 5 / 15 Start Time: 1500 End Time: 1640

INSTRUMENT

Intoxilyzer Model

Serial #

8000 80-005136

GAS CANISTER

LOT #

EXP Date

13314080A1 6-5-2016

REFERENCE


Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR3591	DR3753	DR3594	DR3754
Concentration			
.02	.04	.10	.20
LOT #			
14100	14010	14110	14130
Manufactured Date			
04/15/2014	01/06/2014	05/01/2014	05/12/2014
Expiration Date			
04/15/2016	01/06/2016	05/01/2016	05/12/2016
Solution Commission Date			
<u>7-31-15</u>	<u>7-31-15</u>	<u>7-31-15</u>	<u>7-31-15</u>

REASON FOR BENCH CHECK

- POST REPAIR CYLINDER REPLACEMENT
 TROUBLESHOOTING ROUTINE MAINTENANCE

COMMENTS

Full system inspection completed. Replaced defective ETHERNET CARD AND TESTED EXTENSIVELY. NO OTHER FAULTS FOUND.

 16837
 Technician Signature / Permit #

Josh Smith
 Printed Name

Dry Gas Certificate of Analysis

Board of Tests for Alcohol and Drug Influence
 Enhancing public safety by ensuring the accuracy and scientific reliability
 of blood and breath testing.



ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 08/05/2015 15:21

Test	g/210L	Time
Air Blank	0.000	15:21
Cal Check	0.080	15:22
Air Blank	0.000	15:22
Cal Check	0.080	15:23
Air Blank	0.000	15:23
Cal Check	0.080	15:23
Air Blank	0.000	15:24
Cal Check	0.080	15:24
Air Blank	0.000	15:25
Cal Check	0.080	15:25
Air Blank	0.000	15:26
Cal Check	0.080	15:26
Air Blank	0.000	15:27
Cal Check	0.080	15:27
Air Blank	0.000	15:27
Cal Check	0.080	15:28
Air Blank	0.000	15:28
Cal Check	0.080	15:29
Air Blank	0.000	15:29
Cal Check	0.080	15:29
Air Blank	0.000	15:30
Cal Check Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

 Operator's Signature

Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (PPM) 0.0798 g/210L = 208 PPM ÷ 2605*		
Manufacturer: ILMO	Manufacturer Cert. No.: 6717	Cylinder Size: 105L
Lot No.: 13314080A1	Cylinder No.: 004	Expiration Date: 06-05-2016
The above tests were performed in accordance with procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. Based upon the attached manufacturer's certificate of analysis and the above tests, the above referenced unit meets the requirements for use as an alcohol breath reference method in accordance with 40 OAC 25-1-3, and Board of Tests resolutions dated June 12, 2008 and February 22, 2013. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM ÷ 2605].		
Signature	Date: 08/05/2015	
Name Joshua Smith	Permit No.: 16837	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 6717
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 13314080A1
Expiration: 6/5/2016

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)


Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

*NIST Standard Reference Material
Cylinder No. CCI4290 / Job No. 09160202
Certified 212.8 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition
and direct sunlight. Do not allow storage area to
exceed 52 °C (125 °F).


Specialty Gas Lab Tech


Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com





Material Safety Data Sheet

Ethanol in Nitrogen

www.ilmoproducts.com

Section 1: Product and Company Identification

ILMO Products Company
 7 Eastgate Drive, Jacksonville, Illinois 62850
 217-245-2189 888-243-8393
 Fax 217-243-7634
 E-mail: info@ilmoproducts.com
 Web: www.ilmoproducts.com

Product Code: Ethanol in Nitrogen
 Part Number: BAC
 Gas Name | Concentration
 Ethanol | 5-500 ppm
 Nitrogen | Balance

Chemical Substance	Chemical Family	Trade Names
Ethanol ETHYL ALCOHOL, 100%	hydroxyls, aliphatic, alcoholic, aliphatic	ETHANOL; ETHYL ALCOHOL; ALCOHOL; ANHYDROUS; ALGRAN; AMYL ALCOHOL; Absolute alcohol; Anhydrous ethanol; Ethanol denatured; Fermentation alcohol; Grain alcohol; 1-Hydroxyethane; Methyl carbinol; Ethyl alcohol anhydrous; Absolute ethanol; Denatured ethanol; ETHYL HYDRATE; ETHYL HYDROXIDE; JAYSOL; TEGSOL; STOC 409189; UN 1170; C2H6O
Nitrogen NITROGEN COMPRESSED GAS	Inorganic, gas	DIATOMIC NITROGEN; DINITROGEN; NITROGEN; NITROGEN-14; NITROGEN GAS; UN 1066; N2

Section 2: Hazards Identification

Ethanol	Description	Main Health Hazard
	Colorless, pleasant odor Flammable liquid and vapor. Vapor may cause flash fire.	Respiratory tract irritation, skin irritation, eye irritation, liver damage, central nervous system depression
Nitrogen	Colorless, odorless Containers may rupture or explode if exposed to heat.	Difficulty breathing

Likely Routes of Exposure:

	Inhalation	Ingestion	Eye	Skin	Health Effects	Target Organs	Medical Condition Aggravated by
Ethanol	Irritation, difficulty breathing, headache, drowsiness, symptoms of drunkenness	Rash, low body temperature, vomiting, digestive disorders, irregular heartbeat, headache, drowsiness, symptoms of drunkenness, disorientation, dilated pupils, lung congestion, liver damage, convulsions, coma	Irritation, tearing	Mild irritation, rash	Respiratory tract irritation, skin irritation, eye irritation, liver damage, central nervous system depression	Central nervous system, liver	Central nervous system disorders, kidney disorders, liver disorders

ILMO Products Company
 Generated by the MSDS Manager from AsterRisk, LLC. All Rights Reserved.

page 1 of 8
 Generated: 07/20/2012 09:38:09

	Inhalation	Ingestion	Eye	Skin	Health Effects	Target Organs	Medical Condition Aggravated by
Nitrogen	Nausea, vomiting, difficulty breathing, headache, drowsiness, dizziness, tingling sensation, loss of coordination, convulsions, coma	Ingestion of a gas is unlikely	Contact with cornea exposing gas may cause burns or frostbite	No information on significant adverse effects	Difficulty breathing	Respiratory system	Pre-existing conditions of respiratory system.

This material is considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910.1200).

Section 3: Composition/Information on Ingredients

	CAS #	% by Weight
Ethanol	64-17-5	5-500 ppm
Nitrogen	7727-37-9	Balance

Section 4: First Aid Measures

	Skin Contact	Eye Contact	Ingestion	Inhalation	Note to Physicians
Ethanol	Wash skin with soap and water for at least 15 minutes while removing contaminated clothing and shoes. Get medical attention if needed. Thoroughly clean dry contaminated clothing and shoes before reuse.	Flush eyes with plenty of water for at least 15 minutes. Then get immediate medical attention.	Contact local poison control center or physician immediately. Never make an unconscious person vomit or drink fluids. When vomiting occurs, keep head lower than hips to help prevent aspiration. If person is unconscious, turn head to side. Get medical attention immediately.	If adverse effects occur, remove to unconfined area. Give artificial respiration if not breathing. If breathing is difficult, oxygen should be administered by qualified personnel. Get immediate medical attention.	For inhalation, consider glove hygiene.
Nitrogen	Wash exposed skin with soap and water.	Flush eyes with plenty of water.	If a large amount is swallowed, get medical attention.	If adverse effects occur, remove to unconfined area. Give artificial respiration if not breathing. If breathing is difficult, oxygen should be administered by qualified personnel. Get immediate medical attention.	

Section 5: Fire Fighting Measures

	Suitable Extinguishing Media	Products of Combustion	Protection of Firefighters
Ethanol	Alcohol resistant foam, carbon dioxide, regular dry chemical, water, alcohol resistant foam Large fires: Use alcohol-resistant foam or foam with fine water spray.	Carbon monoxide, carbon dioxide, and toxic and irritating fumes	Any supplied air respirator with full facepiece and operated in a pressure-demand or other positive-pressure mode in combination with a suitable escape supply. Any supplied air respirator with full facepiece and operated in a pressure-demand or other positive-pressure mode in combination with a suitable escape supply. Respiratory protection may be needed for frequent or heavy exposure.
Nitrogen	Non-flammable. Use suitable extinguishing media for surrounding fire. Cylinders may rupture or explode if exposed to heat.	Non-flammable	

ILMO Products Company
 Generated by the MSDS Manager from AsterRisk, LLC. All Rights Reserved.

page 2 of 6
 Generated: 07/20/2012 09:38:09

Section 12: Ecological Information

Fate and Transport

Ecotoxicity	Persistence / Degradability	Bioaccumulation / Accumulation	Mobility in Environment
Ethanol Fish toxicity: 93 ug/L 96 hour(s) LC50 (Moranally) Drosophila (Lepomis macrochirus) Invertebrate toxicity: 24 ug/L 48 hour(s) EC50 (Immobilization) Water flea (Daphnia pulex) Algal toxicity: 10000-25000 ug/L 1-2 hour(s) (Photosynthesis) Green algae (Acrospira sp.) Phyco-synthesis: Not available Other toxicity: Not available	Highly toxic to aquatic life.	Not available	Not available
Nitrogen Fish toxicity: Not available Invertebrate toxicity: Not available Algal toxicity: Not available Phyco toxicity: Not available Other toxicity: Not available	Not available	Not available	Not available

Section 13: Disposal Considerations

Ethanol	Subject to disposal regulations: U.S. EPA 410 CFR 262, Hazardous Waste Number(s): D001. Dispose in accordance with all applicable regulations.
Nitrogen	Dispose in accordance with all applicable regulations.

Section 14: Transportation Information

U.S. DOT 49 CFR 172.101

Proper Shipping Name	ID Number	Hazard Class or Division	Packing Group	Labeling Requirements	Passenger Aircraft or Quantity Limitations	Cargo Aircraft Quantity Limitations	Additional Shipping Description
Ethanol ALCOHOL or ETHANOL SOLUTIONS or ETHYL ALCOHOL SOLUTIONS	UN1170	3	II, III	3	N/A	N/A	N/A
Nitrogen compressed	UN1066	2.2	Not applicable	2.2	75 kg or L	150 kg	N/A

Canadian Transportation of Dangerous Goods

Shipping Name	UN Number	Class	Packing Group / Risk Group
Ethanol	UN1170	3	II
Nitrogen	Nitrogen, compressed	UN1066	2.2 Not applicable

Section 15: Regulatory Information

U.S. Regulations

Chemical	CFR Section	SARA 302	SARA 303	SARA 304
Ethanol	Not regulated.	Not regulated.	Not regulated.	Not regulated.
Nitrogen	Not regulated.	Not regulated.	Not regulated.	Not regulated.

SARA 370.21

Acute	Chronic	Fate	Reactive	Stability
Ethanol	Yes	Yes	No	No
Nitrogen	Yes	No	No	Yes

ILMO Products Company
Generated by the MSDS Manager from Astorisk, LLC All Rights Reserved

SARA 372.65

Ethanol	Not regulated.
Nitrogen	Not regulated.

OSHA Process Safety

Ethanol	Not regulated.
Nitrogen	Not regulated.

State Regulations

Ethanol	CA Proposition 65
Nitrogen	Not regulated.

Canadian Regulations

Ethanol	WHMIS Classification
Nitrogen	A

National Inventory Status

US Inventory (TSCA)	TSCA 12b Export Notification	Canada Inventory (DSL/NDSL)
Ethanol	Listed on inventory.	Not listed.
Nitrogen	Listed on inventory.	Listed on inventory.

Section 16: Other Information

NFPA Rating

Ethanol	HEALTH=2 FIRE=3 REACTIVITY=0
Nitrogen	HEALTH=1 FIRE=0 REACTIVITY=0

0 = minimal hazard, 1 = slight hazard, 2 = moderate hazard, 3 = severe hazard, 4 = extreme hazard

ILMO Products Company
Generated by the MSDS Manager from Astorisk, LLC All Rights Reserved

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 08/06/2015		ARREST TIME 0800		COUNTY # 55	CITY # 12	CITATION # 1234		
ARREST LOCATION MOCK				CITY MOCK		COUNTY OKLAHOMA		
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 11/29/1979	SEX M	RACE W	HEIGHT 511	WEIGHT 190
ADDRESS MOCK			CITY MOCK		STATE OK	ZIP CODE 00000		
DRIVER LICENSE # MOCK		EXPIRATION DATE 08/08/2016		STATE OK	CLASS D	RESTRICTIONS NONE	ENDORSEMENTS NONE	
VEHICLE MAKE NONE		MODEL NONE		TAG # NONE		STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. POST REPAIR. CANISTER REINSTALLED.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-005136	SPECIALIST BOARD OF TESTS																																								
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 16837																																								
MANUFACTURER ILMO		LOT # 13314080A1	DATE CYLINDER INSTALLED 08/05/2015	DATE CYLINDER EXPIRES 06/05/2016																																							
OBSERVATION START 0810	TEST DATE 08/06/2015	MOUTHPIECE G	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td colspan="3">-----</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:35</td></tr> <tr><td>Subject Test</td><td>0.10</td><td>08:35</td></tr> <tr><td> Breath Volume</td><td>1.902 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:36</td></tr> <tr><td>Wait</td><td></td><td>08:37</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:38</td></tr> <tr><td>Subject Test</td><td>0.10</td><td>08:38</td></tr> <tr><td> Breath Volume</td><td>1.550 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:39</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>08:39</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:40</td></tr> </tbody> </table>					Test	g/210L	Time	-----			Air Blank	0.00	08:35	Subject Test	0.10	08:35	Breath Volume	1.902 LITERS		Air Blank	0.00	08:36	Wait		08:37	Air Blank	0.00	08:38	Subject Test	0.10	08:38	Breath Volume	1.550 LITERS		Air Blank	0.00	08:39	Cal Check	0.08	08:39	Air Blank	0.00	08:40
Test	g/210L	Time																																									

Air Blank	0.00	08:35																																									
Subject Test	0.10	08:35																																									
Breath Volume	1.902 LITERS																																										
Air Blank	0.00	08:36																																									
Wait		08:37																																									
Air Blank	0.00	08:38																																									
Subject Test	0.10	08:38																																									
Breath Volume	1.550 LITERS																																										
Air Blank	0.00	08:39																																									
Cal Check	0.08	08:39																																									
Air Blank	0.00	08:40																																									
RESULT TABLE			NAME SMITH, JOSHUA																																								
			BADGE # 0000000435	PERMIT # 0000016837																																							
			AGENCY BOARD OF TESTS																																								

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification.** Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **08/06/2015** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **08/06/2015** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

No Temporary License Issued:
DL Invalid

OFFICER/DPS COPY
BOT Form AFF01 07/2008

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 08/06/2015		ARREST TIME 0800		COUNTY # 55	CITY # 12	CITATION # 1234		
ARREST LOCATION MOCK				CITY MOCK		COUNTY OKLAHOMA		
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 11/29/1979	SEX M	RACE W	HEIGHT 511	WEIGHT 190
ADDRESS MOCK			CITY MOCK		STATE OK	ZIP CODE 00000		
DRIVER LICENSE # MOCK		EXPIRATION DATE 08/08/2016	STATE OK	CLASS D	RESTRICTIONS NONE	ENDORSEMENTS NONE		
VEHICLE MAKE NONE		MODEL NONE	TAG # NONE		STATE OK	CMV N	PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. POST REPAIR. CANISTER REINSTALLED.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-005136	SPECIALIST BOARD OF TESTS																																								
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 16837																																								
MANUFACTURER ILMO		LOT # 13314080A1	DATE CYLINDER INSTALLED 08/05/2015	DATE CYLINDER EXPIRES 06/05/2016																																							
OBSERVATION START 0810	TEST DATE 08/06/2015	MOUTHPIECE G	<p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td colspan="3">-----</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:35</td></tr> <tr><td>Subject Test</td><td>0.10</td><td>08:35</td></tr> <tr><td> Breath Volume</td><td>1.902 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:36</td></tr> <tr><td>Wait</td><td></td><td>08:37</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:38</td></tr> <tr><td>Subject Test</td><td>0.10</td><td>08:38</td></tr> <tr><td> Breath Volume</td><td>1.550 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:39</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>08:39</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>08:40</td></tr> </tbody> </table>					Test	g/210L	Time	-----			Air Blank	0.00	08:35	Subject Test	0.10	08:35	Breath Volume	1.902 LITERS		Air Blank	0.00	08:36	Wait		08:37	Air Blank	0.00	08:38	Subject Test	0.10	08:38	Breath Volume	1.550 LITERS		Air Blank	0.00	08:39	Cal Check	0.08	08:39	Air Blank	0.00	08:40
Test	g/210L	Time																																									

Air Blank	0.00	08:35																																									
Subject Test	0.10	08:35																																									
Breath Volume	1.902 LITERS																																										
Air Blank	0.00	08:36																																									
Wait		08:37																																									
Air Blank	0.00	08:38																																									
Subject Test	0.10	08:38																																									
Breath Volume	1.550 LITERS																																										
Air Blank	0.00	08:39																																									
Cal Check	0.08	08:39																																									
Air Blank	0.00	08:40																																									
RESULT TABLE			<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> Test Result BrAC g/210L 0.10 </div>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>NAME SMITH, JOSHUA</td> <td>PERMIT # 0000016837</td> </tr> <tr> <td>BADGE # 0000000435</td> <td>AGENCY BOARD OF TESTS</td> </tr> </table>					NAME SMITH, JOSHUA	PERMIT # 0000016837	BADGE # 0000000435	AGENCY BOARD OF TESTS																																			
NAME SMITH, JOSHUA	PERMIT # 0000016837																																										
BADGE # 0000000435	AGENCY BOARD OF TESTS																																										

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **08/06/2015** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **08/06/2015** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

No Temporary License Issued:
DL Invalid

ARRESTEE'S COPY
BOT Form AFF02 07/2008



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 17, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0232% (w/vol) ethyl alcohol. The expiration date for this lot number is April 15, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN070212-01 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14010 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 8, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0496% (w/vol) ethyl alcohol. The expiration date for this lot number is January 6, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.040 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN120110-04 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



GUTH LABORATORIES, INC.

599 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14130 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 13, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2418% (w/vol) ethyl alcohol. The expiration date for this lot number is May 12, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN032712-01 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Wet Bath Solution Certificate of Analysis

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing.



ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-002591

11/21/2014

08:43

Test	g/210L	Time
Air Blank	0.000	08:43
Cal Check	0.019	08:44
Air Blank	0.000	08:45
Cal Check	0.020	08:45
Air Blank	0.000	08:46
Cal Check	0.019	08:47
Air Blank	0.000	08:47
Cal Check	0.020	08:48
Air Blank	0.000	08:48
Cal Check	0.020	08:49
Air Blank	0.000	08:50
Cal Check	0.020	08:50
Air Blank	0.000	08:51
Cal Check	0.021	08:52
Air Blank	0.000	08:52
Cal Check	0.020	08:53
Air Blank	0.000	08:53
Cal Check	0.020	08:54
Air Blank	0.000	08:55
Cal Check	0.021	08:55
Air Blank	0.000	08:56
Cal Check Stats		
Average	0.0200	
Std Dev	0.0006	
Rel Std Dev (%)	3.3331	

Operator's Signature

Manufacturer: GUTH LABORATORIES	Manufactured Date: 04/15/2014	Concentration: 0.02
Lot No.: 14100	Expiration Date: 04/15/2016	
The above tests were performed in accordance with procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. Based upon the attached manufacturer's certificate of analysis and the above tests, the referenced lot of simulator solution meets the standards for use in the maintenance procedures prescribed by the State Director of Tests.		
Signature	Date: 11/21/2014	
Name JOSHUA SMITH	Permit No.: 16837	

Wet Bath Solution Certificate of Analysis

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by assuring the accuracy and scientific reliability
of blood and breath testing.



ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-002591

11/21/2014


08:59

Test	g/210L	Time
Air Blank	0.000	08:59
Cal Check	0.041	09:00
Air Blank	0.000	09:00
Cal Check	0.042	09:01
Air Blank	0.000	09:02
Cal Check	0.041	09:02
Air Blank	0.000	09:03
Cal Check	0.042	09:04
Air Blank	0.000	09:04
Cal Check	0.041	09:05
Air Blank	0.000	09:05
Cal Check	0.041	09:06
Air Blank	0.000	09:07
Cal Check	0.040	09:07
Air Blank	0.000	09:08
Cal Check	0.041	09:09
Air Blank	0.000	09:09
Cal Check	0.041	09:10
Air Blank	0.000	09:10
Cal Check	0.042	09:11
Air Blank	0.000	09:12
Cal Check Stats		
Average	0.0412	
Std Dev	0.0006	
Rel Std Dev(%)	1.5350	

Operator's Signature

Manufacturer: GUTH LABORATORIES	Manufactured Date: 01/06/2014	Concentration: 0.04
Lot No.: 14010	Expiration Date: 01/16/2016	
The above tests were performed in accordance with procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. Based upon the attached manufacturer's certificate of analysis and the above tests, the referenced lot of simulator solution meets the standards for use in the maintenance procedures prescribed by the State Director of Tests.		
Signature	Date: 11/21/2014	
Name JOSHUA SMITH	Permit No.: 16837	

Wet Bath Solution Certificate of Analysis




ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 09/02/2014 09:26

Test	g/210L	Time
Air Blank	0.000	09:27
Cal Check	0.102	09:27
Air Blank	0.000	09:28
Cal Check	0.103	09:29
Air Blank	0.000	09:29
Cal Check	0.102	09:30
Air Blank	0.000	09:31
Cal Check	0.102	09:31
Air Blank	0.000	09:32
Cal Check	0.102	09:33
Air Blank	0.000	09:33
Cal Check	0.102	09:34
Air Blank	0.000	09:34
Cal Check	0.103	09:35
Air Blank	0.000	09:36
Cal Check	0.103	09:36
Air Blank	0.000	09:37
Cal Check	0.103	09:38
Air Blank	0.000	09:38
Cal Check	0.103	09:39
Air Blank	0.000	09:39
Cal Check Stats		
Average	0.1025	
Std Dev	0.0005	
Rel Std Dev(%)	0.5141	

Operator's Signature

Manufacturer: GUTH LABORATORIES	Manufactured Date: 05/01/2014	Concentration: 0.10
Lot No.: 14110	Expiration Date: 05/01/2016	
The above tests were performed in accordance with procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. Based upon the attached manufacturer's certificate of analysis and the above tests, the referenced lot of simulator solution meets the standards for use in the maintenance procedures prescribed by the State Director of Tests.		
Signature	Date: 09/02/2014	
Name JOSHUA SMITH	Permit No.: 16837	

Wet Bath Solution Certificate of Analysis



ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 08/29/2014 15:40

Test	g/210L	Time
Air Blank	0.000	15:41
Cal Check	0.204	15:41
Air Blank	0.000	15:42
Cal Check	0.204	15:43
Air Blank	0.000	15:43
Cal Check	0.204	15:44
Air Blank	0.000	15:44
Cal Check	0.203	15:45
Air Blank	0.000	15:46
Cal Check	0.203	15:46
Air Blank	0.000	15:47
Cal Check	0.202	15:48
Air Blank	0.000	15:48
Cal Check	0.203	15:49
Air Blank	0.000	15:49
Cal Check	0.203	15:50
Air Blank	0.000	15:51
Cal Check	0.203	15:51
Air Blank	0.000	15:52
Cal Check	0.203	15:53
Air Blank	0.000	15:53
Cal Check Stats		
Average	0.2032	
Std Dev	0.0006	
Rel Std Dev(%)	0.3112	

Operator's Signature

Manufacturer: GUTH LABORATORIES	Manufactured Date: 05/12/2014	Concentration: 0.20
Lot No.: 14130	Expiration Date: 05/12/2016	
The above tests were performed in accordance with procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. Based upon the attached manufacturer's certificate of analysis and the above tests, the referenced lot of simulator solution meets the standards for use in the maintenance procedures prescribed by the State Director of Tests.		
Signature	Date: 08/29/2014	
Name JOSHUA SMITH	Permit No.: 16837	