



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Joshua Smith  
Director

**ATTESTATION**

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number \_\_\_\_\_, in compliance with the agency's Maintenance Policy and Procedures.

This packet consists of \_\_\_\_\_ pages.

Signed 

Name of Position: Director/Records Custodian

Date of Attestation:



# Service Overview

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

**BENCH CHECK DATE:** 06-28-2021

**SERIAL:** 80-005145

## **DOCUMENTATION WITHIN SERVICE PACKET:**

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)



### STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Administrative Office at:  
Board of Tests for Alcohol and Drug Influence  
Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307

## Certificate of Calibration and Operation

This is to certify that the calibration of **INTOXILYZER 8000**, serial number 80-005145, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R 126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.

In addition the operational procedures and functions of this Intoxilyzer have also been tested and found to be in compliance with the Oklahoma Board of Tests for Alcohol and Drug Influence - Rules and Regulations OAC Title 40: 25-1-2 and 30-1-3 and is hereby certified for use as an evidential breath testing device within this state.

CERTIFICATION TYPE	DATE
Manufacturer Calibration	08-24-10
Board of Tests Bench Check	09-01-10
Board of Tests Bench Check	07-10-12
Manufacturer Calibration	02-25-14
Board of Tests Bench Check	03-04-14
Board of Tests Bench Check	11-18-15
Manufacturer Calibration	11-14-17
Board of Tests Bench Check	03-01-18
Board of Tests Bench Check	07-03-19
Board of Tests Bench Check	06-28-21



Presentation of this form certifies that the Intoxilyzer, SN: 80-005145, functioned properly at the time of the breath test and hereby certifies the breath test result as valid.

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)  
State Director of Tests for Alcohol  
and Drug Influence

**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE  
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 06/28/2021	TIME PROCEDURE BEGAN 13:47:5	INSTRUMENT LOCATION 80-005145
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-005145

**TEST RESULTS**

**Diagnostics**

PASS

**Mouth Alcohol Test**

PASS

**RFI Test**

PASS

**Abort Test**

PASS

**Improper Sample Test**

PASS

**Interferent Detect Test**

PASS

**No Sample Given Test**

PASS

**Insufficient Sample Test**

PASS

**Air Blank Contamination Test**

PASS

**Range Exceeded Test**

PASS

**Duplicate Breath Sampling Results**


Value	Test 1	Test 2	Control Test
.02	0.017	0.017	0.018
.04	0.037	0.037	0.037
.10	0.096	0.097	0.096
.20	0.195	0.194	0.194

PASS

**CERTIFYING TECHNICIAN**

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE



DATE

06/28/2021

NAME

SMITH, JOSHUA C

PERMIT #

16837





3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

## BENCH CHECK TECHNICIAN'S REPORT

**Date:** 06/28/2021

**Start Time:** 13:40

**End Time:** 15:30

### INSTRUMENT

Intoxilyzer Model

Serial #

8000	80-005145
------	-----------

### GAS CANISTER

LOT #

EXP Date

11821080A2-006	06-05-2023
----------------	------------

### REFERENCE

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR2834	DR3700	DR2455	DR3470
Concentration			
.02	.04	.10	.20
LOT #			
20070	20060	20190	20160
Manufactured Date			
02/13/2020	02/10/2020	04/06/2020	03/18/2020
Expiration Date			
02/13/2022	02/10/2022	04/06/2022	03/18/2022
Solution Commission Date			
6/3/2021	6/3/2021	6/18/2021	6/18/2021

### REASON FOR BENCH CHECK

- POST REPAIR     CYLINDER REPLACEMENT  
 TROUBLESHOOTING     ROUTINE MAINTENANCE

### COMMENTS


Technician Signature / Permit #

Joshua Smith

#16837

Printed Name

**Certificate of Analysis – Dry Gas**

6/17/2021	0.080
Date of Analysis	Labelled target value (g/210L)
006	0.081
Cylinder #	Average test result
Christopher Page <i>Christopher Page</i> 23020	
30T Technician Name and Signature	



7/13

ALPHA INSTRUMENT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002591  
06/17/2021 08:59

In accordance with BOT Policy and Procedure Statement BRT-2.2.0, Maintenance, the above referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.  
7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Test	g/210L	Time
Air Blank	0.000	09:00
Cal Check	0.081	09:00
Air Blank	0.000	09:00
Cal Check	0.081	09:01
Air Blank	0.000	09:01
Cal Check	0.081	09:02
Air Blank	0.000	09:02
Cal Check	0.081	09:02
Air Blank	0.000	09:03
Cal Check	0.080	09:03
Air Blank	0.000	09:04
Cal Check	0.081	09:04
Air Blank	0.000	09:05
Cal Check	0.081	09:05
Air Blank	0.000	09:05
Cal Check	0.081	09:06
Air Blank	0.000	09:06
Cal Check	0.080	09:07
Air Blank	0.000	09:07
Cal Check	0.082	09:07
Air Blank	0.000	09:08
Cal Check Stats		
Average	0.0809	
Std Dev	0.0005	
Rel Std Dev(%)	0.7016	

**Certificate of Analysis**

**Certificate ID:** 13921  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 11821080A2  
**Expiration:** 6/5/2023

*RCWD  
6/10/21  
CP*

**0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)**

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

**Distributed by:**  
CMI Inc.  
316 East Ninth St  
Owensboro, KY 42  
Phone 866-835-069  
[www.alcoholtest.co](http://www.alcoholtest.co)

*Christopher Page*  
Operator's Signature

\*Traceable to:  
Certified Reference Material - 262.4 µmol/mol  
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

*Josh Lewis*  
Specialty Gas Lab Tech  
*Travis Nelson*

05-27-2021  
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

8/13

Section 1

ARREST DATE <b>06/28/2021</b>		ARREST TIME <b>1400</b>		COUNTY # <b>55</b>	CITY # <b>29</b>	CITATION # <b>MOCK</b>		
ARREST LOCATION <b>NONE</b>				CITY <b>16</b>	COUNTY <b>OKLAHOMA</b>			
SUBJECT NAME <b>MOCK, SUBJECT TEST</b>				DATE OF BIRTH <b>09/10/2001</b>	SEX <b>F</b>	RACE <b>W</b>	HEIGHT <b>510</b>	WEIGHT <b>210</b>
ADDRESS <b>MOCK</b>			CITY <b>MOCK</b>	STATE <b>OK</b>	ZIP CODE <b>99999</b>			
DRIVER LICENSE # <b>MOCK</b>		EXPIRATION DATE <b>08/01/2021</b>		STATE <b>OK</b>	CLASS <b>D</b>	RESTRICTIONS <b>NONE</b>	ENDORSEMENTS <b>NONE</b>	
VEHICLE MAKE <b>MOCK</b>		MODEL <b>MOCK</b>		TAG # <b>MOCK</b>	STATE <b>OK</b>	CMV <b>N</b>	PLACARD <b>N</b>	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

**MOCK SUBJECT TEST**


Describe person's condition (odor, actions, etc.):

**ROUTINE MAINTENANCE, NEW GAS CANISTER INSTALL.**

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD

Section 2

EQUIPMENT <b>I-8000</b>		SERIAL NUMBER <b>80-005145</b>	SPECIALIST <b>BOARD OF TESTS</b>	
STANDARD <b>Dry</b>		TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>16837</b>	
MANUFACTURER <b>ILMO</b>		LOT # <b>11821080A2</b>	DATE CYLINDER INSTALLED <b>06/28/2021</b>	DATE CYLINDER EXPIRES <b>06/05/2023</b>
OBSERVATION START <b>1430</b>	TEST DATE <b>06/28/2021</b>	MOUTHPIECE <b>G</b>	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.	
RESULT TABLE				
Test	g/210L	Time	 <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p>SIGNATURE OF OPERATOR _____</p> <p>NAME <b>SMITH, JOSHUA C</b></p> <p>BADGE # <b>435</b></p> <p>PERMIT # <b>16837</b></p> <p>AGENCY <b>BOT</b></p>	
Air Blank	0.00	15:11		
Subject Test	0.00	15:12		
Breath Volume	1.511 LITERS			
Air Blank	0.00	15:14		
Wait		15:14		
Air Blank	0.00	15:14		
Subject Test	0.00	15:15		
Breath Volume	1.800 LITERS			
Air Blank	0.00	15:15		
Cal Check	0.00	15:15		
Air Blank	0.00	15:16		

Test Result BrAC g/210L <h2 style="margin: 0;">0.00</h2>
--

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED.** Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

Section 3

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **06/28/2021** Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **06/28/2021** Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME <b>SMITH, JOSHUA C</b>		AGENCY <b>BOT</b>	
BADGE <b>435</b>	AGENCY ADDRESS	ZIP CODE	PHONE

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

No Temporary License Issued:  
Test Below Legal Limit

OFFICER/DPS COPY  
BOT Form AFF01 03/2020



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

9/13

Section 1

ARREST DATE <b>06/28/2021</b>		ARREST TIME <b>1400</b>		COUNTY # <b>55</b>	CITY # <b>29</b>	CITATION # <b>MOCK</b>		
ARREST LOCATION <b>NONE</b>				CITY <b>16</b>	COUNTY <b>OKLAHOMA</b>			
SUBJECT NAME <b>MOCK, SUBJECT TEST</b>				DATE OF BIRTH <b>09/10/2001</b>	SEX <b>F</b>	RACE <b>W</b>	HEIGHT <b>510</b>	WEIGHT <b>210</b>
ADDRESS <b>MOCK</b>			CITY <b>MOCK</b>	STATE <b>OK</b>		ZIP CODE <b>99999</b>		
DRIVER LICENSE # <b>MOCK</b>			EXPIRATION DATE <b>08/01/2021</b>	STATE <b>OK</b>	CLASS <b>D</b>	RESTRICTIONS <b>NONE</b>	ENDORSEMENTS <b>NONE</b>	
VEHICLE MAKE <b>MOCK</b>		MODEL <b>MOCK</b>		TAG # <b>MOCK</b>	STATE <b>OK</b>	CMV <b>N</b>	PLACARD <b>N</b>	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

**MOCK SUBJECT TEST**


Describe person's condition (odor, actions, etc.):

**ROUTINE MAINTENANCE, NEW GAS CANISTER INSTALL.**

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD

Section 2

EQUIPMENT <b>I-8000</b>		SERIAL NUMBER <b>80-005145</b>	SPECIALIST <b>BOARD OF TESTS</b>
STANDARD <b>Dry</b>		TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>16837</b>
MANUFACTURER <b>ILMO</b>		LOT # <b>11821080A2</b>	DATE CYLINDER INSTALLED <b>06/28/2021</b>
			DATE CYLINDER EXPIRES <b>06/05/2023</b>
OBSERVATION START <b>1430</b>	TEST DATE <b>06/28/2021</b>	MOUTHPIECE <b>G</b>	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.
RESULT TABLE			
Test	g/210L	Time	 I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.  SIGNATURE OF OPERATOR  NAME <b>SMITH, JOSHUA C</b> BADGE # <b>435</b> PERMIT # <b>16837</b> AGENCY <b>BOT</b>
Air Blank	0.00	15:13	
Subject Test	0.00	15:12	
Breath Volume	1.511 LITERS		
Air Blank	0.00	15:13	
Wait		15:14	
Air Blank	0.00	15:14	
Subject Test	0.00	15:15	
Breath Volume	1.890 LITERS		
Air Blank	0.00	15:15	
Cal Check	0.07	15:15	
Air Blank	0.00	15:16	



Test Result BrAC g/210L <h2 style="margin: 0;">0.00</h2>
--

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED.** Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

Section 3

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **06/28/2021** Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4


Date: **06/28/2021** Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME <b>SMITH, JOSHUA C</b>		AGENCY <b>BOT</b>	
BADGE <b>435</b>	AGENCY ADDRESS	ZIP CODE	PHONE

No Temporary License Issued:  
Test Below Legal Limit

ARRESTEE'S COPY  
BOT Form AFF02 03/2020



Certificate of Analysis - Wet Bath	
8/11/2020	0.020
Date of Analysis	Labelled Value (g/210L)
20070	0.020
Lot Number	Average Credit (g/210L)
	
BOT Technician Name and Signature	

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

*ACV'D  
6/1/2020  
CP*

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 00-004923  
08/11/2020 09:42


**Certified Alcohol Reference Solution for Simulator**


Test	g/210L	Time
Air Blank	0.000	09:42
Cal Check	0.020	09:43
Air Blank	0.000	09:43
Cal Check	0.020	09:44
Air Blank	0.000	09:45
Cal Check	0.020	09:45
Air Blank	0.000	09:46
Cal Check	0.020	09:47
Air Blank	0.000	09:47
Cal Check	0.021	09:48
Air Blank	0.000	09:48
Cal Check	0.020	09:49
Air Blank	0.000	09:49
Cal Check	0.020	09:50
Air Blank	0.000	09:51
Cal Check	0.020	09:51
Air Blank	0.000	09:52
Cal Check	0.020	09:53
Air Blank	0.000	09:53
Cal Check	0.019	09:54
Air Blank	0.000	09:54
Cal Check Stats		
Average	0.0200	
Std Dev	0.0004	
Rel Std Dev(%)	2.3569	

Random Samples of Lot Number **20070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 17, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0242% (w/vol)** ethyl alcohol. The expiration date for this lot number is **February 13, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.020 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.



  
Operator's Signature

  
Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number **FN08031603** whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Certificate of Analysis - Wet Bath	
7/22/2020	0.040
Date of Analysis	Labelled target value (g/210L)
20060	0.040
Lot Number	Average test result (g/210L)
	
	



**GUTH LABORATORIES, INC.**  
 200 NORTH 57th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

Operator Name and Signature: Christopher Pope  
 BRT Form 6.0  
 In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

*RCV'D  
6/1/20  
CP*

ALPHA UNIT SUB  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003400  
 07/22/2020 14:57

Random Samples of Lot Number **20060** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 11, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0483%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 10, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Test	g/210L	Time
Air Blank	0.000	14:57
Cal Check	0.040	14:58
Air Blank	0.000	14:59
Cal Check	0.039	14:59
Air Blank	0.000	15:00
Cal Check	0.039	15:01
Air Blank	0.000	15:01
Cal Check	0.039	15:02
Air Blank	0.000	15:02
Cal Check	0.039	15:03
Air Blank	0.000	15:04
Cal Check	0.039	15:04
Air Blank	0.000	15:05
Cal Check	0.040	15:06
Air Blank	0.000	15:06
Cal Check	0.040	15:07
Air Blank	0.000	15:07
Cal Check	0.040	15:08
Air Blank	0.000	15:09
Cal Check	0.040	15:09
Air Blank	0.000	15:10
Cal Check Stats		
Average	0.0395	
Std Dev	0.0005	
Rel Std Dev(%)	1.3342	

*Ted L. Pauley*

Ted L. Pauley, President  
 GUTH LABORATORIES, INC.

*Christopher Pope*  
 Operator's Signature

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number **FN12181501** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis - Wet Bath



12/13

06/16/20	.100
Date of Analysis	Labelled target value (g/210L)
20190	.098
Lot Number	Actual test result (g/210L)

John L. Smith  
BOT Technician Name and Signature

GUTH LABORATORIES, INC.

BRT-Form 6.0 USE JRG, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

Rec'd  
06/11/2020  
*[Signature]*

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/16/2020

SN 80-002591  
16:18

Certified Alcohol Reference Solution for Simulator

Test	g/210L	Time
Air Blank	0.000	16:18
Cal Check	0.099	16:19
Air Blank	0.000	16:20
Cal Check	0.098	16:20
Air Blank	0.000	16:21
Cal Check	0.097	16:22
Air Blank	0.000	16:22
Cal Check	0.098	16:23
Air Blank	0.000	16:24
Cal Check	0.098	16:24
Air Blank	0.000	16:25
Cal Check	0.098	16:26
Air Blank	0.000	16:26
Cal Check	0.098	16:27
Air Blank	0.000	16:27
Cal Check	0.097	16:28
Air Blank	0.000	16:29
Cal Check	0.098	16:29
Air Blank	0.000	16:30
Cal Check	0.098	16:31
Air Blank	0.000	16:31
Cal Check Stats		
Average	0.0979	
Std Dev	0.0005	
Rel Std Dev(%)	0.5798	

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

*[Signature]*  
Operator's Signature

*[Signature]*  
Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**  
Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.  
All balances are calibrated annually by an outside agency using NIST traceable weights.  
Calibration verification is done prior to each use utilizing NIST traceable weights.





Certificate of Analysis - Wet Bath	
07/07/2020	0.200
Date of Analysis	Labeled Test Value (g/210L)
20160	0.197
Lot Number	Average Test Result (g/210L)
Christopher Page	0.2030
BOT Technician Name and Signature	

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

### CERTIFICATE OF ANALYSIS

REC'D  
05/2020  
CP

Certified Alcohol Reference Solution for Simulator

INSTRUMENT SUB  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003400  
07/2020 15:25

Test	g/210L	Time
r Blank	0.000	15:26
i Check	0.196	15:26
r Blank	0.000	15:27
i Check	0.196	15:28
r Blank	0.000	15:28
i Check	0.197	15:29
r Blank	0.000	15:29
i Check	0.196	15:30
r Blank	0.000	15:31
i Check	0.198	15:31
r Blank	0.000	15:32
i Check	0.198	15:33
r Blank	0.000	15:33
i Check	0.197	15:34
r Blank	0.000	15:35
i Check	0.198	15:35
r Blank	0.000	15:36
i Check	0.198	15:37
r Blank	0.000	15:37
i Check	0.197	15:38
r Blank	0.000	15:38
i Check Stats		
Average	0.1971	
Std Dev	0.0008	
rel Std Dev(%)	0.4442	

Random Samples of Lot Number **20160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 20, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2427% (w/vol) ethyl alcohol. The expiration date for this lot number is March 18, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

*Christopher Page*  
Operator's Signature

*Ted L. Pauley*

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

#### NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number EN08101505 whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.