



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Joshua Smith  
Director

**ATTESTATION**

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-006139, in compliance with the agency's Maintenance Policy and Procedures.

Signed 

Name of Position: Director/Records Custodian

Date of Attestation: 12/13/2024

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing



2/17

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

**BENCH CHECK DATE:** 12/13/2024

**SERIAL:** 80-006139

## **DOCUMENTATION WITHIN SERVICE PACKET:**

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis on Certified Solutions (4)
- Guth Laboratories Certificate of Analysis (4)



# STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

## CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION

This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.


- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-006139 on 12/13/2024 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 28424080A3-010
  - BOT analysis average test result: 0.082
  - Cylinder expiration date: 11/5/2026
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model	Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #	DR3591	DR3694	DR2455	DR3641
Concentration	.02	.04	.10	.20
LOT #	23360	23400	24110	23340
Manufactured Date	9/25/2023	10/24/2023	3/5/2024	9/18/2023
Expiration Date	9/25/2025	10/24/2025	3/5/2026	9/18/2025
Solution Commission Date	11/27/2024	11/27/2024	12/13/2024	12/13/2024

Completed by BOT Personnel:

  
Signature/Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

  
Signature Joshua Smith  
Name

State Director  
Title

**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE  
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 12/13/2024	TIME PROCEDURE BEGAN 10:27:0	INSTRUMENT LOCATION 80-006139
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-006139

**TEST RESULTS**

**Diagnostics**

PASS

**Mouth Alcohol Test**

PASS

**RFI Test**

PASS

**Abort Test**

PASS

**Improper Sample Test**

PASS

**Interferent Detect Test**

PASS

**No Sample Given Test**

PASS

**Insufficient Sample Test**

PASS

**Air Blank Contamination Test**

PASS

**Range Exceeded Test**

PASS


**Duplicate Breath Sampling Results**

Value	Test 1	Test 2	Control Test
.02	0.018	0.018	0.019
.04	0.036	0.036	0.038
.10	0.096	0.095	0.097
.20	0.195	0.195	0.199

PASS

**CERTIFYING TECHNICIAN**

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 12/13/2024
NAME PAPE, CHRISTOPHER	PERMIT # 000023020



5/17

### STATE OF OKLAHOMA BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Joshua Smith  
Director

#### Certificate of Analysis – Dry Gas

##### ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000


SN 80-002591

11/14/2024

11:41

Test	g/210L	Time
Air Blank	0.000	11:42
Cal Check	0.081	11:42
Air Blank	0.000	11:42
Cal Check	0.082	11:43
Air Blank	0.000	11:43
Cal Check	0.082	11:44
Air Blank	0.000	11:44
Cal Check	0.082	11:44
Air Blank	0.000	11:45
Cal Check	0.081	11:45
Air Blank	0.000	11:46
Cal Check	0.081	11:46
Air Blank	0.000	11:47
Cal Check	0.082	11:47
Air Blank	0.000	11:47
Cal Check	0.082	11:48
Air Blank	0.000	11:48
Cal Check	0.082	11:49
Air Blank	0.000	11:49
Cal Check	0.081	11:50
Air Blank	0.000	11:50
Cal Check Stats		
Average	0.0816	
Std Dev	0.0005	
Rel Std Dev (%)	0.6328	

Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)  
0.08 g/210L = 208 PPM + 2605\*

Manufacturer: <b>ILMO</b>	Manufacturer Cert. No.: <b>16873</b>	Cylinder Size: <b>105L</b>
Lot No.: <b>28424080A3</b>	Cylinder No.: <b>010</b>	Expiration Date: <b>11/05/2026</b>
The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM +2605].		
Signature 	Date: <b>11/14/2024</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
 217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

## Certificate of Analysis

Certificate ID: 16873  
 Part #: BAC105L080T  
 Cylinder Size: 105L  
 Lot Number: 28424080A3  
 Expiration: 11/5/2026

RCVD  
 11/13/24  
 CP

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)


Component:	Reported Concentration:	Analytical Accuracy (U, k=2): abs	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.  
 316 East Ninth Street  
 Owensboro, KY 42303  
 Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

\*Reference Standard:  
 NIST Traceable Certified Reference Material - 262.3 µmol/mol  
 Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

  
 Specialty Gas Lab Tech

10-18-2024  
 Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

# OFFICER'S IMPAIRED DRIVING AFFIDAVIT

7/17

Section 1

ARREST DATE 12/13/2024		ARREST TIME 1230		COUNTY # 55		CITY #		CITATION #	
ARREST LOCATION BOT				CITY				COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, FRANK				DATE OF BIRTH 03/15/1981		SEX M		HEIGHT 509	WEIGHT 175
ADDRESS 3600 MARTIN LUTHER KING				CITY OKLAHOMA CITY				STATE OK	ZIP CODE 73111
DRIVER LICENSE # H083463197			EXPIRATION DATE 10/31/2022		STATE OK	CLASS D	RESTRICTIONS	ENDORSEMENTS	
VEHICLE MAKE MOCK			MODEL MOCK		TAG # MOCK		STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):


ROUTINE MAINTENANCE, INSTALLED NEW CYLINDER

**THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST**

**BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.**

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006139		SPECIALIST BOARD OF TESTS		
STANDARD Dry		TARGET VALUE 0.080		PERMIT NUMBER 23020		
MANUFACTURER ILMO		LOT # 28424080A3		DATE CYLINDER INSTALLED 12/13/2024	DATE CYLINDER EXPIRES 11/05/2026	
OBSERVATION START 1235	TEST DATE 12/13/2024	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.				
RESULT TABLE		<p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p>  <p>SIGNATURE OF OPERATOR</p>				
Test	g/210L					Time
Air Blank	0.00					12:58
Subject Test	0.00					12:59
Breath Volume	1.734 LITERS					
Air Blank	0.00					13:00
Wait						13:01
Air Blank	0.00					13:02
Subject Test	0.00					13:02
Breath Volume	1.675 LITERS					
Air Blank	0.00	13:03				
Cal Check	0.08	13:03				
Air Blank	0.00	13:03				



Test Result  
BrAC g/210L  
**0.00**

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE # 0000000484	PERMIT # 0000023020		
AGENCY BOARD OF TESTS			

Section 3

**What happens next?**

1. Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges.
2. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle.
3. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
4. The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma.  
To confirm or change your address on file, contact Service Oklahoma.
5. If you possess a valid class D license and this is your first impaired driving offense, you may waive your right to appeal and apply for BOT-IDAP within 30 days of arrest to avoid a revocation appearing on your driving record. To begin the process, submit an application with this document at <https://ldap.ok.gov>.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

**OTHER WITNESSES:**

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

# OFFICER'S IMPAIRED DRIVING AFFIDAVIT

8/17

Section 1

ARREST DATE <b>12/13/2024</b>		ARREST TIME <b>1230</b>		COUNTY # <b>55</b>	CITY #	CITATION #	
ARREST LOCATION <b>BOT</b>				CITY		COUNTY <b>OKLAHOMA</b>	
SUBJECT NAME <b>BREATHTEST, FRANK</b>				DATE OF BIRTH <b>03/15/1981</b>		SEX <b>M</b>	HEIGHT <b>509</b>
ADDRESS <b>3600 MARTIN LUTHER KING</b>				CITY <b>OKLAHOMA CITY</b>		STATE <b>OK</b>	ZIP CODE <b>73111</b>
DRIVER LICENSE # <b>H083463197</b>			EXPIRATION DATE <b>10/31/2022</b>		STATE <b>OK</b>	CLASS <b>D</b>	RESTRICTIONS
VEHICLE MAKE <b>MOCK</b>		MODEL <b>MOCK</b>		TAG # <b>MOCK</b>		STATE <b>OK</b>	CMV <b>N</b>
ENDORSEMENTS <b>N</b>							

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

**MOCK SUBJECT TEST**



Describe person's condition (odor, actions, etc.):

**ROUTINE MAINTENANCE, INSTALLED NEW CYLINDER**

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT <b>I-8000</b>		SERIAL NUMBER <b>80-006139</b>	SPECIALIST <b>BOARD OF TESTS</b>	
STANDARD <b>Dry</b>		TARGET VALUE <b>0.080</b>	PERMIT NUMBER <b>23020</b>	
MANUFACTURER <b>ILMO</b>		LOT # <b>28424080A3</b>	DATE CYLINDER INSTALLED <b>12/13/2024</b>	DATE CYLINDER EXPIRES <b>11/05/2026</b>
OBSERVATION START <b>1235</b>	TEST DATE <b>12/13/2024</b>	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"> SIGNATURE OF OPERATOR</p>		
RESULT TABLE				
Test	g/210L	Time		
Air Blank	0.00	12:58		
Subject Test	0.00	12:59		
Breath Volume	1.734 LITERS			
Air Blank	0.00	13:00		
Wait		13:01		
Air Blank	0.00	13:02		
Subject Test	0.00	13:02		
Breath Volume	1.675 LITERS			
Air Blank	0.00	13:03		
Cal Check	0.08	13:03		
Air Blank	0.00	13:03		

### What happens next?

Section 3

1. Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges.
2. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle.
3. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
4. The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma.  
To confirm or change your address on file, contact Service Oklahoma.
5. If you possess a valid class D license and this is your first impaired driving offense, you may waive your right to appeal and apply for BOT-IDAP within 30 days of arrest to avoid a revocation appearing on your driving record. To begin the process, submit an application with this document at <https://idap.ok.gov>.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME <b>PAPE, CHRISTOPHER</b>		AGENCY <b>BOARD OF TESTS</b>	
BADGE <b>0000000484</b>	AGENCY ADDRESS <b>INTER-AGENCY</b>	ZIP CODE <b>00000</b>	PHONE <b>4054252460</b>





**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002591  
04/16/2024 11:24

Test	g/210L	Time
Air Blank	0.000	11:24
Cal Check	0.018	11:25
Air Blank	0.000	11:25
Cal Check	0.019	11:26
Air Blank	0.000	11:27
Cal Check	0.019	11:27
Air Blank	0.000	11:28
Cal Check	0.019	11:29
Air Blank	0.000	11:29
Cal Check	0.019	11:30
Air Blank	0.000	11:30
Cal Check	0.019	11:31
Air Blank	0.000	11:32
Cal Check	0.019	11:32
Air Blank	0.000	11:33
Cal Check	0.019	11:34
Air Blank	0.000	11:34
Cal Check	0.019	11:35
Air Blank	0.000	11:35
Cal Check	0.018	11:36
Air Blank	0.000	11:37
Cal. Check Stats:		
Average	0.0188	
Std Dev	0.0004	
Rel Std Dev (%)	2.2426	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.		
Manufacturer: Guth	Manufactured Date: 9/25/2023	Concentration: 0.02
Lot No.: 23360	Expiration Date: 9/25/2025	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature:	Date: 4/16/2024	
Name: Christopher Pape	Permit No.: 23020	

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**ACVD  
12/19/23  
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23360** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 26, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0240%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 25, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.020 g/210L**  $\pm 3\%$ .

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

Testing was conducted using Cerilliant Reference Standard lot number **FN03122113** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman


Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002591  
05/02/2024 09:26

Test	g/210L	Time
Air Blank	0.000	09:26
Cal Check	0.039	09:27
Air Blank	0.000	09:28
Cal Check	0.039	09:28
Air Blank	0.000	09:29
Cal Check	0.039	09:30
Air Blank	0.000	09:30
Cal Check	0.039	09:31
Air Blank	0.000	09:31
Cal Check	0.039	09:32
Air Blank	0.000	09:33
Cal Check	0.040	09:33
Air Blank	0.000	09:34
Cal Check	0.039	09:35
Air Blank	0.000	09:35
Cal Check	0.039	09:36
Air Blank	0.000	09:37
Cal Check	0.039	09:37
Air Blank	0.000	09:38
Cal Check	0.039	09:38
Air Blank	0.000	09:39
Cal Check Stats		
Average	0.0391	
Std Dev	0.0003	
Rel Std Dev (%)	0.8087	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.		
Manufacturer: <b>Guth</b>	Manufactured Date: <b>10/24/2023</b>	Concentration: <b>0.04</b>
Lot No.: <b>23400</b>	Expiration Date: <b>10/24/2025</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>5/2/2024</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	

**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**RCVD  
12/19/23  
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23400** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 25, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0484%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 24, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

Testing was conducted using Cerilliant Reference Standard lot number **FN06032102** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman


Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT		
Intoxilyzer - Alcohol Analyzer		
Model 8000		SN 80-002591
08/02/2024		12:55
Test	g/210L	Time
Air Blank	0.000	12:55
Cal Check	0.097	12:56
Air Blank	0.000	12:56
Cal Check	0.098	12:57
Air Blank	0.000	12:58
Cal Check	0.098	12:58
Air Blank	0.000	12:59
Cal Check	0.098	13:00
Air Blank	0.000	13:00
Cal Check	0.097	13:01
Air Blank	0.000	13:02
Cal Check	0.098	13:02
Air Blank	0.000	13:03
Cal Check	0.098	13:04
Air Blank	0.000	13:04
Cal Check	0.098	13:05
Air Blank	0.000	13:05
Cal Check	0.098	13:06
Air Blank	0.000	13:07
Cal Check	0.098	13:07
Air Blank	0.000	13:08
Cal Check Stats		
Average	0.0978	
Std Dev	0.0004	
Rel Std Dev (%)	0.4311	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: Guth	Manufactured Date: 3/5/2024	Concentration: 0.100
Lot No.: 24110	Expiration Date: 3/5/2026	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: 8/2/2024	
Name: Christopher Pape	Permit No.: 23020	

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**RCVD  
6/25/24  
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **24110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 6, 2024**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 5, 2026** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT  
Intoxilyzer – Alcohol Analyzer  
Model 8000 SN 80-002591  
10/04/2023 09:22

Test	g/210L	Time
Air Blank	0.000	09:22
Cal Check	0.198	09:23
Air Blank	0.000	09:24
Cal Check	0.198	09:24
Air Blank	0.000	09:25
Cal Check	0.199	09:26
Air Blank	0.000	09:26
Cal Check	0.198	09:27
Air Blank	0.000	09:27
Cal Check	0.198	09:28
Air Blank	0.000	09:29
Cal Check	0.198	09:29
Air Blank	0.000	09:30
Cal Check	0.198	09:31
Air Blank	0.000	09:31
Cal Check	0.198	09:32
Air Blank	0.000	09:33
Cal Check	0.199	09:33
Air Blank	0.000	09:34
Cal Check	0.199	09:35
Air Blank	0.000	09:35
Cal Check Stats		
Average	0.1983	
Std Dev	0.0004	
Rel Std Dev (%)	0.2435	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: <b>Guth</b>	Manufactured Date: <b>9/18/2023</b>	Concentration: <b>0.200</b>
Lot No.: <b>23340</b>	Expiration Date: <b>9/18/2025</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature: 	Date: <b>10/4/2023</b>	
Name: <b>Christopher Pape</b>	Permit No.: <b>23020</b>	

**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**ACVD  
10/2/2023  
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23340 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 19, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2399%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 18, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN09142002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



<u>Location Returned From</u> Mannford PD	<u>Unit Serial #</u> 80-006139	<u>Date:</u> 17/17 12/6/24
--	-----------------------------------	-------------------------------

Inventory Process

Old Cylinder #: 26422080A1-050

Unit checked in to BOT    
  Unlink Gas Canister from Unit    
  Old cylinder disposed in inventory  
 New Unit Checked out    
  Remove old cylinder from unit (After Lab Evaluation)    
  Old cylinder archived

Check in Procedure

Connect unit to ethernet connection 172.019.004, 144  
 Download any stored records 4  
 Change unit location to Serial number  
 Update Unit Date and time  
 Previous time: 12/13/24 0850  
 New Time: 12/13/24 0852  
 Execute a new Forms load on the unit

Notes from the Field

Routine Swap    
  Issues in the Field

Lab Evaluation

Breath Hose Hold Vacuum?:      Test Chamber Hold Vacuum?:

DVM

3µm 12765     Chamber Temp 47.00     Snap Test ok?   
 9µm 12924     Breath Hose Temp 45     RFI Detected?

Peripheral Tests

Speaker (F5)      Power LED Green (F6)      Red (F7)      Fan (F9)   
 Gas Regulator (F11)      Pump (F10)      Receipt paper

Atmospheric Sensor

Pressure 979     Correction Factor 1.03     Flow Sensor Resistance 182     Tank Sensor Value 868  
 Is regulator close to Tank Sensor Value?  Yes /  No

Sent to CMI (If needed)	Reason Sent
Date Sent _____	
Invoice Sent _____	
Date Returned _____	

Repairs needed:

Repairs Made:

Ready for Bench Check?

<u>Inventory Process</u>	To complete if unit sent to CMI
<input type="checkbox"/> Linked Cylinder to Unit <input type="checkbox"/> Changed Unit Status <input type="checkbox"/> Bench Check Report Completed	<input type="checkbox"/> CMI Workorder <input type="checkbox"/> CMI Certificate of Calibration _____ <input type="checkbox"/> CMI Service Evaluation Form
<input type="checkbox"/> Bench Check Technician Report <input type="checkbox"/> Mock Subject Test Completed <input type="checkbox"/> Certificate of Calibration and Operation <input type="checkbox"/> Service Overview Completed	