

Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 07-30-2018

SERIAL: 80-006141

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Administrative Office at:
Board of Tests for Alcohol and Drug Influence
Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307

Certificate of Calibration and Operation

This is to certify that the calibration of **INTOXILYZER 8000**, serial number 80-006141, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R 126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.

In addition the operational procedures and functions of this Intoxilyzer have also been tested and found to be in compliance with the Oklahoma Board of Tests for Alcohol and Drug Influence - Rules and Regulations OAC Title 40: 25-1-2 and 30-1-3 and is hereby certified for use as an evidential breath testing device within this state.

CERTIFICATION TYPE	DATE
Manufacturer Calibration	12-31-13
Board of Tests Bench Check	01-23-14
Board of Tests Bench Check	05-05-14
Manufacturer Calibration	03-24-15
Board of Tests Bench Check	03-30-15
Manufacturer Calibration	05-27-16
Board of Tests Bench Check	06-28-16
Board of Tests Bench Check	07-30-18



Presentation of this form certifies that the Intoxilyzer, SN: 80-006141, functioned properly at the time of the breath test and hereby certifies the breath test result as valid.

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)
State Director of Tests for Alcohol
and Drug Influence

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 07/30/2018	TIME PROCEDURE BEGAN 13:12	INSTRUMENT LOCATION 80-006141
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-006141

TEST RESULTS

Diagnostics
PASS

Mouth Alcohol Test
PASS

RFI Test
PASS

Abort Test
PASS

Improper Sample Test
PASS

Interferent Detect Test
PASS

No Sample Given Test
PASS

Insufficient Sample Test
PASS

Air Blank Contamination Test
PASS

Range Exceeded Test
PASS


Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.018	0.018	0.018
.04	0.038	0.037	0.038
.10	0.097	0.097	0.100
.20	0.199	0.198	0.200

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 07-30-2018
NAME SMITH, JOSHUA	PERMIT # 0000016837



3600 N. Martin Luther King Bldg #9
 Oklahoma City, OK 73111
 405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 07 / 30 / 18 Start Time: 1215 End Time: 1520

INSTRUMENT

Intoxilyzer Model Serial #
8000 80-006141

GAS CANISTER

LOT # EXP Date
10818080A5 # 048 06-05-2020


REFERENCE

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR3591	DR3753	DR3594	DR3754
Concentration			
.02	.04	.10	.20
LOT #			
16400	16320	17130	17090
Manufactured Date			
12-12-2016	10-21-2016	04-03-2017	02-24-2017
Expiration Date			
12-12-2018	10-21-2018	04-03-2019	02-24-2019
Solution Commission Date			
<u>07-30-18</u>	<u>07-30-18</u>	<u>07-30-18</u>	<u>07-30-18</u>

REASON FOR BENCH CHECK


- POST REPAIR CYLINDER REPLACEMENT
 TROUBLESHOOTING ROUTINE MAINTENANCE

COMMENTS

 16837
 Technician Signature / Permit #

Josh Smith
 Printed Name

Certificate of Analysis - Dry Gas	
07-30-18	.080
Date of Analysis	Labelled target value (g/210L)
048	.0791
Cylinder #	Average test result
JOSH SMITH	
BOT Technician Name and Signature	



BRT Form 5.0

In accordance with BOT Policy and Procedure Statement BRT-2.1.0, Maintenance, the above referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.
 7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
 217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 07/30/2018 12:17

Test	g/210L	Time
Air Blank	0.000	12:18
Cal Check	0.079	12:18
Air Blank	0.000	12:18
Cal Check	0.079	12:19
Air Blank	0.000	12:19
Cal Check	0.079	12:20
Air Blank	0.000	12:20
Cal Check	0.079	12:20
Air Blank	0.000	12:21
Cal Check	0.079	12:21
Air Blank	0.000	12:22
Cal Check	0.080	12:22
Air Blank	0.000	12:23
Cal Check	0.079	12:23
Air Blank	0.000	12:24
Cal Check	0.079	12:24
Air Blank	0.000	12:24
Cal Check	0.079	12:25
Air Blank	0.000	12:25
Cal Check	0.079	12:26
Air Blank	0.000	12:26
Cal Check Stats		
Average	0.0791	
Std Dev	0.0003	
Rel Std Dev(%)	0.3997	

Certificate of Analysis

Certificate ID: 11093
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 10818080A5
Expiration: 6/5/2020

REC'D 05/21/18

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/- 0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:
 CMI Inc.
 316 East Ninth Street
 Owensboro, KY 42301
 Phone 866-835-0690
www.alcoholtest.com

Operator's Signature: *[Signature]* 16837

*NIST Traceable Reference Material
 Cylinder No. CC274523 / Job No. 09160306
 Certified 362.2 µmol/mol Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
 Specialty Gas Lab Tech

05-09-18
 Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 07/30/2018		ARREST TIME 1430		COUNTY # 55	CITY #	CITATION #		
ARREST LOCATION MOCK					CITY	COUNTY OKLAHOMA		
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 08/09/1961	SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73133		
DRIVER LICENSE # T083454871			EXPIRATION DATE 06/30/2019	STATE OK	CLASS A	RESTRICTIONS	ENDORSEMENTS M	
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK		STATE OK	CMV N	PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. ROUTINE MAINTENANCE. NEW GAS CANISTER INSTALL.


Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, AND LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006141	SPECIALIST BOARD OF TESTS		
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 16837		
MANUFACTURER ILMO		LOT # 10818080A5	DATE CYLINDER INSTALLED 07/30/2018	DATE CYLINDER EXPIRES 06/05/2020	
OBSERVATION START 1450	TEST DATE 07/30/2018	MOUTHPIECE G	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p>_____ SIGNATURE OF OPERATOR</p> <p>NAME SMITH, JOSHUA</p> <p>BADGE # 0000000435</p> <p>PERMIT # 0000016837</p> <p>AGENCY BOARD OF TESTS</p>		
RESULT TABLE					
Test	g/210L	Time			
Air Blank	0.00	15:13			
Subject Test	0.00	15:14			
Breath Volume	1.546 LITERS				
Air Blank	0.00	15:14			
Wait		15:16			
Air Blank	0.00	15:16			
Subject Test	0.00	15:17			
Breath Volume	1.246 LITERS				
Air Blank	0.00	15:17			
Cal Check	0.08	15:17			
Air Blank	0.00	15:18			

Test Result BrAC g/210L <h2 style="margin: 0;">0.00</h2>
--

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **07/30/2018** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **07/30/2018** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

No Temporary License Issued:
Test Below Legal Limit

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 07/30/2018		ARREST TIME 1430		COUNTY # 55	CITY #	CITATION #		
ARREST LOCATION MOCK				CITY		COUNTY OKLAHOMA		
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 08/09/1961	SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73133		
DRIVER LICENSE # T083454871			EXPIRATION DATE 06/30/2019	STATE OK	CLASS A	RESTRICTIONS	ENDORSEMENTS M	
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK		STATE OK	CMV N	PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. ROUTINE MAINTENANCE. NEW GAS CANISTER INSTALL.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006141	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 16837																																					
MANUFACTURER ILMO		LOT # 10818080A5	DATE CYLINDER INSTALLED 07/30/2018	DATE CYLINDER EXPIRES 06/05/2020																																				
OBSERVATION START 1450	TEST DATE 07/30/2018	MOUTHPIECE G	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p>_____ SIGNATURE OF OPERATOR</p>																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>15:13</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>15:14</td></tr> <tr><td>Breath Volume</td><td>1.546 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:14</td></tr> <tr><td>Wait</td><td></td><td>15:16</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:16</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>15:17</td></tr> <tr><td>Breath Volume</td><td>1.246 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:17</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>15:18</td></tr> </tbody> </table>					Test	g/210L	Time	Air Blank	0.00	15:13	Subject Test	0.00	15:14	Breath Volume	1.546 LITERS		Air Blank	0.00	15:14	Wait		15:16	Air Blank	0.00	15:16	Subject Test	0.00	15:17	Breath Volume	1.246 LITERS		Air Blank	0.00	15:17	Cal Check	0.08	15:17	Air Blank	0.00	15:18
Test	g/210L	Time																																						
Air Blank	0.00	15:13																																						
Subject Test	0.00	15:14																																						
Breath Volume	1.546 LITERS																																							
Air Blank	0.00	15:14																																						
Wait		15:16																																						
Air Blank	0.00	15:16																																						
Subject Test	0.00	15:17																																						
Breath Volume	1.246 LITERS																																							
Air Blank	0.00	15:17																																						
Cal Check	0.08	15:17																																						
Air Blank	0.00	15:18																																						
NAME SMITH, JOSHUA			<div style="border: 1px solid black; padding: 5px;"> <p>Test Result BrAC g/210L 0.00</p> </div>																																					
BADGE # 0000000435		PERMIT # 0000016837																																						
AGENCY BOARD OF TESTS																																								

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **07/30/2018** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **07/30/2018** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

No Temporary License Issued:
Test Below Legal Limit

ARRESTEE'S COPY
BOT Form AFF02 07/2008

Certificate of Analysis - Wet Bath	
05-05-17	.020
Date of Analysis	Labeled target value (g/210L)
16400	.0195
Lot Number	Average test result (g/210L)
Jost Smith	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 G, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

INTOXILYZER 8000
 Instrument Initialization
 8133.12
 14:48 05/05/2017

Certified Alcohol Reference Solution for Simulator

Rec'd
 4/01/2017
 #16837

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 05/05/2017 14:56

Random Samples of Lot Number 16400 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 13, 2016, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0241% (w/vol) ethyl alcohol. The expiration date for this lot number is December 12, 2018 at 11:59 PM.

Test	g/210L	Time
Air Blank	0.000	14:57
Cal Check	0.020	14:58
Air Blank	0.000	14:58
Cal Check	0.020	14:59
Air Blank	0.000	14:59
Cal Check	0.019	15:00
Air Blank	0.000	15:01
Cal Check	0.020	15:01
Air Blank	0.000	15:02
Cal Check	0.020	15:03
Air Blank	0.000	15:03
Cal Check	0.020	15:04
Air Blank	0.000	15:04
Cal Check	0.019	15:05
Air Blank	0.000	15:06
Cal Check	0.019	15:06
Air Blank	0.000	15:07
Cal Check	0.019	15:08
Air Blank	0.000	15:08
Cal Check	0.019	15:09
Air Blank	0.000	15:09
Cal Check Stats		
Average	0.0195	
Std Dev	0.0005	
Rel Std Dev(%)	2.7026	

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN11241503 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

#16837

Certificate of Analysis – Wet Bath	
3-8-2017	.040
Date of Analysis	Labeled target value (g/210L)
16320	.0409
Lot Number	Average test result (g/210L)
Josh Smith <i>[Signature]</i>	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 BURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT – 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

RECD
03/01/17
[Signature]
#16837

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 16320 ✓ of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 25, 2016**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0495% (w/vol)** ethyl alcohol. The expiration date for this lot number is **October 21, 2018** at 11:59 PM.

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 03/08/2017 11:16

Test	g/210L	Time
Air Blank	0.000	11:16
Cal Check	0.041	11:17
Air Blank	0.000	11:17
Cal Check	0.041	11:18
Air Blank	0.000	11:19
Cal Check	0.040	11:19
Air Blank	0.000	11:20
Cal Check	0.041	11:21
Air Blank	0.000	11:21
Cal Check	0.041	11:22
Air Blank	0.000	11:22
Cal Check	0.041	11:23
Air Blank	0.000	11:24
Cal Check	0.041	11:24
Air Blank	0.000	11:25
Cal Check	0.040	11:26
Air Blank	0.000	11:26
Cal Check	0.042	11:27
Air Blank	0.000	11:27
Cal Check	0.041	11:28
Air Blank	0.000	11:29
Cal Check Stats		
Average	0.0409	
Std Dev	0.0005	
Rel Std Dev(%)	1.3878	


When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

[Signature]
#16837

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN1218501 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath	
06/15/2017	.10
Date of Analysis	Labelled target value (g/210L)
17130	.1015
Lot Number	Average test result (g/210L)
 BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT – 2.1.0, *Maintenance*, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

*Rec'd
06/09/2017
[Signature]*

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 06/15/2017

SN 80-004923
 16:25

Certified Alcohol Reference Solution for Simulator

Test	g/210L	Time
Air Blank	0.000	16:26
Cal Check	0.102	16:26
Air Blank	0.000	16:27
Cal Check	0.102	16:28
Air Blank	0.000	16:28
Cal Check	0.101	16:29
Air Blank	0.000	16:29
Cal Check	0.102	16:30
Air Blank	0.000	16:31
Cal Check	0.101	16:31
Air Blank	0.000	16:32
Cal Check	0.101	16:33
Air Blank	0.000	16:33
Cal Check	0.102	16:34
Air Blank	0.000	16:34
Cal Check	0.101	16:35
Air Blank	0.000	16:36
Cal Check	0.102	16:36
Air Blank	0.000	16:37
Cal Check	0.101	16:38
Air Blank	0.000	16:38
Cal Check Stats		
Average	0.1015	
Std Dev	0.0005	
Rel Std Dev(%)	0.5192	

Random Samples of Lot Number **17130** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 6, 2017**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1212%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 3, 2019** at 11:59 PM. ✓


When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

[Signature]
 Operator's Signature

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis - Wet Bath	
06-16-17	.20
Date of Analysis	Labelled target value (g/210L)
17090	.2045
Lot Number	Average test result (g/210L)
 BOT Technician Name and Signature	



LABORATORIES, INC.
 HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

REC'D
 06/09/17


CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 06/16/2017 10:52


Certified Alcohol Reference Solution for Simulator


Test	g/210L	Time
Air Blank	0.000	10:53
Cal Check	0.204	10:53
Air Blank	0.000	10:54
Cal Check	0.205	10:55
Air Blank	0.000	10:55
Cal Check	0.205	10:56
Air Blank	0.000	10:57
Cal Check	0.205	10:57
Air Blank	0.000	10:58
Cal Check	0.205	10:59
Air Blank	0.000	10:59
Cal Check	0.205	11:00
Air Blank	0.000	11:00
Cal Check	0.204	11:01
Air Blank	0.000	11:02
Cal Check	0.204	11:02
Air Blank	0.000	11:03
Cal Check	0.204	11:04
Air Blank	0.000	11:04
Cal Check	0.204	11:05
Air Blank	0.000	11:05
Cal Check Stats		
Average	0.2045	
Std Dev	0.0005	
Rel Std Dev(%)	0.2577	

Random Samples of Lot Number 17090 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on February 27, 2017, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2416% (w/vol) ethyl alcohol. The expiration date for this lot number is February 24, 2019 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.


 Ted L. Pauley, President
 GUTH LABORATORIES, INC.


 16837
 Operator's Signature

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN08101505 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.