



STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT
Chairman

Joshua Smith
Director

ATTESTATION

I, the undersigned [Joshua Smith](#) (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-006155, in compliance with the agency's Maintenance Policy and Procedures.

Signed _____

Name of Position: Director/Records Custodian

Date of Attestation: 07/10/2024

Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



2/20

3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 7/10/2024

SERIAL: 80-006155

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- CMI Certificate of Calibration
- CMI Evaluation Form
- CMI Work Order
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis - Reference Solution (4)
- Guth Laboratories Certificate of Analysis (4)



STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION

This certifies the Intoxilyzer 8000 referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-006155 on 7/10/2024 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
 - Manufacturer: ILMO Unit: 208 ppm / 105 L
 - Cylinder Lot No. - Canister No.: 28423080A2-063
 - BOT analysis average test result: 0.082
 - Cylinder expiration date: 11/5/2025
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR2834	DR3694	DR2455	DR3470
Concentration			
.02	.04	.10	.20
LOT #			
23360	23400	23450	23340
Manufactured Date			
9/25/2023	10/24/2023	12/5/2023	9/18/2023
Expiration Date			
9/25/2025	10/24/2025	12/5/2025	9/18/2025
Solution Commission Date			
7/10/2024	7/10/2024	7/10/2024	7/10/2024

Completed by BOT Personnel:


Christopher Pape
Signature/Name

Breath-Alcohol Testing Program Administrator
Title

Certified by:

Joshua Smith
Signature Name

State Director
Title

STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT

BENCH CHECK DATE 07/10/2024	TIME PROCEDURE BEGAN 12:10:2	INSTRUMENT LOCATION 80-006155
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-006155

TEST RESULTS

Diagnostics
PASS

Mouth Alcohol Test
PASS

RFI Test
PASS

Abort Test
PASS

Improper Sample Test
PASS

Interferent Detect Test
PASS

No Sample Given Test
PASS

Insufficient Sample Test
PASS

Air Blank Contamination Test
PASS

Range Exceeded Test
PASS

Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.018	0.018	0.019
.04	0.039	0.039	0.039
.10	0.101	0.099	0.099
.20	0.199	0.197	0.197

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 7/10/2024
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-006155, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58 No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date June 18, 2024 Signed Christopher Randall
Technician



316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev.A

CMI INC.

Service Evaluation Form

This form **MUST** be completed and enclosed with item to be serviced.
***** Hazardous Material Warning! – DO NOT return gas cylinder with instrument! *****

Note: Please ship items in their original shipping container.

Contact information:

Name Chris Pape Phone: (405) 425-2460
 Email: christopher.pape@bot.ok.gov Customer # 731362 (contact Customer Service)

Your Billing Address

Agency Name: Board of Tests
 Address: 3600 N ML King, Bldg 9
 City: Oklahoma City State: OK Zip: 73111
 Credit Card/PO #:
 Name on Card:
 Expiration Date: _____ CVV _____

Your Shipping Address

Agency Name: Board of Tests
 Address: 3600 N ML King, Bldg 9
 City: Oklahoma City State: OK Zip: 73111
 Contact Person: Chris Pape
 Contact Phone: (405) 425-2460
 Email: christopher.pape@bot.ok.gov

Instrument Serial Number: 80-006155

Detailed Description of Problem:

Instrument's source is unstable. Failed snap test during evaluation after returned from field.

Note: For instruments not under warranty, an evaluation fee of \$83.00 (infrared and I-200) or \$43.50 (fuel cell), plus return shipping cost, will apply to all service items.

I authorize all repairs up to \$500 to be performed.

After repair, add "Certificate of Calibration" for: S-D2, S-D5, and I-800 (\$19)
 I-200, I-240, I-300, I-400, and I-600 (\$39)

Authorized By:

<u>Chris Pape</u>	<u>Breath Testing Admin</u>
Name (Please Print)	Title
	<u>4/12/2024</u>
Signature	Date

Ship item to:

CMI, Inc.
 Attn: Service Dept.
 316 East Ninth Street
 Owensboro, KY 42303



7/20

DATE: 4/24/2024

316 E 9th Street / Owensboro KY 42303 / USA
Phone: 866-835-0690 Fax: 270-685-6268

SERVICE WORK ORDER

408746

Bill To:
Department Of Public Safety
Material Management-Maintenance
PO Box 53004
Oklahoma City, OK 73152-9998
USA
Customer #: 731362
Phone: 405-425-2424
Fax: 405-425-2490 fax

Ship To:
Oklahoma Board Of Tests
Blvd Adcu Annex Bldg #9
3600 Martin Luther King
Oklahoma City OK 73136-3000
USA
Phone: 405-425-2424

MODEL #: 0024800K

Serial Number: 80-006155

BILL CODE: Out of Warranty

EXTRA PARTS RCVD:

8000- BH, KB, BATTERY, GAS SHELF

DESCRIPTION OF PROBLEM

INSTRUMENTS SOURCE IS UNSTABLE. FAILED SNAP TEST DURING EVALUATION AFTER RETURNED FROM FIELD.

WORK PERFORMED:

Replaced parts listed; Calibrated/ Final tested; Software 8133.16.

PARTS USED

Seq. No.	Part	Description	Quantity
100	650517	CERTIFICATE OF CALIBRATION	2.00 EA
110	441169	COVER DUST,5/8in x 1/2in	1.00 EA
120	470145	CAP,PLASTIC,.25IDx.50LG,RED	3.00 EA
130	470154	CAP,PLSTC,.406IDx.438-.562,RED	1.00 EA
140	470218	SEAL,TAMPER EVIDENT,SERVICE	1.00 EA
150	530030	TUBING,TYGON,.375IDx.562OD	2.00 IN
160	210120	SENSOR,SMD,PRESSURE	1.00 EA
170	140112	DISPLAY,2x20 VFD 9MM	1.00 EA
180	320643	HDR,14PIN,2ROW,.100CC,4-SIDED	1.00 EA
190	R020982	SOURCE ASSY, requalified	1.00 EA

LABOR / TESTING

Misc Code	Description	Hours
LABR	Service Repair Labor	
LCAL	Service - Calibration Adjust	0.50
LFT	Service - Cal / Final Test	1.00

Technician Name: Levi Ray

SRV4087460001





STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

Dr. Jarrad Wagner, Ph.D., F-ABFT
Chairman

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov


Joshua Smith
Director

Certificate of Analysis – Dry Gas

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
02/05/2024 09:56

Test	g/210L	Time
Air Blank	0.000	09:57
Cal Check	0.082	09:57
Air Blank	0.000	09:57
Cal Check	0.082	09:58
Air Blank	0.000	09:58
Cal Check	0.081	09:59
Air Blank	0.000	09:59
Cal Check	0.082	10:00
Air Blank	0.000	10:00
Cal Check	0.082	10:00
Air Blank	0.000	10:01
Cal Check	0.081	10:01
Air Blank	0.000	10:02
Cal Check	0.081	10:02
Air Blank	0.000	10:03
Cal Check	0.082	10:03
Air Blank	0.000	10:03
Cal Check	0.082	10:04
Air Blank	0.000	10:04
Cal Check	0.081	10:05
Air Blank	0.000	10:05
Cal Check Stats		
Average	0.0816	
Std Dev	0.0005	
Rel Std Dev (%)	0.6328	

Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)
0.08 g/210L = 208 PPM ÷ 2605*

Manufacturer: ILMO	Manufacturer Cert. No.: 16055	Cylinder Size: 105L
Lot No.: 28423080A2	Cylinder No.: 063	Expiration Date: 11/05/2025
The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM ÷ 2605].		
Signature 	Date: 02/05/2024	
Name Christopher Pape	Permit No.: 23020	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 16055
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 28423080A2
Expiration: 11/5/2025

RCU'D
1/16/2024
CP

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)


Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 261.0 µmol/mol
Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).


Specialty Gas Lab Tech

10-25-2023
Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory

OFFICER'S IMPAIRED DRIVING AFFIDAVIT

10/20

Section 1

ARREST DATE 07/10/2024		ARREST TIME 1330		COUNTY # 55	CITY #	CITATION #	
ARREST LOCATION BOT				CITY		COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, FRANK				DATE OF BIRTH 03/15/1981		SEX M	HEIGHT 509
ADDRESS 3600 MARTIN LUTHER KING				CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73111
DRIVER LICENSE # H083463197			EXPIRATION DATE 10/31/2022		STATE OK	CLASS D	RESTRICTIONS
VEHICLE MAKE MOCK		MODEL MOCK		TAG # MOCK		STATE OK	CMV N
						PLACARD N	ENDORSEMENTS

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006155	SPECIALIST BOARD OF TESTS	
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020	
MANUFACTURER ILMO		LOT # 28423080A2	DATE CYLINDER INSTALLED 07/10/2024	DATE CYLINDER EXPIRES 11/05/2025
OBSERVATION START 1345	TEST DATE 07/10/2024	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p>  <p>_____ SIGNATURE OF OPERATOR</p>		
RESULT TABLE				
Test	g/210L	Time		
Air Blank	0.00	14:09		
Subject Test	0.00	14:09		
Breath Volume	2.121 LITERS			
Air Blank	0.00	14:10		
Wait		14:11		
Air Blank	0.00	14:12		
Subject Test	0.00	14:12		
Breath Volume	2.148 LITERS			
Air Blank	0.00	14:13		
Cal Check	0.08	14:13		
Air Blank	0.00	14:13		
NAME PAPE, CHRISTOPHER		BADGE # 0000000484		
		PERMIT # 0000023020		
AGENCY BOARD OF TESTS				



Test Result BrAC g/210L <h2 style="margin: 0;">0.00</h2>
--

What happens next?

Section 3

1. The Arresting Officer is required to submit the Officer's Impaired Driving Affidavit to the State of Oklahoma.
2. Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
3. The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma. To confirm or change your address on file, contact Service Oklahoma.
4. The Revocation and/or Disqualification will be effective 55 days from the date of the Order, including 10 days allowed for mailing. 47 O.S. §§2-116, 753, 754.
5. The Order of Revocation and/or Disqualification will contain important information about your rights to appeal the revocation and/or disqualification of your driving privileges. To appeal the applicable Order of Revocation and/or Disqualification your petition must be filed with the District Court of the County in which you were arrested within 40 days of the date of the order, including 10 days allowed for mailing. 47 O.S. §§2-116, 6-211.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: _____ Place (location when signed): _____ Signature of arresting officer: _____

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE # 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

OFFICER'S IMPAIRED DRIVING AFFIDAVIT

11/20

Section 1

ARREST DATE 07/10/2024		ARREST TIME 1330		COUNTY # 55	CITY #	CITATION #	
ARREST LOCATION BOT				CITY		COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, FRANK				DATE OF BIRTH 03/15/1981		SEX M	HEIGHT 509
ADDRESS 3600 MARTIN LUTHER KING				CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73111
DRIVER LICENSE # H083463197			EXPIRATION DATE 10/31/2022		STATE OK	CLASS D	RESTRICTIONS
VEHICLE MAKE MOCK		MODEL MOCK		TAG # MOCK		STATE OK	CMV N
						PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST

Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006155	SPECIALIST BOARD OF TESTS	
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020	
MANUFACTURER ILMO		LOT # 28423080A2	DATE CYLINDER INSTALLED 07/10/2024	DATE CYLINDER EXPIRES 11/05/2025
OBSERVATION START 1345	TEST DATE 07/10/2024	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p>_____ SIGNATURE OF OPERATOR</p>		
RESULT TABLE				
Test	g/210L	Time		
Air Blank	0.00	14:09		
Subject Test	0.00	14:09		
Breath Volume	2.121 LITERS			
Air Blank	0.00	14:10		
Wait		14:11		
Air Blank	0.00	14:12		
Subject Test	0.00	14:12		
Breath Volume	2.148 LITERS			
Air Blank	0.00	14:13		
Cal Check	0.08	14:13		
Air Blank	0.00	14:13		
NAME PAPE, CHRISTOPHER		BADGE # 0000000484		
		PERMIT # 0000023020		
AGENCY BOARD OF TESTS				



Test Result BrAC g/210L <h2 style="margin: 0;">0.00</h2>
--

What happens next?

Section 3

1. The Arresting Officer is required to submit the Officer's Impaired Driving Affidavit to the State of Oklahoma.
2. Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
3. The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma. To confirm or change your address on file, contact Service Oklahoma.
4. The Revocation and/or Disqualification will be effective 55 days from the date of the Order, including 10 days allowed for mailing. 47 O.S. §§2-116, 753, 754.
5. The Order of Revocation and/or Disqualification will contain important information about your rights to appeal the revocation and/or disqualification of your driving privileges. To appeal the applicable Order of Revocation and/or Disqualification your petition must be filed with the District Court of the County in which you were arrested within 40 days of the date of the order, including 10 days allowed for mailing. 47 O.S. §§2-116, 6-211.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: _____ Place (location when signed): _____ Signature of arresting officer: _____

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT
Chairman

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov


Joshua Smith
Director

Certificate of Analysis – Reference Solution

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
04/16/2024 11:24

Test	g/210L	Time
Air Blank	0.000	11:24
Cal Check	0.018	11:25
Air Blank	0.000	11:25
Cal Check	0.019	11:26
Air Blank	0.000	11:27
Cal Check	0.019	11:27
Air Blank	0.000	11:28
Cal Check	0.019	11:29
Air Blank	0.000	11:29
Cal Check	0.019	11:30
Air Blank	0.000	11:30
Cal Check	0.019	11:31
Air Blank	0.000	11:32
Cal Check	0.019	11:32
Air Blank	0.000	11:33
Cal Check	0.019	11:34
Air Blank	0.000	11:34
Cal Check	0.019	11:35
Air Blank	0.000	11:35
Cal Check	0.018	11:36
Air Blank	0.000	11:37
Cal Check Stats		
Average	0.0188	
Std Dev	0.0004	
Rel Std Dev (%)	2.2426	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: Guth	Manufactured Date: 9/25/2023	Concentration: 0.02
Lot No.: 23360	Expiration Date: 9/25/2025	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: 4/16/2024	
Name Christopher Pape	Permit No.: 23020	

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSISRCVD
12/19/23
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23360** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 26, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0240%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 25, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.020 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03122113 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT
Chairman


Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

Joshua Smith
Director

Certificate of Analysis – Reference Solution

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
05/02/2024 09:26

Test	g/210L	Time
Air Blank	0.000	09:26
Cal Check	0.039	09:27
Air Blank	0.000	09:28
Cal Check	0.039	09:28
Air Blank	0.000	09:29
Cal Check	0.039	09:30
Air Blank	0.000	09:30
Cal Check	0.039	09:31
Air Blank	0.000	09:31
Cal Check	0.039	09:32
Air Blank	0.000	09:33
Cal Check	0.040	09:33
Air Blank	0.000	09:34
Cal Check	0.039	09:35
Air Blank	0.000	09:35
Cal Check	0.039	09:36
Air Blank	0.000	09:37
Cal Check	0.039	09:37
Air Blank	0.000	09:38
Cal Check	0.039	09:38
Air Blank	0.000	09:39
Cal Check Stats		
Average	0.0391	
Std Dev	0.0003	
Rel Std Dev (%)	0.8087	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.		
Manufacturer: Guth	Manufactured Date: 10/24/2023	Concentration: 0.04
Lot No.: 23400	Expiration Date: 10/24/2025	Equipment Model: Guth 2100
<p>The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.</p>		
Signature 	Date: 5/2/2024	
Name Christopher Pape	Permit No.: 23020	

**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

RCVD
12/19/23
CP

Random Samples of Lot Number **23400** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 25, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0484%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 24, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN06032102 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT
Chairman


Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

Joshua Smith
Director

Certificate of Analysis – Reference Solution

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002591
01/23/2024 09:07

Test	g/210L	Time
Air Blank	0.000	09:08
Cal Check	0.099	09:08
Air Blank	0.000	09:09
Cal Check	0.099	09:10
Air Blank	0.000	09:10
Cal Check	0.100	09:11
Air Blank	0.000	09:11
Cal Check	0.100	09:12
Air Blank	0.000	09:13
Cal Check	0.100	09:13
Air Blank	0.000	09:14
Cal Check	0.100	09:15
Air Blank	0.000	09:15
Cal Check	0.100	09:16
Air Blank	0.000	09:17
Cal Check	0.100	09:17
Air Blank	0.000	09:18
Cal Check	0.100	09:18
Air Blank	0.000	09:19
Cal Check	0.100	09:20
Air Blank	0.000	09:20
Cal Check Stats		
Average	0.0998	
Std Dev	0.0004	
Rel Std Dev (%)	0.4224	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.		
Manufacturer: Guth	Manufactured Date: 12/5/2023	Concentration: 0.100
Lot No.: 23450	Expiration Date: 12/5/2025	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: 01/23/2024	
Name Christopher Pape	Permit No.: 23020	

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSISRCVP
12/19/23
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23450** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 7, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1211%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 5, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT
Chairman

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

Joshua Smith
Director

Certificate of Analysis – Reference Solution

ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000


SN 80-002591

10/04/2023

09:22

Test	g/210L	Time
Air Blank	0.000	09:22
Cal Check	0.198	09:23
Air Blank	0.000	09:24
Cal Check	0.198	09:24
Air Blank	0.000	09:25
Cal Check	0.199	09:26
Air Blank	0.000	09:26
Cal Check	0.198	09:27
Air Blank	0:000	09:27
Cal Check	0.198	09:28
Air Blank	0.000	09:29
Cal Check	0.198	09:29
Air Blank	0.000	09:30
Cal Check	0.198	09:31
Air Blank	0.000	09:31
Cal Check	0.198	09:32
Air Blank	0.000	09:33
Cal Check	0.199	09:33
Air Blank	0.000	09:34
Cal Check	0.199	09:35
Air Blank	0.000	09:35
Cal Check Stats		
Average	0.1983	
Std Dev	0.0004	
Rel Std Dev(%)	0.2435	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: Guth	Manufactured Date: 9/18/2023	Concentration: 0.200
Lot No.: 23340	Expiration Date: 9/18/2025	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: 10/4/2023	
Name Christopher Pape	Permit No.: 23020	

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSISRCVD
10/2/2023
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23340** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 19, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2399%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 18, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN09142002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Location Returned From Midwest City PD 2	Unit Serial # 80-006155	Date: 20/20 10/11/23
---	----------------------------	--

Inventory Process

Unit checked in to BOT
 Unlink Gas Canister from Unit
 Old Cylinder #: 19021080A2-018
 New Unit Checked out
 Remove old cylinder from unit (After Lab Evaluation)
 Old cylinder disposed in inventory
 Old cylinder archived

Check in Procedure

Connect unit to ethernet connection 172.019.004. 141
 Download any stored records 0
 Change unit location to Serial number
 Update Unit Date and time
 Previous time: 10/12/2023 12:46
 New Time: 10/12/2023 12:48
 Execute a new Forms load on the unit

Notes from the Field

Routine Swap
 Issues in the Field

Lab Evaluation

Breath Hose Hold Vacuum?: Test Chamber Hold Vacuum?:

DVM

3µm 12966 Chamber Temp 47.00 Snap Test ok?
 9µm 13775 Breath Hose Temp 45 RFI Detected?

Peripheral Tests

Speaker (F5) Power LED Green (F6) Red (F7) Fan (F9)
 Gas Regulator (F11) Pump (F10) Receipt paper

Atmospheric Sensor

Pressure 956 Correction Factor 1.05 Flow Sensor Resistance 167 Tank Sensor Value 879
 Is regulator close to Tank Sensor Value? Yes / No

Sent to CMI (If needed) Reason Sent

Date Sent 4/12/24
 Invoice Sent 5/20/24
 Date Returned 7/1/24

Instrument's source is unstable. Failed snap test during evaluation after returned from field.

Repairs needed:

Repairs Made:

Ready for Bench Check?

<u>Inventory Process</u> <input type="checkbox"/> Linked Cylinder to Unit <input type="checkbox"/> Changed Unit Status <input type="checkbox"/> Bench Check Report Completed	<input type="checkbox"/> Bench Check Technician Report <input type="checkbox"/> Mock Subject Test Completed <input type="checkbox"/> Certificate of Calibration and Operation <input type="checkbox"/> Service Overview Completed	<u>To complete if unit sent to CMI</u> <input type="checkbox"/> CMI Workorder <input type="checkbox"/> CMI Certificate of Calibration <input type="checkbox"/> CMI Service Evaluation Form
---	--	---