


Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 10-22-2018

SERIAL: 80-006160

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,

A handwritten signature in black ink, appearing to read "J. Robert Blakeburn". The signature is fluid and cursive, written over a light background.

J. Robert Blakeburn (OBA11059)
State Director of Tests for Alcohol
and Drug Influence

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

| | | |
|--------------------------------|-------------------------------|---------------------------------------|
| BENCH CHECK DATE 10/22/2018 | TIME PROCEDURE BEGAN 07:34 | INSTRUMENT LOCATION 80-006160 |
| INSTRUMENT TYPE Intoxilyzer | INSTRUMENT MODEL I-8000 | INSTRUMENT SERIAL NUMBER 80-006160 |

TEST RESULTS

Diagnostics

PASS

Mouth Alcohol Test

PASS

RFI Test

PASS

Abort Test

PASS

Improper Sample Test

PASS

Interferent Detect Test

PASS

No Sample Given Test

PASS

Insufficient Sample Test

PASS

Air Blank Contamination Test

PASS

Range Exceeded Test

PASS


Duplicate Breath Sampling Results

| Value | Test 1 | Test 2 | Control Test |
|-------|--------|--------|--------------|
| .02 | 0.021 | 0.021 | 0.021 |
| .04 | 0.039 | 0.041 | 0.041 |
| .10 | 0.099 | 0.099 | 0.100 |
| .20 | 0.199 | 0.198 | 0.204 |

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

| | |
|--------------------------------------------------------------------------------------------------|------------------------|
| SIGNATURE  | DATE 10-22-18 |
| NAME SMITH, JOSHUA | PERMIT # 0000016837 |



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 10 / 22 / 18 Start Time: 0730 End Time: 0925

INSTRUMENT

Intoxilyzer Model 8000 Serial # 30-006160

GAS CANISTER

LOT # 19B17080A2 # 040 EXP Date 9-5-2019

REFERENCE

| Simulator Model | | | |
|--------------------------|-----------------|-----------------|-----------------|
| Guth 2100 | Guth 2100 | Guth 2100 | Guth 2100 |
| Serial # | | | |
| DR3591 | DR3753 | DR3594 | DR3754 |
| Concentration | | | |
| .02 | .04 | .10 | .20 |
| LOT # | | | |
| 18020 | 17410 | 18070 | 17340 |
| Manufactured Date | | | |
| 01-09-2018 | 12-06-2017 | 02-26-2018 | 10-09-2017 |
| Expiration Date | | | |
| 01-09-2020 | 12-06-2019 | 02-26-2018 | 10-09-2019 |
| Solution Commission Date | | | |
| <u>10-18-18</u> | <u>10-18-18</u> | <u>10-18-18</u> | <u>10-18-18</u> |

REASON FOR BENCH CHECK


- POST REPAIR CYLINDER REPLACEMENT
 TROUBLESHOOTING ROUTINE MAINTENANCE

COMMENTS

REPLACED AIR PUMP & EXHAUST BLOCK ASSEMBLY

Technician Signature / Permit # 16837

Printed Name Jose Suite

| Certificate of Analysis – Dry Gas | |
|-----------------------------------------------------------------------------------------------------------------------|--------------------------------|
| 9/18/17 | 080 |
| Date of Analysis | Labelled target value (g/210L) |
| 040 | 0806 |
| Cylinder # | Average test result |
|  BOT Technician Name and Signature | |



ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 00-002591
 09/18/2017 13:01

ISO/IEC 17025:2005 Accredited Laboratory

In accordance with BOT Policy and Procedure Statement BRT-2.1.0, Maintenance, the above referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 13:02 |
| Cal Check | 0.080 | 13:02 |
| Air Blank | 0.000 | 13:03 |
| Cal Check | 0.080 | 13:03 |
| Air Blank | 0.000 | 13:04 |
| Cal Check | 0.081 | 13:04 |
| Air Blank | 0.000 | 13:04 |
| Cal Check | 0.081 | 13:05 |
| Air Blank | 0.000 | 13:05 |
| Cal Check | 0.081 | 13:06 |
| Air Blank | 0.000 | 13:06 |
| Cal Check | 0.081 | 13:06 |
| Air Blank | 0.000 | 13:07 |
| Cal Check | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Cal Check | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Cal Check | 0.081 | 13:09 |
| Air Blank | 0.000 | 13:10 |
| Cal Check | 0.080 | 13:10 |
| Air Blank | 0.000 | 13:10 |
| Cal Check Stats | | |
| Average | 0.0806 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 0.6406 | |

Certificate of Analysis

Certificate ID: 10399 ✓
 Part #: BAC105L080T
 Cylinder Size: 105L
 Lot Number: 19817080A2 ✓
 Expiration: 9/5/2019 ✓

REC'D
 09/18/2017


0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

| Component: | Concentration: | Accuracy: | Method: |
|------------|----------------|--------------------------|---------|
| Ethanol | 208 ppm | +/- 0.002 or 2% | NDIR |
| Nitrogen | Balance | BAC whichever is greater | |


 Operator's Signature

*NIST Standard Reference Material
 Cylinder No. CC 274507 / Job No. 09160309
 Certified 362.2 µmol/mol Ethanol in Nitrogen
 for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).


 Specialty Gas Lab Tech

08-16-17
 Date

Distributed by:

CMI Inc.
 316 East Ninth Street
 Owensboro, KY 42303
 Phone 866-835-0690
www.alcoholtest.com



OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

| | | | | | | | | |
|---------------------------------------------|--|--------------------------------------|------------------------------|------------------------------------|-----------------------------|-----------------------------|---------------------------|----------------------|
| ARREST DATE 10/22/2018 | | ARREST TIME 0800 | | COUNTY # 55 | CITY # 60 | CITATION # MOCK | | |
| ARREST LOCATION MOCK | | | | | CITY | | COUNTY OKLAHOMA | |
| SUBJECT NAME MOCK, SUBJECT TEST | | | | DATE OF BIRTH 08/09/1961 | SEX M | RACE W | HEIGHT 510 | WEIGHT 165 |
| ADDRESS 3600 MARTIN LUTHER KING A | | | CITY OKLAHOMA CITY | | STATE OK | ZIP CODE 73133 | | |
| DRIVER LICENSE # T083454871 | | EXPIRATION DATE 06/30/2019 | STATE OK | CLASS A | RESTRICTIONS NONE | ENDORSEMENTS NONE | | |
| VEHICLE MAKE NONE | | MODEL NONE | TAG # NONE | | STATE OK | CMV N | PLACARD N | |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK

Describe person's condition (odor, actions, etc.):

MOCK SUBJECT TEST. POST REPAIR.

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

| EQUIPMENT I-8000 | | SERIAL NUMBER 80-006160 | SPECIALIST BOARD OF TESTS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------|-----------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------|------|--------|------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|------|--|-------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|-----------|------|-------|-----------|------|-------|
| STANDARD Dry | | TARGET VALUE 0.080 | PERMIT NUMBER 16837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| MANUFACTURER ILMO | | LOT # 19817080A2 | DATE CYLINDER INSTALLED 10/22/2018 | DATE CYLINDER EXPIRES 09/05/2019 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| OBSERVATION START 0825 | TEST DATE 10/22/2018 | MOUTHPIECE G | <p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p>_____ SIGNATURE OF OPERATOR</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>09:11</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>09:12</td></tr> <tr><td>Breath Volume</td><td>2.621 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:12</td></tr> <tr><td>Wait</td><td></td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:14</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>09:15</td></tr> <tr><td>Breath Volume</td><td>1.566 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:15</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>09:16</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:16</td></tr> </tbody> </table> | | | | | Test | g/210L | Time | Air Blank | 0.00 | 09:11 | Subject Test | 0.00 | 09:12 | Breath Volume | 2.621 LITERS | | Air Blank | 0.00 | 09:12 | Wait | | 09:14 | Air Blank | 0.00 | 09:14 | Subject Test | 0.00 | 09:15 | Breath Volume | 1.566 LITERS | | Air Blank | 0.00 | 09:15 | Cal Check | 0.08 | 09:16 | Air Blank | 0.00 | 09:16 |
| Test | g/210L | Time | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:11 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:12 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 2.621 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:12 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Wait | | 09:14 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:14 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:15 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 1.566 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:15 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Cal Check | 0.08 | 09:16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RESULT TABLE | | | NAME SMITH, JOSHUA | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | | BADGE # 0000000435 | PERMIT # 0000016837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | | AGENCY BOARD OF TESTS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a

Section 3

period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **10/22/2018** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **10/22/2018** Place (location when signed): _____ Signature of arresting officer: _____

| | | | |
|------------------------------|---------------------------------------|---------------------------------|------------------------------|
| NAME SMITH, JOSHUA | | AGENCY BOARD OF TESTS | |
| BADGE 0000000435 | AGENCY ADDRESS INTER-AGENCY | ZIP CODE 00000 | PHONE "4054252460" |

OTHER WITNESSES:

Section 5

| | | | | |
|---|------|-------|---------|-------|
| 1 | NAME | TITLE | ADDRESS | PHONE |
| 2 | NAME | TITLE | ADDRESS | PHONE |

No Temporary License Issued:
Test Below Legal Limit

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

| | | | | | | | | |
|--------------------------------------|--|---------------------|-------------------------------|-----------------------------|--------------|----------------------|----------------------|---------------|
| ARREST DATE 10/22/2018 | | ARREST TIME 0800 | | COUNTY # 55 | CITY # 60 | CITATION # MOCK | | |
| ARREST LOCATION MOCK | | | | | CITY | COUNTY OKLAHOMA | | |
| SUBJECT NAME MOCK, SUBJECT TEST | | | | DATE OF BIRTH 08/09/1961 | SEX M | RACE W | HEIGHT 510 | WEIGHT 165 |
| ADDRESS 3600 MARTIN LUTHER KING A | | | CITY OKLAHOMA CITY | | STATE OK | ZIP CODE 73133 | | |
| DRIVER LICENSE # T083454871 | | | EXPIRATION DATE 06/30/2019 | STATE OK | CLASS A | RESTRICTIONS NONE | ENDORSEMENTS NONE | |
| VEHICLE MAKE NONE | | MODEL NONE | | TAG # NONE | | STATE OK | CMV N | PLACARD N |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK


Describe person's condition (odor, actions, etc.):

MOCK SUBJECT TEST. POST REPAIR.

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

| EQUIPMENT I-8000 | | SERIAL NUMBER 80-006160 | SPECIALIST BOARD OF TESTS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------|----------------------------|---------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------|--|------|--------|------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|------|--|-------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|-----------|------|-------|-----------|------|-------|
| STANDARD Dry | | TARGET VALUE 0.080 | PERMIT NUMBER 16837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| MANUFACTURER ILMO | | LOT # 19817080A2 | DATE CYLINDER INSTALLED 10/22/2018 | DATE CYLINDER EXPIRES 09/05/2019 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| OBSERVATION START 0825 | TEST DATE 10/22/2018 | MOUTHPIECE G | A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration. | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RESULT TABLE <table style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Test</th> <th style="text-align: left;">g/210L</th> <th style="text-align: left;">Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>09:11</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>09:12</td></tr> <tr><td style="padding-left: 20px;">Breath Volume</td><td>2.621 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:12</td></tr> <tr><td>Wait</td><td></td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:14</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>09:15</td></tr> <tr><td style="padding-left: 20px;">Breath Volume</td><td>1.566 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:15</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>09:16</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:16</td></tr> </tbody> </table> | | | | | | Test | g/210L | Time | Air Blank | 0.00 | 09:11 | Subject Test | 0.00 | 09:12 | Breath Volume | 2.621 LITERS | | Air Blank | 0.00 | 09:12 | Wait | | 09:14 | Air Blank | 0.00 | 09:14 | Subject Test | 0.00 | 09:15 | Breath Volume | 1.566 LITERS | | Air Blank | 0.00 | 09:15 | Cal Check | 0.08 | 09:16 | Air Blank | 0.00 | 09:16 |
| Test | g/210L | Time | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:11 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:12 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 2.621 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:12 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Wait | | 09:14 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:14 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:15 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 1.566 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:15 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Cal Check | 0.08 | 09:16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SIGNATURE OF OPERATOR | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| NAME SMITH, JOSHUA | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| BADGE # 0000000435 | PERMIT # 0000016837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| AGENCY BOARD OF TESTS | | |  Test Result BrAC g/210L 0.00 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **10/22/2018** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **10/22/2018** Place (location when signed): _____ Signature of arresting officer: _____

| | | | |
|-----------------------|--------------------------------|--------------------------|-----------------------|
| NAME SMITH, JOSHUA | | AGENCY BOARD OF TESTS | |
| BADGE 0000000435 | AGENCY ADDRESS INTER-AGENCY | ZIP CODE 00000 | PHONE "4054252460" |

No Temporary License Issued:
Test Below Legal Limit

ARRESTEE'S COPY
BOT Form AFF02 07/2008



GUTH LABORATORIES, INC.
 590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

| Certificate of Analysis – Wet Bath | |
|------------------------------------|--------------------------------|
| 09-26-2018 | 0.0200 |
| Date of Analysis | Labelled target value (g/210L) |
| 18020 | 0.0204 |
| Lot Number | Average test result (g/210L) |
| Just Smith | |
| BOT Technician Name and Signature | |



BRT Form 6.0

In accordance with BRT – 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

RECD
 08/25/18
[Signature]

Random Samples of Lot Number 18020 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 11, 2018, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0238% (w/vol) ethyl alcohol. The expiration date for this lot number is January 9, 2020 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 09/26/2018 15:36

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 15:36 |
| Cal Check | 0.020 | 15:37 |
| Air Blank | 0.000 | 15:37 |
| Cal Check | 0.020 | 15:38 |
| Air Blank | 0.000 | 15:39 |
| Cal Check | 0.020 | 15:39 |
| Air Blank | 0.000 | 15:40 |
| Cal Check | 0.020 | 15:41 |
| Air Blank | 0.000 | 15:41 |
| Cal Check | 0.020 | 15:42 |
| Air Blank | 0.000 | 15:42 |
| Cal Check | 0.020 | 15:43 |
| Air Blank | 0.000 | 15:44 |
| Cal Check | 0.021 | 15:44 |
| Air Blank | 0.000 | 15:45 |
| Cal Check | 0.021 | 15:46 |
| Air Blank | 0.000 | 15:46 |
| Cal Check | 0.021 | 15:47 |
| Air Blank | 0.000 | 15:47 |
| Cal Check | 0.021 | 15:48 |
| Air Blank | 0.000 | 15:49 |
| Cal Check Stats | | |
| Average | 0.0204 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 2.5312 | |

[Signature]
 16837
 Operator's Signature

[Signature]
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard, lot number FN08031603 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4814 • TELEPHONE: 717-664-8470

Certificate of Analysis – Wet Bath

| | |
|-----------------------------------|--------------------------------|
| 09-26-18 | .040 |
| Date of Analysis | Labelled target value (g/210L) |
| 17410 | 0.0396 |
| Lot Number | Average test result (g/210L) |
| Josh Smith | <i>[Signature]</i> |
| BOT Technician Name and Signature | |



BRT Form 6.0

In accordance with BRT – 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 17410 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 7, 2017, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0481%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 6, 2019** at 11:59 PM.

*RECD
8/25/18
[Signature]*

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.040 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000
09/26/2018

SN 80-002591
15:54


| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 15:54 |
| Cal Check | 0.040 | 15:55 |
| Air Blank | 0.000 | 15:56 |
| Cal Check | 0.039 | 15:56 |
| Air Blank | 0.000 | 15:57 |
| Cal Check | 0.039 | 15:58 |
| Air Blank | 0.000 | 15:58 |
| Cal Check | 0.040 | 15:59 |
| Air Blank | 0.000 | 15:59 |
| Cal Check | 0.040 | 16:00 |
| Air Blank | 0.000 | 16:01 |
| Cal Check | 0.040 | 16:01 |
| Air Blank | 0.000 | 16:02 |
| Cal Check | 0.039 | 16:03 |
| Air Blank | 0.000 | 16:03 |
| Cal Check | 0.040 | 16:04 |
| Air Blank | 0.000 | 16:04 |
| Cal Check | 0.040 | 16:05 |
| Air Blank | 0.000 | 16:06 |
| Cal Check | 0.039 | 16:06 |
| Air Blank | 0.000 | 16:07 |
| Cal Check Stats | | |
| Average | 0.0396 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 1.3040 | |

[Signature]
Operator's Signature
16837

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN12181501 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

| Certificate of Analysis - Wet Bath | |
|----------------------------------------------------------------------------------|--------------------------------|
| 09-27-2018 | .100 |
| Date of Analysis | Labelled Target value (g/210L) |
| 18070 | 0.0996 |
| Lot Number | Average Test result (g/210L) |
|  | |
| BOT Technician Name and Signature | |



GUTH LABORATORIES, INC.
 590 NORTH 67th STREET • HARRISBURG, PA 17111- 4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

*REC'D
08/25/18
JK*

Random Samples of Lot Number 18070 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 28, 2018**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1219% (w/vol)** ethyl alcohol. The expiration date for this lot number is **February 26, 2020** at 11:59 PM.

ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 09/27/2018 08:28

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 08:28 |
| Cal Check | 0.100 | 08:29 |
| Air Blank | 0.000 | 08:29 |
| Cal Check | 0.100 | 08:30 |
| Air Blank | 0.000 | 08:31 |
| Cal Check | 0.099 | 08:31 |
| Air Blank | 0.000 | 08:32 |
| Cal Check | 0.100 | 08:33 |
| Air Blank | 0.000 | 08:33 |
| Cal Check | 0.100 | 08:34 |
| Air Blank | 0.000 | 08:34 |
| Cal Check | 0.100 | 08:35 |
| Air Blank | 0.000 | 08:36 |
| Cal Check | 0.099 | 08:36 |
| Air Blank | 0.000 | 08:37 |
| Cal Check | 0.099 | 08:38 |
| Air Blank | 0.000 | 08:38 |
| Cal Check | 0.099 | 08:39 |
| Air Blank | 0.000 | 08:39 |
| Cal Check | 0.100 | 08:40 |
| Air Blank | 0.000 | 08:41 |
| Cal Check Stats | | |
| Average | 0.0996 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 0.5184 | |

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley

Ted L. Pauley, President
 GUTH LABORATORIES, INC.

JK 16837
 Operator's Signature

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

| Certificate of Analysis - Wet Bath | |
|------------------------------------|--------------------------------|
| 09-27-2018 | .200 |
| Date of Analysis | Labelled target value (g/210L) |
| 17340 | 0.1968 |
| Lot Number | Test result (g/210L) |
| Josh Smith | |
| BOT Technician Name and Signature | |



GUTH LABORATORIES, INC.
 500 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470
 BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

REC'D
08/25/18
[Signature]

Random Samples of Lot Number 17340 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 10, 2017**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2410%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 9, 2019** at 11:59 PM.

| ALPHA INSTRUMENT | | |
|--------------------------------|--------------|-------|
| Intoxilyzer - Alcohol Analyzer | | |
| Model 8000 | SN 80-002591 | |
| 09/27/2018 | 15:27 | |
| Test | g/210L | Time |
| Air Blank | 0.000 | 15:27 |
| Cal Check | 0.196 | 15:28 |
| Air Blank | 0.000 | 15:29 |
| Cal Check | 0.196 | 15:29 |
| Air Blank | 0.000 | 15:30 |
| Cal Check | 0.197 | 15:31 |
| Air Blank | 0.000 | 15:31 |
| Cal Check | 0.197 | 15:32 |
| Air Blank | 0.000 | 15:33 |
| Cal Check | 0.197 | 15:33 |
| Air Blank | 0.000 | 15:34 |
| Cal Check | 0.197 | 15:35 |
| Air Blank | 0.000 | 15:35 |
| Cal Check | 0.197 | 15:36 |
| Air Blank | 0.000 | 15:36 |
| Cal Check | 0.197 | 15:37 |
| Air Blank | 0.000 | 15:38 |
| Cal Check | 0.197 | 15:38 |
| Air Blank | 0.000 | 15:39 |
| Cal Check | 0.197 | 15:40 |
| Air Blank | 0.000 | 15:40 |
| Cal Check Stats | | |
| Average | 0.1968 | |
| Std Dev | 0.0004 | |
| Rel Std Dev(%) | 0.2142 | |

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

[Signature] 16837
Operator's Signature

[Signature]

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN08101505 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.