



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Joshua Smith  
Director

**ATTESTATION**


I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-006163, in compliance with the agency's Maintenance Policy and Procedures.

Signed \_\_\_\_\_

Joshua Smith

Name of Position: Director/Records Custodian



State of Oklahoma

County of Oklahoma

Signed or attested before me on 10/01/2025 by

Joshua Smith.

Jordan Faith Thompson  
(Signature of notarial officer)

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing.



2/17

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

BENCH CHECK DATE: 04/01/2022

SERIAL: 80-006163

## DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis on Certified Solutions (4)
- Guth Laboratories Certificate of Analysis (4)



3/17

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

**CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION**


This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-006163 on 04/01/2022 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 35521080A3-002
  - BOT analysis average test result: 0.081
  - Cylinder expiration date: 03/05/2024
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR3685	DR3694	DR3594	DR3754
Concentration			
.02	.04	.10	.20
LOT #			
21120	21070	21190	20510
Manufactured Date			
4/7/2021	3/2/2021	6/9/2021	12/3/2020
Expiration Date			
4/7/2023	3/1/2023	6/8/2023	12/3/2022
Solution Commission Date			
03/31/2022	03/31/2022	03/31/2022	03/31/2022

Completed by BOT Personnel:

  
Signature/Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

Certified by:

  
Signature Name Joshua Smith

State Director  
Title



# STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

## CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT

BENCH CHECK DATE <b>04/01/2022</b>	TIME PROCEDURE BEGAN <b>12:16:3</b>	INSTRUMENT LOCATION <b>80-006163</b>
INSTRUMENT TYPE <b>Intoxilyzer</b>	INSTRUMENT MODEL <b>I-8000</b>	INSTRUMENT SERIAL NUMBER <b>80-006163</b>

### TEST RESULTS

**Diagnostics****PASS****Mouth Alcohol Test****PASS****RFI Test****PASS****Abort Test****PASS****Improper Sample Test****PASS****Interferent Detect Test****PASS****No Sample Given Test****PASS****Insufficient Sample Test****PASS****Air Blank Contamination Test****PASS****Range Exceeded Test****PASS**


### Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.019	0.017	0.018
.04	0.037	0.037	0.037
.10	0.097	0.096	0.096
.20	0.192	0.193	0.192

**PASS**

### CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE <b>4/1/2022</b>
NAME <b>PAPE, CHRISTOPHER</b>	PERMIT # <b>0000023020</b>



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Dry Gas**

ALPHA INSTRUMENT 2

Intoxilyzer – Alcohol Analyzer

Model 8000


SN 80-004923

03/25/2022

10:34

Test	g/210L	Time
Air Blank	0.000	10:35
Cal Check	0.081	10:35
Air Blank	0.000	10:36
Cal Check	0.081	10:36
Air Blank	0.000	10:36
Cal Check	0.080	10:37
Air Blank	0.000	10:37
Cal Check	0.081	10:37
Air Blank	0.000	10:38
Cal Check	0.081	10:38
Air Blank	0.000	10:39
Cal Check	0.081	10:39
Air Blank	0.000	10:40
Cal Check	0.081	10:40
Air Blank	0.000	10:41
Cal Check	0.081	10:41
Air Blank	0.000	10:41
Cal Check	0.081	10:42
Air Blank	0.000	10:42
Cal Check	0.081	10:43
Air Blank	0.000	10:43
Cal Check Stats		
Average	0.0809	
Std Dev	0.0003	
Rel Std Dev(%)	0.3908	

**Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)  
0.08 g/210L = 208 PPM ÷ 2605\***

Manufacturer: <b>ILMO</b>	Manufacturer Cert. No.: <b>14549</b>	Cylinder Size: <b>105L</b>
Lot No.: <b>35521080A3</b>	Cylinder No.: <b>002</b>	Expiration Date: <b>03/05/2024</b>
<p>The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM ÷ 2605].</p>		
Signature: 		Date: <b>03/25/2022</b>
Name: <b>Christopher Pape</b>		Permit No.: <b>23020</b>



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • [www.ilmoproducts.com](http://www.ilmoproducts.com)

## Certificate of Analysis

**Certificate ID:** 14549  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 35521080A3  
**Expiration:** 3/5/2024

RCV'D  
2/16/2022  
CP

**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L)	NDIR
Nitrogen	Balance	[5.2 ppm]	

**Distributed by:**

**CMI Inc.**  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

\*Traceable to:

Certified Reference Material - 262.4 µmol/mol

Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

02-01-2022  
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

7/17

Section 1

ARREST DATE 04/01/2022	ARREST TIME 1315	COUNTY # 55	CITY #	CITATION #
ARREST LOCATION BOT		CITY	COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, ALPHA		DATE OF BIRTH 01/01/1980	SEX M	HEIGHT 510 WEIGHT 150
ADDRESS 3600 MARTIN LUTHER KING A		CITY OKLAHOMA CITY	STATE OK	ZIP CODE 73111
DRIVER LICENSE # B083462869		EXPIRATION DATE 10/31/2022	STATE OK	CLASS D RESTRICTIONS ENDORSEMENTS
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK CMV N PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):


POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

## THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000	SERIAL NUMBER 80-006163	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry	TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO	LOT # 35521080A3	DATE CYLINDER INSTALLED 04/01/2022	DATE CYLINDER EXPIRES 03/05/2024																																				
OBSERVATION START 1330	TEST DATE 04/01/2022	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.																																					
RESULT TABLE		I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.																																					
<table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.00</td><td>13:57</td></tr><tr><td>Subject Test</td><td>0.00</td><td>13:59</td></tr><tr><td>Breath Volume</td><td>1.617 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:00</td></tr><tr><td>Wait</td><td></td><td>14:01</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:02</td></tr><tr><td>Subject Test</td><td>0.00</td><td>14:02</td></tr><tr><td>Breath Volume</td><td>1.570 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:03</td></tr><tr><td>Cal Check</td><td>0.08</td><td>14:03</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:04</td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.00	13:57	Subject Test	0.00	13:59	Breath Volume	1.617 LITERS		Air Blank	0.00	14:00	Wait		14:01	Air Blank	0.00	14:02	Subject Test	0.00	14:02	Breath Volume	1.570 LITERS		Air Blank	0.00	14:03	Cal Check	0.08	14:03	Air Blank	0.00	14:04	 SIGNATURE OF OPERATOR	
Test	g/210L	Time																																					
Air Blank	0.00	13:57																																					
Subject Test	0.00	13:59																																					
Breath Volume	1.617 LITERS																																						
Air Blank	0.00	14:00																																					
Wait		14:01																																					
Air Blank	0.00	14:02																																					
Subject Test	0.00	14:02																																					
Breath Volume	1.570 LITERS																																						
Air Blank	0.00	14:03																																					
Cal Check	0.08	14:03																																					
Air Blank	0.00	14:04																																					
		NAME PAPE, CHRISTOPHER																																					
		BADGE # 0000000484	PERMIT # 0000023020																																				
		AGENCY BOARD OF TESTS																																					

  
1907

Test Result  
BrAC g/210L  
**0.00**

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED.** Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER	AGENCY BOARD OF TESTS
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY
ZIP CODE 00000	PHONE 4054252460

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

8/17

Section 1

ARREST DATE 04/01/2022	ARREST TIME 1315	COUNTY # 55	CITY #	CITATION #		
ARREST LOCATION BOT		CITY	COUNTY OKLAHOMA			
SUBJECT NAME BREATHTEST, ALPHA		DATE OF BIRTH 01/01/1980	SEX M	HEIGHT 510	WEIGHT 150	
ADDRESS 3600 MARTIN LUTHER KING A		CITY OKLAHOMA CITY	STATE OK	ZIP CODE 73111		
DRIVER LICENSE # B083462869		EXPIRATION DATE 10/31/2022	STATE OK	CLASS D	RESTRICTIONS ENDORSEMENTS	
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):


POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

## THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000	SERIAL NUMBER 80-006163	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry	TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO	LOT # 35521080A3	DATE CYLINDER INSTALLED 04/01/2022	DATE CYLINDER EXPIRES 03/05/2024																																				
OBSERVATION START 1330	TEST DATE 04/01/2022	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.																																					
RESULT TABLE		I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.																																					
<table border="1"><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.00</td><td>13:57</td></tr><tr><td>Subject Test</td><td>0.00</td><td>13:59</td></tr><tr><td>Breath Volume</td><td>1.617 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:00</td></tr><tr><td>Wait</td><td></td><td>14:01</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:02</td></tr><tr><td>Subject Test</td><td>0.00</td><td>14:02</td></tr><tr><td>Breath Volume</td><td>1.570 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:03</td></tr><tr><td>Cal Check</td><td>0.08</td><td>14:03</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:04</td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.00	13:57	Subject Test	0.00	13:59	Breath Volume	1.617 LITERS		Air Blank	0.00	14:00	Wait		14:01	Air Blank	0.00	14:02	Subject Test	0.00	14:02	Breath Volume	1.570 LITERS		Air Blank	0.00	14:03	Cal Check	0.08	14:03	Air Blank	0.00	14:04	 SIGNATURE OF OPERATOR	
Test	g/210L	Time																																					
Air Blank	0.00	13:57																																					
Subject Test	0.00	13:59																																					
Breath Volume	1.617 LITERS																																						
Air Blank	0.00	14:00																																					
Wait		14:01																																					
Air Blank	0.00	14:02																																					
Subject Test	0.00	14:02																																					
Breath Volume	1.570 LITERS																																						
Air Blank	0.00	14:03																																					
Cal Check	0.08	14:03																																					
Air Blank	0.00	14:04																																					
NAME PAPE, CHRISTOPHER		BADGE # 0000000484																																					
AGENCY BOARD OF TESTS		PERMIT # 0000023020																																					



Test Result  
BrAC g/210L  
**0.00**

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180)

Section 3

days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460





9/17

**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 11:33

Test	g/210L	Time
Air Blank	0.000	11:34
Cal Check	0.022	11:34
Air Blank	0.000	11:35
Cal Check	0.021	11:36
Air Blank	0.000	11:36
Cal Check	0.021	11:37
Air Blank	0.000	11:37
Cal Check	0.021	11:38
Air Blank	0.000	11:38
Cal Check	0.021	11:39
Air Blank	0.000	11:40
Cal Check	0.021	11:40
Air Blank	0.000	11:41
Cal Check	0.020	11:42
Air Blank	0.000	11:42
Cal Check	0.021	11:43
Air Blank	0.000	11:43
Cal Check	0.021	11:44
Air Blank	0.000	11:45
Cal Check	0.020	11:45
Air Blank	0.000	11:46
Cal Check Stats		
Average	0.0209	
Std Dev	0.0005	
Rel Std Dev(%)	2.7158	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: <b>Guth</b>	Manufactured Date: <b>04/07/2021</b>	Concentration: <b>0.02</b>
Lot No.: <b>21120</b>	Expiration Date: <b>04/07/2023</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature: 		Date: <b>03/08/2022</b>
Name: <b>Christopher Pape</b>		Permit No.: <b>23020</b>



10/17

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 11:53

Test	g/210L	Time
Air Blank	0.000	11:53
Cal Check	0.041	11:54
Air Blank	0.000	11:54
Cal Check	0.041	11:55
Air Blank	0.000	11:56
Cal Check	0.042	11:56
Air Blank	0.000	11:57
Cal Check	0.042	11:58
Air Blank	0.000	11:58
Cal Check	0.042	11:59
Air Blank	0.000	11:59
Cal Check	0.041	12:00
Air Blank	0.000	12:01
Cal Check	0.041	12:01
Air Blank	0.000	12:02
Cal Check	0.041	12:02
Air Blank	0.000	12:03
Cal Check	0.041	12:04
Air Blank	0.000	12:04
Cal Check	0.042	12:05
Air Blank	0.000	12:05
Cal Check Stats		
Average	0.0414	
Std Dev	0.0005	
Rel Std Dev(%)	1.2473	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: Guth	Manufactured Date: 03/01/2021	Concentration: 0.04
Lot No.: 21070	Expiration Date: 03/01/2023	Equipment Model: Guth 2100
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: 03/08/2022	
Name: Christopher Pape	Permit No.: 23020	





11/17

**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 12:11

Test	g/210L	Time
Air Blank	0.000	12:11
Cal Check	0.101	12:12
Air Blank	0.000	12:13
Cal Check	0.101	12:13
Air Blank	0.000	12:14
Cal Check	0.101	12:15
Air Blank	0.000	12:15
Cal Check	0.102	12:16
Air Blank	0.000	12:16
Cal Check	0.101	12:17
Air Blank	0.000	12:18
Cal Check	0.101	12:18
Air Blank	0.000	12:19
Cal Check	0.101	12:20
Air Blank	0.000	12:20
Cal Check	0.101	12:21
Air Blank	0.000	12:21
Cal Check	0.101	12:22
Air Blank	0.000	12:23
Cal Check	0.101	12:23
Air Blank	0.000	12:24
Cal Check Stats		
Average	0.1011	
Std Dev	0.0003	
Rel Std Dev(%)	0.3127	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: **Guth**

Manufactured Date: **06/08/2021**

Concentration: **0.10**

Lot No.: **21190**

Expiration Date: **06/08/2023**

Equipment Model: **Guth 2100**

The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.

Signature

Date:

**03/08/2022**

Name **Christopher Pape**

Permit No.: **23020**



12/17

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov


Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT 2  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-004923  
03/08/2022 12:28

Test	g/210L	Time
Air Blank	0.000	12:29
Cal Check	0.201	12:29
Air Blank	0.000	12:30
Cal Check	0.200	12:31
Air Blank	0.000	12:31
Cal Check	0.201	12:32
Air Blank	0.000	12:33
Cal Check	0.201	12:33
Air Blank	0.000	12:34
Cal Check	0.201	12:35
Air Blank	0.000	12:35
Cal Check	0.201	12:36
Air Blank	0.000	12:36
Cal Check	0.201	12:37
Air Blank	0.000	12:38
Cal Check	0.200	12:38
Air Blank	0.000	12:39
Cal Check	0.201	12:40
Air Blank	0.000	12:40
Cal Check	0.200	12:41
Air Blank	0.000	12:41
Cal Check Stats		
Average	0.2007	
Std Dev	0.0004	
Rel Std Dev(%)	0.2406	

NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.

Manufacturer: <b>Guth</b>	Manufactured Date: <b>12/03/2020</b>	Concentration: <b>0.20</b>
Lot No.: <b>20510</b>	Expiration Date: <b>12/03/2022</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature: 	Date: <b>03/08/2022</b>	
Name: <b>Christopher Pape</b>	Permit No.: <b>23020</b>	



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 21120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 16, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0237% (w/vol) ethyl alcohol. The expiration date for this lot number is April 7, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08031603 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

RCV'D  
8/26/21  
CP

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**ACV'D  
8/26/21  
CP**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number **21070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 2, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0488%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 1, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**RCV'D  
8/26/21  
CP**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number **21190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **June 9, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **June 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

**Ted L. Pauley, President**  
**GUTH LABORATORIES, INC.**

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 20510 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 7, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2419% (w/vol) ethyl alcohol. The expiration date for this lot number is December 3, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN06231703 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

RCV'D  
8/26/21  
CP



Location Returned From Marlow PD	Unit Serial # 80-006163	Date: 17/17 3/31/22
-------------------------------------	----------------------------	------------------------

Inventory Process		Old Cylinder #: 05620080A2-024
<input checked="" type="checkbox"/> Unit checked in to BOT	<input checked="" type="checkbox"/> Unlink Gas Canister from Unit	<input type="checkbox"/> Old cylinder disposed in inventory
<input checked="" type="checkbox"/> New Unit Checked out	<input checked="" type="checkbox"/> Remove old cylinder from unit (After Lab Evaluation)	<input type="checkbox"/> Old cylinder archived

Check in Procedure	Notes from the Field
<input checked="" type="checkbox"/> Connect unit to ethernet connection 172.019.004, 134 <input checked="" type="checkbox"/> Download any stored records 5 <input checked="" type="checkbox"/> Change unit location to Serial number <input checked="" type="checkbox"/> Update Unit Date and time Previous time: 4/1/22 1053 New Time: 4/1/22 1055 <input checked="" type="checkbox"/> Execute a new Forms load on the unit	<input checked="" type="checkbox"/> Routine Swap <input type="checkbox"/> Issues in the Field

Lab Evaluation	
Breath Hose Hold Vacuum?: <input checked="" type="checkbox"/>	Test Chamber Hold Vacuum?: <input checked="" type="checkbox"/>
DVM	
3µm 12861	Chamber Temp 47.00
9µm 13218	Breath Hose Temp 45
	Snap Test ok? <input checked="" type="checkbox"/>
	RFI Detected? <input checked="" type="checkbox"/>

Peripheral Tests	
Speaker (F5) <input checked="" type="checkbox"/>	Power LED Green (F6) <input checked="" type="checkbox"/>
Gas Regulator (F11) <input checked="" type="checkbox"/>	Pump (F10) <input checked="" type="checkbox"/>
Red (F7) <input checked="" type="checkbox"/>	Fan (F9) <input checked="" type="checkbox"/>
	Receipt paper <input checked="" type="checkbox"/>
Atmospheric Sensor	
Pressure 970	Correction Factor 1.04
Flow Sensor Resistance 220	Tank Sensor Value 980
	Is regulator close to Tank Sensor Value? <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No

Sent to CMI (If needed)	Reason Sent
Date Sent _____	
Invoice Sent _____	
Date Returned _____	

Repairs needed: Replace Breath hose	Repairs Made: Breath hose replaced
	Ready for Bench Check? <input checked="" type="checkbox"/>

Inventory Process	<input type="checkbox"/> Bench Check Technician Report <input type="checkbox"/> Mock Subject Test Completed <input type="checkbox"/> Certificate of Calibration and Operation <input type="checkbox"/> Service Overview Completed	To complete if unit sent to CMI <input type="checkbox"/> CMI Workorder <input type="checkbox"/> CMI Certificate of Calibration <input type="checkbox"/> CMI Service Evaluation Form
<input type="checkbox"/> Linked Cylinder to Unit <input type="checkbox"/> Changed Unit Status <input type="checkbox"/> Bench Check Report Completed		