


Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 05-04-2018

SERIAL: 80-006164

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)
State Director of Tests for Alcohol
and Drug Influence

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 05/04/2018	TIME PROCEDURE BEGAN 10:12	INSTRUMENT LOCATION 80-006164
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-006164

TEST RESULTS

Diagnostics
PASS

Mouth Alcohol Test
PASS

RFI Test
PASS

Abort Test
PASS

Improper Sample Test
PASS

Interferent Detect Test
PASS

No Sample Given Test
PASS

Insufficient Sample Test
PASS

Air Blank Contamination Test
PASS

Range Exceeded Test
PASS


Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.017	0.018	0.019
.04	0.037	0.038	0.039
.10	0.099	0.098	0.101
.20	0.198	0.198	0.200

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 05/04/2018
NAME SMITH, JOSHUA	PERMIT # 0000016837



3600 N. Martin Luther King Bldg #9
 Oklahoma City, OK 73111
 405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 5 / 4 / 18 Start Time: 1015 End Time: 1145

INSTRUMENT

Intoxilyzer Model	Serial #
8000	80-006164

GAS CANISTER

LOT #	EXP Date
02618080A3 # 007	3-5-2020

REFERENCE

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR3591	DR3753	DR3594	DR3754
Concentration			
.02	.04	.10	.20
LOT #			
16400	16320	17130	17090
Manufactured Date			
12-12-2016	10-21-2016	04-03-2017	02-24-2017
Expiration Date			
12-12-2018	10-21-2018	04-03-2019	02-24-2019
Solution Commission Date			
5-4-18	5-4-18	5-4-18	5-4-18

REASON FOR BENCH CHECK

- POST REPAIR CYLINDER REPLACEMENT
 TROUBLESHOOTING ROUTINE MAINTENANCE

COMMENTS

 16837
 Technician Signature / Permit #
 Josh Smith
 Printed Name

Certificate of Analysis - Dry Gas	
8-4-18	.080
Date of Analysis	Labelled target value (g/210L)
007	.0782
Cylinder #	Average test result
Josh Smith <i>[Signature]</i>	
BOT Technician Name and Signature	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
 217-243-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 05/04/2018 09:30

Test	g/210L	Time
Air Blank	0.000	09:30
Cal Check	0.078	09:31
Air Blank	0.000	09:31
Cal Check	0.078	09:32
Air Blank	0.000	09:32
Cal Check	0.079	09:32
Air Blank	0.000	09:33
Cal Check	0.078	09:33
Air Blank	0.000	09:34
Cal Check	0.078	09:34
Air Blank	0.000	09:35
Cal Check	0.078	09:35
Air Blank	0.000	09:36
Cal Check	0.078	09:36
Air Blank	0.000	09:36
Cal Check	0.078	09:37
Air Blank	0.000	09:37
Cal Check	0.079	09:38
Air Blank	0.000	09:38
Cal Check	0.078	09:38
Air Blank	0.000	09:39
Cal Check Stats		
Average	0.0782	
Std Dev	0.0004	
Rel Std Dev(%)	0.5391	

Certificate of Analysis

Certificate ID: 10864
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 02618080A3 ✓
Expiration: 3/5/2020 ✓

[Signature] **REC'D**
 03/10/18

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:
 CMI Inc.
 316 East Ninth Street
 Owensboro, KY 4230
 Phone 866-835-0690
www.alcoholtest.com

[Signature] 10864
 Operator's Signature

*NIST Traceable Reference Material
 Cylinder No. CC274523 / Job No. 09160306
 Certified 362.2 µmol/mol Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
 Specialty Gas Lab Tech

02-16-18
 Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 05/04/2018		ARREST TIME 1040		COUNTY # 55	CITY # 9	CITATION # MOCK				
ARREST LOCATION MOCK				CITY MOCK		COUNTY OKLAHOMA				
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 03/20/2001		SEX M	RACE W	HEIGHT 511	WEIGHT 218	
ADDRESS MOCK			CITY MOCK			STATE MO		ZIP CODE 91212		
DRIVER LICENSE # MOCK			EXPIRATION DATE 09/09/2018		STATE MO	CLASS D	RESTRICTIONS NONE	ENDORSEMENTS NONE		
VEHICLE MAKE NONE			MODEL NONE			TAG # NONE		STATE OK	CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. ROUTINE MAINTENANCE. NEW GAS CANISTER INSTALLED.


Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006164	SPECIALIST BOARD OF TESTS	
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 16837	
MANUFACTURER ILMO		LOT # 02618080A3	DATE CYLINDER INSTALLED 05/04/2018	DATE CYLINDER EXPIRES 03/05/2020
OBSERVATION START 1100	TEST DATE 05/04/2018	MOUTHPIECE G	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.	
RESULT TABLE			 I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.	
Test	g/210L	Time		
Air Blank	0.00	11:27		
Subject Test	0.00	11:27		
Breath Volume	2.539 LITERS			
Air Blank	0.00	11:28		
Wait		11:29		
Air Blank	0.00	11:30		
Subject Test	0.00	11:30		
Breath Volume	2.167 LITERS			
Air Blank	0.00	11:31		
Cal Check	0.07	11:31		
Air Blank	0.00	11:32		
SIGNATURE OF OPERATOR			Test Result BrAC g/210L 0.00	
NAME SMITH, JOSHUA				
BADGE # 0000000435	PERMIT # 0000016837			
AGENCY BOARD OF TESTS				

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a

Section 3

period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **05/04/2018** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **05/04/2018** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

No Temporary License Issued:
Test Below Legal Limit

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

Section 1

ARREST DATE 05/04/2018		ARREST TIME 1040		COUNTY # 55	CITY # 9	CITATION # MOCK		
ARREST LOCATION MOCK				CITY MOCK		COUNTY OKLAHOMA		
SUBJECT NAME MOCK, SUBJECT TEST				DATE OF BIRTH 03/20/2001	SEX M	RACE W	HEIGHT 511	WEIGHT 218
ADDRESS MOCK			CITY MOCK		STATE MO	ZIP CODE 91212		
DRIVER LICENSE # MOCK		EXPIRATION DATE 09/09/2018	STATE MO	CLASS D	RESTRICTIONS NONE	ENDORSEMENTS NONE		
VEHICLE MAKE NONE		MODEL NONE	TAG # NONE		STATE OK	CMV N	PLACARD, N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. ROUTINE MAINTENANCE. NEW GAS CANISTER INSTALLED.


Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-006164	SPECIALIST BOARD OF TESTS	
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 16837	
MANUFACTURER ILMO		LOT # 02618080A3	DATE CYLINDER INSTALLED 05/04/2018	DATE CYLINDER EXPIRES 03/05/2020
OBSERVATION START 1100	TEST DATE 05/04/2018	MOUTHPIECE G	A 15 minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.	
RESULT TABLE			 I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.	
Test	g/210L	Time		
Air Blank	0.00	11:27		
Subject Test	0.00	11:27		
Breath Volume	2.533 LITERS			
Air Blank	0.00	11:28		
Wait		11:29		
Air Blank	0.00	11:30		
Subject Test	0.00	11:30		
Breath Volume	2.167 LITERS			
Air Blank	0.00	11:31		
Cal Check	0.07	11:31		
Air Blank	0.00	11:32		
SIGNATURE OF OPERATOR			NAME SMITH, JOSHUA	
BADGE # 0000000435		PERMIT # 0000016837		
AGENCY BOARD OF TESTS				
			Test Result BrAC g/210L 0.00	

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a

Section 3

period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **05/04/2018** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **05/04/2018** Place (location when signed): _____ Signature of arresting officer: _____

NAME SMITH, JOSHUA		AGENCY BOARD OF TESTS	
BADGE 0000000435	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE "4054252460"

No Temporary License Issued:
Test Below Legal Limit

ARRESTEE'S COPY
BOT Form AFF02 07/2008

Certificate of Analysis - Wet Bath	
05-05-17	.020
Date of Analysis	Labelled target value (g/210L)
16400	.0195
Lot Number	Average test result (g/210L)
Jst Smith	
BOT Technician Name and Signature	



ORIES, INC.
 G, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

INTOXILYZER 8000
 Instrument Initialization
 8133.12
 14:48 05/05/2017

Certified Alcohol Reference Solution for Simulator

Rec'd
 4/01/2017
 #16837

Random Samples of Lot Number 16400 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 13, 2016, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 05/05/2017 14:56

Test	g/210L	Time
Air Blank	0.000	14:57
Cal Check	0.020	14:58
Air Blank	0.000	14:58
Cal Check	0.020	14:59
Air Blank	0.000	14:59
Cal Check	0.019	15:00
Air Blank	0.000	15:01
Cal Check	0.020	15:01
Air Blank	0.000	15:02
Cal Check	0.020	15:03
Air Blank	0.000	15:03
Cal Check	0.020	15:04
Air Blank	0.000	15:04
Cal Check	0.019	15:05
Air Blank	0.000	15:06
Cal Check	0.019	15:06
Air Blank	0.000	15:07
Cal Check	0.019	15:08
Air Blank	0.000	15:08
Cal Check	0.019	15:09
Air Blank	0.000	15:09
Cal Check Stats		
Average	0.0195	
Std Dev	0.0005	
Rel Std Dev(%)	2.7026	

0.0241% (w/vol) ethyl alcohol. The expiration date for this lot number is December 12, 2018 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN11241503 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

#16837

Certificate of Analysis – Wet Bath	
3-8-2017	.040
Date of Analysis	Labelled target value (g/210L)
16320	.0409
Lot Number	Average test result (g/210L)
Josh Smith <i>[Signature]</i>	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 BURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT – 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

REC'D
03/01/17
[Signature]
#16837

Random Samples of Lot Number **16320** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 25, 2016**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0495%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 21, 2018** at 11:59 PM.

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/08/2017
 SN 80-004923
 11:16

Test	g/210L	Time
Air Blank	0.000	11:16
Cal Check	0.041	11:17
Air Blank	0.000	11:17
Cal Check	0.041	11:18
Air Blank	0.000	11:19
Cal Check	0.040	11:19
Air Blank	0.000	11:20
Cal Check	0.041	11:21
Air Blank	0.000	11:21
Cal Check	0.041	11:22
Air Blank	0.000	11:22
Cal Check	0.041	11:23
Air Blank	0.000	11:24
Cal Check	0.041	11:24
Air Blank	0.000	11:25
Cal Check	0.040	11:26
Air Blank	0.000	11:26
Cal Check	0.042	11:27
Air Blank	0.000	11:27
Cal Check	0.041	11:28
Air Blank	0.000	11:29
Cal Check Stats		
Average	0.0409	
Std Dev	0.0005	
Rel Std Dev(%)	1.3878	


When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

[Signature]
#16837

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN1218501 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis - Wet Bath	
06/15/2017	.10
Date of Analysis	Labeled target value (g/210L)
17130	.1015
Lot Number	Average test result (g/210L)
 BOT Technician Name and Signature	




GUTH LABORATORIES, INC.
 HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

*Rec'd
06/09/2017*


ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 06/15/2017

SN 80-004923
 16:25

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **17130** of

Test	g/210L	Time
Air Blank	0.000	16:26
Cal Check	0.102	16:26
Air Blank	0.000	16:27
Cal Check	0.102	16:28
Air Blank	0.000	16:28
Cal Check	0.101	16:29
Air Blank	0.000	16:29
Cal Check	0.102	16:30
Air Blank	0.000	16:31
Cal Check	0.101	16:31
Air Blank	0.000	16:32
Cal Check	0.101	16:33
Air Blank	0.000	16:33
Cal Check	0.102	16:34
Air Blank	0.000	16:34
Cal Check	0.101	16:35
Air Blank	0.000	16:36
Cal Check	0.102	16:36
Air Blank	0.000	16:37
Cal Check	0.101	16:38
Air Blank	0.000	16:38
Cal Check Stats		
Average	0.1015	
Std Dev	0.0005	
Rel Std Dev(%)	0.5192	

Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 6, 2017**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1212%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 3, 2019** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**. ✓


The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley

Ted L. Pauley, President
 GUTH LABORATORIES, INC.


 Operator's Signature

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis - Wet Bath	
06-16-17	.20
Date of Analysis	Labelled target value (g/210L)
17090	.2045
Lot Number	Average test result (g/210L)
 Josh Smith	
BOT Technician Name and Signature	



ORATORIES, INC.
 HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

REC'D
 06/09/17

 16837

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 06/16/2017

SN 80-004923
 10:52


Certified Alcohol Reference Solution for Simulator


Test	g/210L	Time
Air Blank	0.000	10:53
Cal Check	0.204	10:53
Air Blank	0.000	10:54
Cal Check	0.205	10:55
Air Blank	0.000	10:55
Cal Check	0.205	10:56
Air Blank	0.000	10:57
Cal Check	0.205	10:57
Air Blank	0.000	10:58
Cal Check	0.205	10:59
Air Blank	0.000	10:59
Cal Check	0.205	11:00
Air Blank	0.000	11:00
Cal Check	0.204	11:01
Air Blank	0.000	11:02
Cal Check	0.204	11:02
Air Blank	0.000	11:03
Cal Check	0.204	11:04
Air Blank	0.000	11:04
Cal Check	0.204	11:05
Air Blank	0.000	11:05
Cal Check	0.000	11:05
Cal Check Stats		
Average	0.2045	
Std Dev	0.0005	
Rel Std Dev(%)	0.2577	

Random Samples of Lot Number 17090 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on February 27, 2017, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2416% (w/vol) ethyl alcohol. The expiration date for this lot number is February 24, 2019 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.


 Ted L. Pauley, President
 GUTH LABORATORIES, INC.


 Operator's Signature
 16837

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number FN08101505 whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.