

Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 06-16-2017

SERIAL: 80-006165

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- Bench Check
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)
State Director of Tests for Alcohol
and Drug Influence

**STATE OF OKLAHOMA - DEPARTMENT OF PUBLIC SAFETY
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

| | | |
|--------------------------------|-------------------------------|---------------------------------------|
| BENCH CHECK DATE 06/16/2017 | TIME PROCEDURE BEGAN 10:53 | INSTRUMENT LOCATION 80-006165 |
| INSTRUMENT TYPE Intoxilyzer | INSTRUMENT MODEL I-8000 | INSTRUMENT SERIAL NUMBER 80-006165 |

TEST RESULTS

Diagnostics
PASS

Mouth Alcohol Test
PASS

RFI Test
PASS

Abort Test
PASS

Improper Sample Test
PASS

Interferent Detect Test
PASS

No Sample Given Test
PASS

Insufficient Sample Test
PASS

Air Blank Contamination Test
PASS

Range Exceeded Test
PASS


Duplicate Breath Sampling Results

| Value | Test 1 | Test 2 | Control Test |
|-------|--------|--------|--------------|
| .02 | 0.017 | 0.018 | 0.019 |
| .04 | 0.037 | 0.036 | 0.037 |
| .10 | 0.100 | 0.100 | 0.101 |
| .20 | 0.201 | 0.200 | 0.201 |

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

| | |
|--|------------------------|
| SIGNATURE  | DATE 06/16/2017 |
| NAME SMITH, JOSHUA | PERMIT # 0000016837 |



3600 N. Martin Luther King Bldg #9
 Oklahoma City, OK 73111
 405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 06 / 16 / 17 Start Time: 1000 End Time: 1300

INSTRUMENT

Intoxilyzer Model 8000 Serial # 80-006165

GAS CANISTER

LOT # 09117080A2 # 017 EXP Date 6-5-2019

REFERENCE

| Simulator Model | | | |
|--------------------------|----------------|----------------|----------------|
| Guth 2100 | Guth 2100 | Guth 2100 | Guth 2100 |
| Serial # | | | |
| DR3591 | DR3753 | DR3594 | DR3754 |
| Concentration | | | |
| .02 | .04 | .10 | .20 |
| LOT # | | | |
| 16400 | 16320 | 15220 | 15240 |
| Manufactured Date | | | |
| 12-12-2016 | 10-21-2016 | 09-28-2015 | 11-11-2015 |
| Expiration Date | | | |
| 12-12-2018 | 10-21-2018 | 09-28-2017 | 11-11-2017 |
| Solution Commission Date | | | |
| <u>5-30-17</u> | <u>5-30-17</u> | <u>6-15-17</u> | <u>6-15-17</u> |

REASON FOR BENCH CHECK


- POST REPAIR
 CYLINDER REPLACEMENT
 TROUBLESHOOTING
 ROUTINE MAINTENANCE

COMMENTS

LEAKING REGULATOR REPLACED.

 16837
 Technician Signature / Permit #

Josh Smith
 Printed Name

| Certificate of Analysis - Dry Gas | |
|---|--------------------------------|
| 06/02/2017 | .080 |
| Date of Analysis | Labelled target value (g/210L) |
| 017 | .0797 |
| Cylinder # | Average test result |
|  BOT Technician Name and Signature | |



ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 06/02/2017 13:34

BRT Form 5.0
ISO/IEC 17025:2005 Accredited Laboratory

In accordance with BOT Policy and Procedure Statement BRT-2.1.0, Maintenance, the above referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 13:35 |
| Cal Check | 0.080 | 13:35 |
| Air Blank | 0.000 | 13:36 |
| Cal Check | 0.080 | 13:36 |
| Air Blank | 0.000 | 13:37 |
| Cal Check | 0.079 | 13:37 |
| Air Blank | 0.000 | 13:37 |
| Cal Check | 0.080 | 13:38 |
| Air Blank | 0.000 | 13:38 |
| Cal Check | 0.080 | 13:39 |
| Air Blank | 0.000 | 13:39 |
| Cal Check | 0.080 | 13:39 |
| Air Blank | 0.000 | 13:40 |
| Cal Check | 0.080 | 13:40 |
| Air Blank | 0.000 | 13:41 |
| Cal Check | 0.080 | 13:41 |
| Air Blank | 0.000 | 13:42 |
| Cal Check | 0.079 | 13:42 |
| Air Blank | 0.000 | 13:42 |
| Cal Check | 0.079 | 13:43 |
| Air Blank | 0.000 | 13:43 |
| Cal Check Stats | | |
| Average | 0.0797 | |
| Std Dev | 0.0004 | |
| Rel Std Dev(%) | 0.6060 | |

Certificate of Analysis

Certificate ID: 10144 ✓
 Part #: BAC105L080T
 Cylinder Size: 105L
 Lot Number: 09117080A2 ✓
 Expiration: 6/5/2019 ✓

*Rec'd
 06/01/17
 JS*

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)


Contents: 105 Liters @ 1000 psig 70°F (21°C)

| Component: | Concentration: | Accuracy: | Method: |
|------------|----------------|--------------------------|---------|
| Ethanol | 208 ppm | +/- 0.002 or 2% | NDIR |
| Nitrogen | Balance | BAC whichever is greater | |


 Operator's Signature 16837

*NIST Standard Reference Material
 Cylinder No. CC 274507 / Job No. 09160309
 Certified 362.2 µmol/mol Ethanol in Nitrogen
 for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).


 Specialist Gas Lab Tech

04/25/17
 Date



Distributed by:

CMI Inc.
 316 East Ninth Street
 Owensboro, KY 42303
 Phone 866-835-0690
www.alcoholtest.com

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCAION/DISQUALIFICATION

Section 1

| | | | | | | | | |
|---|--|----------------------------|--------------------------------------|------------------------------------|---------------------|---------------------------|--------------------------|----------------------|
| ARREST DATE 06/21/2017 | | ARREST TIME 0800 | | COUNTY # 55 | CITY # 52 | CITATION # MOCK | | |
| ARREST LOCATION MOCK | | | | CITY | | COUNTY OKLAHOMA | | |
| SUBJECT NAME MOCK, SUBJECT TEST | | | | DATE OF BIRTH 08/09/1961 | SEX M | RACE W | HEIGHT 510 | WEIGHT 165 |
| ADDRESS 3600 MARTIN LUTHER KING A | | | CITY OKLAHOMA CITY | | STATE OK | ZIP CODE 73133 | | |
| DRIVER LICENSE # T083454871 | | | EXPIRATION DATE 06/30/2019 | STATE OK | CLASS A | RESTRICTIONS | ENDORSEMENTS M | |
| VEHICLE MAKE MOCK | | MODEL MOCK | | TAG # MOCK | | STATE OK | CMV N | PLACARD N |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. POST REPAIR MAINTENANCE. NEW GAS CANISTER INSTALLED.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

| EQUIPMENT I-8000 | | SERIAL NUMBER 80-006165 | SPECIALIST BOARD OF TESTS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|--------------------------------|-----------------------------------|--|--|------|--------|------|-------|--|--|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|------|--|-------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|-----------|------|-------|-----------|------|-------|
| STANDARD Dry | | TARGET VALUE 0.080 | PERMIT NUMBER 16837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| MANUFACTURER ILMO | | LOT # 09117080A2 | DATE CYLINDER INSTALLED 06/16/2017 | DATE CYLINDER EXPIRES 06/05/2019 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| OBSERVATION START 0900 | TEST DATE 06/21/2017 | MOUTHPIECE G | <p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td colspan="3">-----</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:20</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>09:20</td></tr> <tr><td> Breath Volume</td><td>1.316 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:20</td></tr> <tr><td>Wait</td><td></td><td>09:22</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:23</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>09:23</td></tr> <tr><td> Breath Volume</td><td>2.390 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:23</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>09:24</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>09:24</td></tr> </tbody> </table> | | | | | Test | g/210L | Time | ----- | | | Air Blank | 0.00 | 09:20 | Subject Test | 0.00 | 09:20 | Breath Volume | 1.316 LITERS | | Air Blank | 0.00 | 09:20 | Wait | | 09:22 | Air Blank | 0.00 | 09:23 | Subject Test | 0.00 | 09:23 | Breath Volume | 2.390 LITERS | | Air Blank | 0.00 | 09:23 | Cal Check | 0.08 | 09:24 | Air Blank | 0.00 | 09:24 |
| Test | g/210L | Time | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| ----- | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 1.316 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Wait | | 09:22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 2.390 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Cal Check | 0.08 | 09:24 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:24 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RESULT TABLE | | | NAME SMITH, JOSHUA | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | | BADGE # 0000000435 | PERMIT # 0000016837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | | AGENCY BOARD OF TESTS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification.** Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **06/21/2017** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **06/21/2017** Place (location when signed): _____ Signature of arresting officer: _____

| | | | |
|------------------------------|---------------------------------------|---------------------------------|------------------------------|
| NAME SMITH, JOSHUA | | AGENCY BOARD OF TESTS | |
| BADGE 0000000435 | AGENCY ADDRESS INTER-AGENCY | ZIP CODE 00000 | PHONE "4054252460" |

OTHER WITNESSES:

Section 5

| | | | | |
|---|------|-------|---------|-------|
| 1 | NAME | TITLE | ADDRESS | PHONE |
| 2 | NAME | TITLE | ADDRESS | PHONE |

No Temporary License Issued:
Test Below Legal Limit

OFFICER/DPS COPY
BOT Form AFF01 07/2008

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCAION/DISQUALIFICATION

Section 1

| | | | | | | | | |
|---|--|----------------------------|--------------------------------------|------------------------------------|---------------------|---------------------------|--------------------------|----------------------|
| ARREST DATE 06/21/2017 | | ARREST TIME 0800 | | COUNTY # 55 | CITY # 52 | CITATION # MOCK | | |
| ARREST LOCATION MOCK | | | | CITY | | COUNTY OKLAHOMA | | |
| SUBJECT NAME MOCK, SUBJECT TEST | | | | DATE OF BIRTH 08/09/1961 | SEX M | RACE W | HEIGHT 510 | WEIGHT 165 |
| ADDRESS 3600 MARTIN LUTHER KING A | | | CITY OKLAHOMA CITY | | STATE OK | ZIP CODE 73133 | | |
| DRIVER LICENSE # T083454871 | | | EXPIRATION DATE 06/30/2019 | STATE OK | CLASS A | RESTRICTIONS | ENDORSEMENTS M | |
| VEHICLE MAKE MOCK | | MODEL MOCK | | TAG # MOCK | | STATE OK | CMV N | PLACARD N |

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST. POST REPAIR MAINTENANCE. NEW GAS CANISTER INSTALLED.

Describe person's condition (odor, actions, etc.):

NONE

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

| EQUIPMENT I-8000 | | SERIAL NUMBER 80-006165 | SPECIALIST BOARD OF TESTS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|--------------------------------|-----------------------------------|--|--|------|--------|------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|------|--|-------|-----------|------|-------|--------------|------|-------|---------------|--------------|--|-----------|------|-------|-----------|------|-------|-----------|------|-------|
| STANDARD Dry | | TARGET VALUE 0.080 | PERMIT NUMBER 16837 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| MANUFACTURER ILMO | | LOT # 09117080A2 | DATE CYLINDER INSTALLED 06/16/2017 | DATE CYLINDER EXPIRES 06/05/2019 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| OBSERVATION START 0900 | TEST DATE 06/21/2017 | MOUTHPIECE G | <p style="text-align: center;">A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;">SIGNATURE OF OPERATOR _____</p> <p>NAME SMITH, JOSHUA</p> <p>BADGE # 0000000435</p> <p>PERMIT # 0000016837</p> <p>AGENCY BOARD OF TESTS</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr> <td>Air Blank</td> <td>0.00</td> <td>09:20</td> </tr> <tr> <td>Subject Test</td> <td>0.00</td> <td>09:20</td> </tr> <tr> <td> Breath Volume</td> <td>1.316 LITERS</td> <td></td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>09:20</td> </tr> <tr> <td>Wait</td> <td></td> <td>09:22</td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>09:23</td> </tr> <tr> <td>Subject Test</td> <td>0.00</td> <td>09:23</td> </tr> <tr> <td> Breath Volume</td> <td>2.390 LITERS</td> <td></td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>09:23</td> </tr> <tr> <td>Cal Check</td> <td>0.08</td> <td>09:24</td> </tr> <tr> <td>Air Blank</td> <td>0.00</td> <td>09:24</td> </tr> </tbody> </table> | | | | | Test | g/210L | Time | Air Blank | 0.00 | 09:20 | Subject Test | 0.00 | 09:20 | Breath Volume | 1.316 LITERS | | Air Blank | 0.00 | 09:20 | Wait | | 09:22 | Air Blank | 0.00 | 09:23 | Subject Test | 0.00 | 09:23 | Breath Volume | 2.390 LITERS | | Air Blank | 0.00 | 09:23 | Cal Check | 0.08 | 09:24 | Air Blank | 0.00 | 09:24 |
| Test | g/210L | Time | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 1.316 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Wait | | 09:22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Subject Test | 0.00 | 09:23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Breath Volume | 2.390 LITERS | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Cal Check | 0.08 | 09:24 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Air Blank | 0.00 | 09:24 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. To appeal the driver's license revocation/disqualification, the Department of Public Safety must receive your written request for an Administrative Hearing **within fifteen (15) days from the date you received notice of revocation or disqualification**. Submit your **written** request in person at 3600 N. MLK Ave. or by mail to Legal Division, Department of Public Safety, P.O. Box 11415, OKC, OK, 73136

Section 3

SERVING OF NOTICE: I personally hand-delivered a copy of the above containing the Notice of Revocation to the person named in Section 1 above.

Date served: **06/21/2017** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under **penalty of perjury** under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date: **06/21/2017** Place (location when signed): _____ Signature of arresting officer: _____

| | | | |
|------------------------------|---------------------------------------|---------------------------------|------------------------------|
| NAME SMITH, JOSHUA | | AGENCY BOARD OF TESTS | |
| BADGE 0000000435 | AGENCY ADDRESS INTER-AGENCY | ZIP CODE 00000 | PHONE "4054252460" |

No Temporary License Issued:
Test Below Legal Limit

ARRESTEE'S COPY
BOT Form AFF02 07/2008

Certificate of Analysis - Wet Bath



| | |
|-----------------------------------|--------------------------------|
| 05-05-17 | .020 |
| Date of Analysis | Labelled target value (g/210L) |
| 16488 | .0195 |
| Lot Number | Average test result (g/210L) |
| Josh Smith | |
| BOT Technician Name and Signature | |

BRT Form 6.0

GUTH LABORATORIES, INC.

16, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

INTOXILYZER 8000
Instrument Initialization
8133.12
14:48 05/05/2017

Certified Alcohol Reference Solution for Simulator

Rec'd
4/01/2017
#16837

Random Samples of Lot Number 16400 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 13, 2016, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain

0.0241% (w/vol) ethyl alcohol. The expiration date for this lot number is December 12, 2018 at 11:59 PM.

ALPHA INSTRUMENT 2
Intoxilyzer - Alcohol Analyzer
Model 8000
05/05/2017
SN 80-004923
14:56

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 14:57 |
| Cal Check | 0.020 | 14:58 |
| Air Blank | 0.000 | 14:58 |
| Cal Check | 0.020 | 14:59 |
| Air Blank | 0.000 | 14:59 |
| Cal Check | 0.019 | 15:00 |
| Air Blank | 0.000 | 15:01 |
| Cal Check | 0.020 | 15:01 |
| Air Blank | 0.000 | 15:02 |
| Cal Check | 0.020 | 15:03 |
| Air Blank | 0.000 | 15:03 |
| Cal Check | 0.020 | 15:04 |
| Air Blank | 0.000 | 15:04 |
| Cal Check | 0.019 | 15:05 |
| Air Blank | 0.000 | 15:06 |
| Cal Check | 0.019 | 15:06 |
| Air Blank | 0.000 | 15:07 |
| Cal Check | 0.019 | 15:08 |
| Air Blank | 0.000 | 15:08 |
| Cal Check | 0.019 | 15:09 |
| Air Blank | 0.000 | 15:09 |
| Cal Check Stats | | |
| Average | 0.0195 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 2.7026 | |

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.020 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11241503 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

#16837

| Certificate of Analysis - Wet Bath | |
|------------------------------------|-------------------------------|
| 3-8-2017 | .040 |
| Date of Analysis | Labeled target value (g/210L) |
| 16320 | .0409 |
| Lot Number | Average test result (g/210L) |
| Josh Smith <i>[Signature]</i> | |
| BOT Technician Name and Signature | |



GUTH LABORATORIES, INC.
 BURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

RECD
03/01/17
[Signature]
#16337

Random Samples of Lot Number 16320 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 25, 2016, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0495% (w/vol) ethyl alcohol. The expiration date for this lot number is October 21, 2018 at 11:59 PM.

ALPHA INSTRUMENT 2
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004923
 03/08/2017 11:16

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 11:16 |
| Cal Check | 0.041 | 11:17 |
| Air Blank | 0.000 | 11:17 |
| Cal Check | 0.041 | 11:18 |
| Air Blank | 0.000 | 11:19 |
| Cal Check | 0.040 | 11:19 |
| Air Blank | 0.000 | 11:20 |
| Cal Check | 0.041 | 11:21 |
| Air Blank | 0.000 | 11:21 |
| Cal Check | 0.041 | 11:22 |
| Air Blank | 0.000 | 11:22 |
| Cal Check | 0.041 | 11:23 |
| Air Blank | 0.000 | 11:24 |
| Cal Check | 0.041 | 11:24 |
| Air Blank | 0.000 | 11:25 |
| Cal Check | 0.040 | 11:26 |
| Air Blank | 0.000 | 11:26 |
| Cal Check | 0.042 | 11:27 |
| Air Blank | 0.000 | 11:27 |
| Cal Check | 0.041 | 11:28 |
| Air Blank | 0.000 | 11:29 |
| Cal Check Stats | | |
| Average | 0.0409 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 1.3878 | |

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.040 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

[Signature]
#16337

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FNI218501 whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



BRT Form 6.0

| | |
|------------------------------------|------------------|
| Certificate of Analysis - Wet Bath | |
| Date of Analysis | 12-12-2016 |
| Labelled target value (g/210L) | .100 |
| Lot Number | 15220 |
| Average test result (g/210L) | 0.0999 |
| BOIT Technician Name and Signature | JOSH SMITH 16837 |

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.



GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

REC'D
11-4-2015
#16837

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 30, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is September 28, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Intoxilyzer - Alcohol Analyzer
Model 8000
12/12/2016

SN 80-002591
14:04

| Test | g/210L | Time |
|-------------------|--------|-------|
| Air Blank | 0.000 | 14:04 |
| Cal Check | 0.100 | 14:05 |
| Air Blank | 0.000 | 14:05 |
| Cal Check | 0.100 | 14:06 |
| Air Blank | 0.000 | 14:07 |
| Cal Check | 0.100 | 14:07 |
| Air Blank | 0.000 | 14:08 |
| Cal Check | 0.100 | 14:09 |
| Air Blank | 0.000 | 14:09 |
| Cal Check | 0.099 | 14:10 |
| Air Blank | 0.000 | 14:10 |
| Cal Check | 0.100 | 14:11 |
| Air Blank | 0.000 | 14:12 |
| Cal Check | 0.100 | 14:12 |
| Air Blank | 0.000 | 14:13 |
| Cal Check | 0.100 | 14:14 |
| Air Blank | 0.000 | 14:14 |
| Cal Check | 0.100 | 14:15 |
| Air Blank | 0.000 | 14:16 |
| Cal Check | 0.100 | 14:16 |
| Air Blank | 0.000 | 14:17 |
| Cal Check Stats | | |
| Average | 0.0999 | |
| Std. Dev | 0.0003 | |
| Rel. Std. Dev (%) | 0.3165 | |

212-16



BRT Form 6.0

Certificate of Analysis - Wet Bath

12-12-16

Date of Analysis

15240

Lot Number

0.200

Labelled target value (g/210L)

0.2006

Average test result (g/210L)

#16837

Ted Pauley

BOI Technician Name and Signature

In accordance with BRT - 2.1.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

REC'D
11-4-2015
#16837

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15240 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on November 12, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.2418% (w/vol) ethyl alcohol. The expiration date for this lot number is November 11, 2017 at 11:59 PM.

ALPHA INSTRUMENT
Intoxilyzer - Alcohol Analyzer
Model 8000
12/12/2016

SN 80-002591
13:42

| Test | g/210L | Time |
|-----------------|--------|-------|
| Air Blank | 0.000 | 13:42 |
| Cal Check | 0.202 | 13:43 |
| Air Blank | 0.000 | 13:43 |
| Cal Check | 0.201 | 13:44 |
| Air Blank | 0.000 | 13:45 |
| Cal Check | 0.201 | 13:45 |
| Air Blank | 0.000 | 13:46 |
| Cal Check | 0.201 | 13:47 |
| Air Blank | 0.000 | 13:47 |
| Cal Check | 0.200 | 13:48 |
| Air Blank | 0.000 | 13:49 |
| Cal Check | 0.200 | 13:49 |
| Air Blank | 0.000 | 13:50 |
| Cal Check | 0.201 | 13:51 |
| Air Blank | 0.000 | 13:51 |
| Cal Check | 0.200 | 13:52 |
| Air Blank | 0.000 | 13:52 |
| Cal Check | 0.199 | 13:53 |
| Air Blank | 0.000 | 13:54 |
| Cal Check | 0.201 | 13:54 |
| Air Blank | 0.000 | 13:55 |
| Cal Check Stats | | |
| Average | 0.2006 | |
| Std Dev | 0.0008 | |
| Rel Std Dev(%) | 0.4203 | |

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.200 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03211401 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

12-12-16