



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT
Chairman


Joshua Smith
Director

ATTESTATION

I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 (11)(a) this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-007493 , in compliance with the agency's Maintenance Policy and Procedures.

This packet consists of 17 pages.

Signed 

Name of Position: Director/Records Custodian

Date of Attestation: 09/03/2021

Service Overview

Board of Tests for Alcohol and Drug Influence
Enhancing public safety by ensuring the accuracy and scientific reliability
of blood and breath testing



3600 N. Martin Luther King Bldg #9
Oklahoma City, OK 73111
405-425-2460

BENCH CHECK DATE: 09-03-2021

SERIAL: 80-007493

DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Calibration and Operation
- Bench Check
- CMI Certificate of Calibration - 08/25/2021
- Bench Check Technician's Report
- BOT Dry Gas Certificate of Analysis
- Mock Subject Test
- BOT Certificates of Analysis on Certified Solutions (4)
- CMI Packing Slip 44120
- Statement of Warranty



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Administrative Office at:
Board of Tests for Alcohol and Drug Influence
Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307

Certificate of Calibration and Operation

This is to certify that the calibration of **INTOXILYZER 8000**, **serial number 80-007493**, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R 126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.

In addition the operational procedures and functions of this Intoxilyzer have also been tested and found to be in compliance with the Oklahoma Board of Tests for Alcohol and Drug Influence - Rules and Regulations OAC Title 40: 25-1-2 and 30-1-3 and is hereby certified for use as an evidential breath testing device within this state.

CERTIFICATION TYPE	DATE
Manufacturer Calibration	08/25/21
Board of Tests Bench Check	09/03/21



Presentation of this form certifies that the Intoxilyzer, SN: 80-007493, functioned properly at the time of the breath test and hereby certifies the breath test result as valid.

Date:

Re: Request for Log of Test on Intoxilyzer 8000

Name:

Dear Sir:

The Log of Test information from a breath test conducted on a Intoxilyzer 8000 is printed in Section 2 of the Affidavit and is no longer retained as a separate form as with the Intoxilyzer 5000. The 8000 uses a nitrogen/ethanol dry gas for the external control test. The dry gas cylinder has the capacity to perform approximately 325 tests and does not require maintenance until its expiration date or can no longer provide the required pressure to perform a control test. The cylinder has an expiration date of two years from date filled at the manufacturer and is programmed in the 8000 upon installation, both dates are printed in Section 2 of the Affidavit. The Intoxilyzer is programmed to disable itself when the date of expiration is reached or the pressure monitor indicates 25 lbs psi.

Due to the reliability and stability of dry gas, maintenance on the 8000 is not required every 30 days or 25 tests, therefore maintenance is no longer performed by officers in the field. The Board of Tests for Alcohol and Drug Influence certifies each Intoxilyzer 8000 for field use by performing a bench check. The bench check is designed to check the operation and procedures programmed in the 8000 to ensure it is functioning properly for use in Oklahoma. A bench check is performed every time a new dry gas cylinder is installed or the instrument has returned from the manufacturer for repair. Once the instrument passes the bench check, all that is required to set it up for field use is to plug it in and turn it on and does not require a permit to set up. The Board of Tests has the ability to monitor each instrument once it is set up in the field.

Each Intoxilyzer has a database of valid permit numbers that it checks prior to allowing a breath test to be performed. The database is updated on a regular basis, if a permit number is invalid that officer will not be able to conduct a breath test on the 8000. If the 8000 prints an affidavit the breath test operator listed in Section 2 had a valid permit at the time of the test and the breath test is valid.

If you have any further questions please feel free to call me.

Sincerely,



J. Robert Blakeburn (OBA11059)
State Director of Tests for Alcohol
and Drug Influence

**STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE
CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT**

BENCH CHECK DATE 09/03/2021	TIME PROCEDURE BEGAN 12:35:4	INSTRUMENT LOCATION 80-007493
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-007493

TEST RESULTS

Diagnostics

PASS

Mouth Alcohol Test

PASS

RFI Test

PASS

Abort Test

PASS

Improper Sample Test

PASS

Interferent Detect Test

PASS

No Sample Given Test

PASS

Insufficient Sample Test

PASS

Air Blank Contamination Test

PASS

Range Exceeded Test

PASS


Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.017	0.018	0.019
.04	0.038	0.039	0.040
.10	0.100	0.099	0.100
.20	0.200	0.198	0.197

PASS

CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 9/3/2021
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-007493, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58 No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).


Date Aug. 25, 2021 Signed [Signature]

Technician



316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev.A



3600 N. Martin Luther King Bldg #9
 Oklahoma City, OK 73111
 405-425-2460

BENCH CHECK TECHNICIAN'S REPORT

Date: 09/03/2021 **Start Time:** 12:35 **End Time:** 13:45

INSTRUMENT

Intoxilyzer Model	Serial #
8000	80-007493

GAS CANISTER

LOT #	EXP Date
15421080A1-027	08/05/2021

REFERENCE

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR2834	DR3700	DR2455	DR3470
Concentration			
.02	.04	.10	.20
LOT #			
20070	20060	20190	20160
Manufactured Date			
02/13/2020	02/10/2020	04/06/2020	03/18/2020
Expiration Date			
02/13/2022	02/10/2022	04/06/2022	03/18/2022
Solution Commission Date			
9/3/2021	9/3/2021	9/3/2021	9/3/2021

REASON FOR BENCH CHECK

- POST REPAIR CYLINDER REPLACEMENT
 TROUBLESHOOTING ROUTINE MAINTENANCE

COMMENTS

Christopher Pape 23020
 Technician Signature / Permit #

Christopher Pape #23020
 Printed Name

8/17

Certificate of Analysis - Dry Gas	
09/02/2021	0.080
Date of Analysis	Labelled target value (g/210L)
027	0.081
Cylinder #	Average test result
Christopher Pope <i>[Signature]</i>	
BOT Technician Name and Signature	



ALPHA INSTRUMENT
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-002591
 09/02/2021 08:51

In accordance with BOT Maintenance Policy and Procedure, this ILMO referenced dry gas cylinder is suitable for use as an external reference with the Intoxilyzer 8000.
 7 Eastgate Dr., P.O. Box 790 • Jacksonville, IL 62651-0790
 217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Test	g/210L	Time
Air Blank	0.000	08:51
Cal Check	0.081	08:52
Air Blank	0.000	08:52
Cal Check	0.080	08:53
Air Blank	0.000	08:53
Cal Check	0.081	08:53
Air Blank	0.000	08:54
Cal Check	0.081	08:54
Air Blank	0.000	08:55
Cal Check	0.081	08:55
Air Blank	0.000	08:56
Cal Check	0.081	08:56
Air Blank	0.000	08:57
Cal Check	0.081	08:57
Air Blank	0.000	08:57
Cal Check	0.081	08:58
Air Blank	0.000	08:58
Cal Check	0.081	08:59
Air Blank	0.000	08:59
Cal Check	0.080	09:00
Air Blank	0.000	09:00
Cal Check Stats		
Average	0.0808	
Std Dev	0.0004	
Rel Std Dev(%)	0.5218	

Certificate of Analysis

Certificate ID: 14000 ✓
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 15421080A1 ✓
Expiration: 8/5/2023 ✓

RCVD
8/31/21
CP

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.
 316 East Ninth Street
 Owensboro, KY 423
 Phone 866-835-0690
www.alcoholtest.com

[Signature]
 Operator's Signature

*Traceable to:
 Certified Reference Material - 262.4 µmol/mol
 Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
 Specialty Gas Lab Tech

[Signature]

07-15-2021
 Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

9/17

Section 1

ARREST DATE 09/03/2021		ARREST TIME 1300		COUNTY # 55	CITY #	CITATION #		
ARREST LOCATION BOT				CITY	COUNTY OKLAHOMA			
SUBJECT NAME NORTH, GAYLON LEVERNE				DATE OF BIRTH 08/09/1961	SEX M	RACE W	HEIGHT 510	WEIGHT 165
ADDRESS 3600 MARTIN LUTHER KING A			CITY OKLAHOMA CITY	STATE OK	ZIP CODE 73133			
DRIVER LICENSE # T083454871		EXPIRATION DATE 06/30/2019	STATE OK	CLASS A	RESTRICTIONS	ENDORSEMENTS M		
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK	CMV N	PLACARD N		

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):

ROUTINE MAINTNEANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-007493	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO		LOT # 15421080A1	DATE CYLINDER INSTALLED 09/03/2021	DATE CYLINDER EXPIRES 08/05/2023																																				
OBSERVATION START 1315	TEST DATE 09/03/2021	MOUTHPIECE G	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration. I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.  SIGNATURE OF OPERATOR																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>13:40</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:40</td></tr> <tr><td> Breath Volume</td><td>2.195 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:41</td></tr> <tr><td>Wait</td><td></td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:43</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:43</td></tr> <tr><td> Breath Volume</td><td>2.351 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:44</td></tr> <tr><td>Cal Check</td><td>0.07</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:45</td></tr> </tbody> </table>					Test	g/210L	Time	Air Blank	0.00	13:40	Subject Test	0.00	13:40	Breath Volume	2.195 LITERS		Air Blank	0.00	13:41	Wait		13:42	Air Blank	0.00	13:43	Subject Test	0.00	13:43	Breath Volume	2.351 LITERS		Air Blank	0.00	13:44	Cal Check	0.07	13:44	Air Blank	0.00	13:45
Test	g/210L	Time																																						
Air Blank	0.00	13:40																																						
Subject Test	0.00	13:40																																						
Breath Volume	2.195 LITERS																																							
Air Blank	0.00	13:41																																						
Wait		13:42																																						
Air Blank	0.00	13:43																																						
Subject Test	0.00	13:43																																						
Breath Volume	2.351 LITERS																																							
Air Blank	0.00	13:44																																						
Cal Check	0.07	13:44																																						
Air Blank	0.00	13:45																																						
NAME PAPE, CHRISTOPHER			AGENCY BOARD OF TESTS																																					
BADGE # 0000000484		PERMIT # 0000023020																																						

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

SERVING OF NOTICE: I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **09/03/2021** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Date: **09/03/2021** Place (location when signed): _____ Signature of arresting officer: _____

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

OTHER WITNESSES:

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

10/17

Section 1

ARREST DATE 09/03/2021		ARREST TIME 1300		COUNTY # 55	CITY #	CITATION #	
ARREST LOCATION BOT				CITY		COUNTY OKLAHOMA	
SUBJECT NAME NORTH, GAYLON LEVERNE				DATE OF BIRTH 08/09/1961	SEX M	RACE W	HEIGHT 510
ADDRESS 3600 MARTIN LUTHER KING A				CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73133
DRIVER LICENSE # T083454871		EXPIRATION DATE 06/30/2019	STATE OK	CLASS A	RESTRICTIONS	ENDORSEMENTS M	
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK	CMV N	PLACARD N	

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST

Describe person's condition (odor, actions, etc.):

ROUTINE MAINTNEANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-007493	SPECIALIST BOARD OF TESTS																																					
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO		LOT # 15421080A1	DATE CYLINDER INSTALLED 09/03/2021	DATE CYLINDER EXPIRES 08/05/2023																																				
OBSERVATION START 1315	TEST DATE 09/03/2021	MOUTHPIECE G	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"><i>Christopher Pape</i> 23020 SIGNATURE OF OPERATOR</p>																																					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>13:40</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:40</td></tr> <tr><td> Breath Volume</td><td>2.195 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:41</td></tr> <tr><td>Wait</td><td></td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:43</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>13:43</td></tr> <tr><td> Breath Volume</td><td>2.351 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:44</td></tr> <tr><td>Cal Check</td><td>0.07</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:45</td></tr> </tbody> </table>					Test	g/210L	Time	Air Blank	0.00	13:40	Subject Test	0.00	13:40	Breath Volume	2.195 LITERS		Air Blank	0.00	13:41	Wait		13:42	Air Blank	0.00	13:43	Subject Test	0.00	13:43	Breath Volume	2.351 LITERS		Air Blank	0.00	13:44	Cal Check	0.07	13:44	Air Blank	0.00	13:45
Test	g/210L	Time																																						
Air Blank	0.00	13:40																																						
Subject Test	0.00	13:40																																						
Breath Volume	2.195 LITERS																																							
Air Blank	0.00	13:41																																						
Wait		13:42																																						
Air Blank	0.00	13:43																																						
Subject Test	0.00	13:43																																						
Breath Volume	2.351 LITERS																																							
Air Blank	0.00	13:44																																						
Cal Check	0.07	13:44																																						
Air Blank	0.00	13:45																																						
RESULT TABLE			<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td>NAME</td><td colspan="2">PAPE, CHRISTOPHER</td></tr> <tr><td>BADGE #</td><td>0000000484</td><td>PERMIT # 0000023020</td></tr> <tr><td>AGENCY</td><td colspan="2">BOARD OF TESTS</td></tr> </table>		NAME	PAPE, CHRISTOPHER		BADGE #	0000000484	PERMIT # 0000023020	AGENCY	BOARD OF TESTS																												
NAME	PAPE, CHRISTOPHER																																							
BADGE #	0000000484	PERMIT # 0000023020																																						
AGENCY	BOARD OF TESTS																																							

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

SERVING OF NOTICE: I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: **09/03/2021** Signature of officer: _____ Badge #: _____

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct." Section 4

Date: **09/03/2021** Place (location when signed): _____ Signature of arresting officer: _____

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

No Temporary License Issued:
Test Below Legal Limit

ARRESTEE'S COPY
BOT Form AFF02 03/2020

Certificate of Analysis - Wet Bath	
8/11/2020	0.020
Date of Analysis	Labelled tank Value (g/210L)
20070	0.020
Lot Number	Average test result (g/210L)
christopher Pauley	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.

220 NORTH 5th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

ACV'D
6/1/2020
CP

ALPHA INSTRUMENT 2
Intoxilyzer - Alcohol Analyzer
Model 8000 SH 80-004923
08/11/2020 09:42

Test	g/210L	Time
Air Blank	0.000	09:42
Cal Check	0.020	09:43
Air Blank	0.000	09:43
Cal Check	0.020	09:44
Air Blank	0.000	09:45
Cal Check	0.020	09:45
Air Blank	0.000	09:46
Cal Check	0.020	09:47
Air Blank	0.000	09:47
Cal Check	0.021	09:48
Air Blank	0.000	09:48
Cal Check	0.020	09:49
Air Blank	0.000	09:49
Cal Check	0.020	09:50
Air Blank	0.000	09:51
Cal Check	0.020	09:51
Air Blank	0.000	09:52
Cal Check	0.020	09:53
Air Blank	0.000	09:53
Cal Check	0.019	09:54
Air Blank	0.000	09:54
Cal Check Stats		
Average	0.0200	
Std Dev	0.0004	
Rel Std Dev(%)	2.3569	

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 17, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0242%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 13, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.020 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Christopher Pauley
Operator's Signature

Ted L. Pauley
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN08031603** whose values are traceable to NIST. All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath	
7/22/2020	0.040
Date of Analysis	Labelled target value (g/210L)
20060	0.040
Lot Number	Average test result (g/210L)
Christopher Pauley	
BOT Technician Name and Signature	



GUTH LABORATORIES, INC.
 230 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

In accordance with BRT – 2.2.0, *Maintenance*, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

RCV'D
6/1/20
CP

ALPHA UNIT SUB
 Intoxilyzer – Alcohol Analyzer
 Model 8000 SN 80-003400
 07/22/2020 14:57

Random Samples of Lot Number **20060** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 11, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0483% (w/vol)** ethyl alcohol. The expiration date for this lot number is **February 10, 2022** at 11:59 PM.

Test	g/210L	Time
Air Blank	0.000	14:57
Cal Check	0.040	14:58
Air Blank	0.000	14:59
Cal Check	0.039	14:59
Air Blank	0.000	15:00
Cal Check	0.039	15:01
Air Blank	0.000	15:01
Cal Check	0.039	15:02
Air Blank	0.000	15:02
Cal Check	0.039	15:03
Air Blank	0.000	15:04
Cal Check	0.039	15:04
Air Blank	0.000	15:05
Cal Check	0.040	15:06
Air Blank	0.000	15:06
Cal Check	0.040	15:07
Air Blank	0.000	15:07
Cal Check	0.040	15:08
Air Blank	0.000	15:09
Cal Check	0.040	15:09
Air Blank	0.000	15:10
Cal Check Stats		
Average	0.0395	
Std Dev	0.0005	
Rel Std Dev(%)	1.3342	

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
 GUTH LABORATORIES, INC.

Operator's Signature

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN12181501** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis – Wet Bath



13/17

06/16/20	.100
Date of Analysis	Labelled target value (g/210L)
20190	.098
Lot Number	Average test result (g/210L)
John Smith	
BOD Technician Name and Signature	

GUTH LABORATORIES, INC.

BRT Form 6.0 GUTH LABORATORIES, INC. 17511-4511 • TELEPHONE: 717-564-5470

In accordance with BRT – 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

Rec'd
06/11/2020
[Signature]

CERTIFICATE OF ANALYSIS

ALPHA INSTRUMENT

Intoxilyzer – Alcohol Analyzer
Model 8000
06/16/2020

SN 80-002591
16:18

Certified Alcohol Reference Solution for Simulator

Test	g/210L	Time
Air Blank	0.000	16:18
Cal Check	0.099	16:19
Air Blank	0.000	16:20
Cal Check	0.098	16:20
Air Blank	0.000	16:21
Cal Check	0.097	16:22
Air Blank	0.000	16:22
Cal Check	0.098	16:23
Air Blank	0.000	16:24
Cal Check	0.098	16:24
Air Blank	0.000	16:25
Cal Check	0.098	16:26
Air Blank	0.000	16:26
Cal Check	0.098	16:27
Air Blank	0.000	16:27
Cal Check	0.097	16:28
Air Blank	0.000	16:29
Cal Check	0.098	16:29
Air Blank	0.000	16:30
Cal Check	0.098	16:31
Air Blank	0.000	16:31
Cal Check Stats		
Average	0.0979	
Std Dev	0.0005	
Rel Std Dev(%)	0.5798	

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

[Signature]
Operator's Signature 16037

[Signature]
Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Certificate of Analysis - Wet Bath	
07/07/2020	0.200
Date of Analysis	Labeled Test Value (g/210L)
20160	0.197
Lot Number	Average Test Result (g/210L)
Christopher Pope	0.2000
BRT Technician Name and Signature	



GUTH LABORATORIES, INC.
 590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-5470

BRT Form 6.0

In accordance with BRT - 2.2.0, Maintenance, the above referenced simulator solution is suitable for use as an external reference in maintenance of the Intoxilyzer 8000.

CERTIFICATE OF ANALYSIS

*Rec'd
05/2020
CP*

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 20, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.2427%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 18, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.200 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

LPRA UNIT SUB
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003400
 7/07/2020 15:25

ast	g/210L	Time
lr Blank	0.000	15:26
al Check	0.196	15:26
lr Blank	0.000	15:27
al Check	0.196	15:28
lr Blank	0.000	15:28
al Check	0.197	15:29
lr Blank	0.000	15:29
al Check	0.196	15:30
lr Blank	0.000	15:31
al Check	0.198	15:31
lr Blank	0.000	15:32
al Check	0.198	15:33
lr Blank	0.000	15:33
al Check	0.197	15:34
lr Blank	0.000	15:35
al Check	0.198	15:35
lr Blank	0.000	15:36
al Check	0.198	15:37
lr Blank	0.000	15:37
al Check	0.197	15:38
lr Blank	0.000	15:38
al Check Stats		
Average	0.1971	
Std Dev	0.0008	
Rel Std Dev(%)	0.4442	

[Signature]
 Operator's Signature

[Signature]
 Ted L. Pauley, President
 GUTH LABORATORIES, INC.

NIST Traceability:
 Testing was conducted using Cerilliant Reference Standard lot number **FN08101505** whose values are traceable to NIST.
 All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

CMI Inc
 316 E 9th Street
 Owensboro KY 42303
 USA
 Phone: 866-835-0690
 Fax: 270-685-6268



Packing Slip: 44120

Ship To: Oklahoma Board of Test 3600 N Martin Luther King Ave Attn: Joshua Smith Oklahoma City OK 73136 USA Phone: Fax: Email:	Sold To: Charles Danker Iowa Tribe Police Dept 335468 East 750 Road Perkins OK 74059 USA Phone: 405-547-5355 Fax: Email: cdanker@iowanation.org
--	--

Ship Date: 8/26/2021 F.O.B.: Origin Tracking #: 1Z4X04930349990795
 Ship Via: UPS Ground Carrier: United Parcel Service
 SCAC: UPSN

Line	Part Number/Description	Rev.	Planned Qty	Shipped Qty	PO Line
Sales Order: 933514		Your PO: PO032051			

LPaul@iowanation.org

Instrument Pricing includes ground transportation charges.

Instrument to be shipped to Oklahoma Board of Test (Oklahoma City) for registration and inspection before evidential use. The printer, printer cable, mouthpieces, printer paper, and gas cylinder may ship direct to the end user.

Notwithstanding any provisions or terms set forth in other contracts or agency purchasing web sites, the software contained in the Intoxilyzer 8000 is, in its entirety, a proprietary product of CMI, Inc. All intellectual property rights in and to software are retained by CMI, Inc., regardless of whether such software (or any portion thereof) was created prior to or in connection with any purchase. Placement and/or payment of a purchase order based upon this pricing proposal constitutes acceptance of this condition, which shall supersede any other agreements or understandings relating to the purchase of the Intoxilyzer 8000 and the use of its software and components.

9	0024800K / UNIT ASSY,I8000,IMAGER/ETH/IMA	A	1.00EA	1.00EA	
---	---	---	--------	--------	--

Mark For:

Phone:
Fax:

Serial Number: 80-007493

1	015111 / MOUTHPIECES,VALVED,BAG OF 100	-	1.00EA	1.00EA	
---	--	---	--------	--------	--

RCVD
 9/1/21 CP

Mark For:

Phone:
Fax:

2	015088 / PAPER,ROLL,THERMAL,58MM X 25MM	ND	3.00EA	3.00EA	
---	---	----	--------	--------	--

RCVD
 9/1/21 CP

Mark For:

Phone:
Fax:

3	690079 / PRINTER,LASER,BLACK,W/WIFI,I9000 ONLY		1.00EA	1.00EA	
---	--	--	--------	--------	--

RCVD
 8/31/21 CP

Mark For:

Phone:
Fax:

4	330540 / CABLE ASSY,USB,AMALE-BMALE,3 M	ND	1.00EA	1.00EA	
---	---	----	--------	--------	--

RCVD
 9/1/21 CP

CMI Inc
316 E 9th Street
Owensboro KY 42303
USA
Phone: 866-835-0690
Fax: 270-685-6268



Packing Slip: 44120

2 of 2

Line 6 Rel 1
10 011148 / GAS SYSTEM,MOBILE FLORIDA

Mark For:
Phone:
Fax:

D 1.00EA 1.00EA

Mark For:
Phone:
Fax:

Serial Number: GR-7493
Line 7 Rel 1
5 ✓ 340234 / CYL,EGS,.080g/210L,105L,ST

REV D
8/31/21
CP

1.00EA 1.00EA

Mark For:
Phone:
Fax:

Lot #: 15421080A1 Qty: 1.00
Line 8 Rel 1
6 03813312 / PROM,I8000,OKLAHOMA

ND 1.00EA 1.00EA

Mark For:
Phone:
Fax:

Line 9 Rel 1
7 02136216 / SOFTWARE, I8000 DSP, US MARKET

1.00EA 1.00EA

Mark For:
Phone:
Fax:

Line 10 Rel 1
8 02136107 / SOFTWARE,I8000,CPLD REV D

ND 1.00EA 1.00EA

Mark For:
Phone:
Fax:

Statement of Warranty

New Product Warranty

CMI, Inc. warrants that each new Intoxilyzer Instrument will be free from defects in material and workmanship, under normal use and service, for a period of one year from the date of invoice to the initial purchaser. CMI's obligation is limited to repairing or replacing, as CMI may elect, any part or parts of such product, which CMI determines to be defective in material or workmanship. Warranty repairs will be performed at the factory or at a factory authorized service center, freight prepaid.

The product, or part of the product, considered to be covered by the conditions of this warranty shall be returned, freight prepaid, in its original shipping container or similar protective container, to the factory, only after receipt of a Returned Material Authorization number from CMI. The repair or replacement part or product will be returned from CMI or the authorized service center, freight prepaid.

Warranty coverage extends only to the original purchaser and does not include abuse, misuse, cables, switches, internal batteries, peripheral accessories or use of the product for other than its intended purpose. This warranty also does not apply if the product is adversely affected by attaching a feature or device to it, or is on any way tampered with or modified, without express written permission from CMI, Inc.

Repair Product Warranty

Out of warranty product repairs are warranted for 90 days from the date of repair. This includes labor and those parts, which are replaced. If additional repair is required within the 90 day period, there will be a charge for any parts that were not originally replaced. Repairs made during the 90 day period that are unrelated to the original repair are not covered under the warranty.

There are no warranties expressed or implied, either oral or written, other than those contained in this warranty. In no event shall CMI be liable for any loss of profits or any indirect or consequential damages arising out of any such defect in material or workmanship.

CMI INC.
Intoxilyzer[®]
Breath Alcohol Testing Is Our Business

316 East 9th Street
Owensboro, KY 42303
1-866-835-0690
www.alcoholtest.com