



OSDH KN95 Mask Information

Manufacture	H-D #	Model	Sent	Tested	Results rec	% Eff
Nine Particles Medical Equipment co., ltd	HD-01	KN95- ear hanging type	6-Apr-20	7-Apr-20	7-Apr-20	98.4
ZHONGCHEN	HD-04	ZC959	13-Apr-20	14-Apr-20	14-Apr-20	96.3
Shenzhen Weifa Technology Co, LTD	HD-03	STMK-N9W01	13-Apr-20	14-Apr-20	14-Apr-20	97.3
HXFH100-003	HD-02	HXFH100-003	13-Apr-20	14-Apr-20	14-Apr-20	92.9
No Brand, GB2626-2006	HD-06	GB2626-2006/"CW" cert	15-Apr-20	16-Apr-20	16-Apr-20	99
Maozheng	HD-05	8600	15-Apr-20	16-Apr-20	16-Apr-20	98.8
Shenzhen Tianlong Century Tech	HD-10	EN 149.200/+A1:309	20-Apr-20	22-Apr-20	22-Apr-20	97.9
LEEPHC< (KN95)	HD-09	FFP2/1P	20-Apr-20	22-Apr-20	22-Apr-20	98.7
Lamdown, KN95Protection Level	HD-15	SD-KN95	21-Apr-20	23-Apr-20	24-Apr-20	98.1
3M	N/A	N951860	20-Apr-20	N/A	24-Apr-20	N/A
Makrite	HD-14	9500-N95	20-Apr-20	23-Apr-20	24-Apr-20	99.1
Gerson 1730	HD-12	1730/D1081730TY164A	20-Apr-20	23-Apr-20	24-Apr-20	97.8
Moldex	N/A	2200N95 M/L; NSN No. 42	20-Apr-20	N/A	24-Apr-20	N/A
Moldex	N/A	2211GN95-S	20-Apr-20	N/A	24-Apr-20	N/A
3M	N/A	8000/NIOSH N95 1289-0	20-Apr-20	N/A	24-Apr-20	N/A
Gerson 2130	HD_13	2130	20-Apr-20	23-Apr-20	24-Apr-20	99.0

The samples shaded in blue represent N95's (NIOSH approved). The samples tested were masks that were past their expiration date and approved to use by NIOSH. Some were not tested because they were in prime condition still.



Manufacture	H-D #	Model/AKA/Marking	Sent	Testing	Received Res	Where is Documentation	Picture	% Eff

Manufacture	HD #	Model	Sent	Tested	Received Res	% eff	Quantity
Dongguan Modern Children's Products Co,Ltd	HD-07	"Cert"	15-Apr-20	16-Apr-20	17-Apr-20	85.2%	18,000
Jusenyan	HD-08	FFP2 N95	16-Apr-20	16-Apr-20	4/16/2020 and 17-Apr-	89.3	4,550
Dongguan Hao Yang carbon filter beating products	HD-11	HY001/FFP2	20-Apr-20			71.3	4010

